

LOT MX 055 L
 Rev 1

CONTROL L

(Exp.)

2015-08-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l		2.1			2.1					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l		2.40			2.40					± 0.15
HB HGB	g/dl		6.0			5.7					± 0.4
	g/l		60			57					± 4
	mmol/l		3.71			3.52					± 0.25
HT HCT	%		16.3			16.6					± 2.0
	l/l		0.163			0.166					± 0.020
VGM MCV	µm ³ ; fl		68			69					± 4
TGMH MCH	pg		24.9			23.6					± 2.0
	fmol		1.55			1.47					± 0.12
CCMH MCHC	g/dl		36.6			34.3					± 3.0
	g/l		366			343					± 30
	mmol/l		22.75			21.27					± 1.86
IDR RDW	%		14.0			13.9					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l		74			76					± 20
VPM MPV	µm ³ ; fl		9.2			8.5					± 2.0
LYMPHO	#		1.34			1.32					± 0.40
	%		63.6			63.0					± 8.0
MONO	#		0.26			0.25					± 0.20
	%		12.5			12.0					± 6.0
GRANULO	#		0.50			0.53					± 0.40
	%		24.0			25.0					± 7.0

LOT

MX 055 N
Rev 1

CONTROL N

(Exp.)

2015-08-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l		7.8			7.7					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l		4.62			4.64					± 0.18
HB HGB	g/dl		13.5			13.2					± 0.5
	g/l		135			132					± 5
	mmol/l		8.38			8.22					± 0.31
HT HCT	%		37.9			38.0					± 2.5
	l/l		0.379			0.380					± 0.025
VGM MCV	µm ³ ; fl		82			82					± 4
TGMH MCH	pg		29.2			28.5					± 2.0
	fmol		1.81			1.77					± 0.12
CCMH MCHC	g/dl		35.6			34.8					± 3.0
	g/l		356			348					± 30
	mmol/l		22.13			21.59					± 1.86
IDR RDW	%		13.2			13.6					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l		260			265					± 40
VPM MPV	µm ³ ; fl		7.9			7.7					± 2.0
LYMPHO	#		2.69			2.46					± 0.40
	%		34.5			32.1					± 5.0
MONO	#		0.56			0.57					± 0.30
	%		7.1			7.4					± 4.0
GRANULO	#		4.55			4.64					± 0.60
	%		58.3			60.5					± 6.0

9930002-A
Ref: TEMP-0830 Rev.36



LOT

MX 055 H

CONTROL

H

(Exp.)

2015-08-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l		CRP 200			60 8P - 60 18P					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l										± 0.20
HB HGB	g/dl		18.1			18.1					± 0.6
	g/l		181			181					± 6
	mmol/l		11.24			11.25					± 0.37
HT HCT	%		51.0			51.5					± 3.0
	l/l		0.510			0.515					± 0.030
VGM MCV	µm ³ ; fl		89			89					± 4
TGMH MCH	pg		31.6			31.3					± 2.0
	fmol		1.96			1.94					± 0.12
CCMH MCHC	g/dl		35.5			35.2					± 3.0
	g/l		355			352					± 30
	mmol/l		22.04			21.83					± 1.86
IDR RDW	%		12.7			13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l		500			515					± 45
VPM MPV	µm ³ ; fl		8.2			8.0					± 2.0
LYMPHO	#		3.57			3.29					± 1.00
	%		17.5			16.5					± 5.0
MONO	#		1.21			1.10					± 0.80
	%		5.9			5.5					± 4.0
GRANULO	#		15.59			15.53					± 1.40
	%		76.4			77.7					± 7.0