






LOT		MX 075 L		CONTROL		L		 (Exp.)		2015-10-05 ( YYYY - MM - DD )	
PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
			CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.1			2.0					± 0.4	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.40			2.40					± 0.15	
HB HGB	g/dl	5.9			5.7					± 0.4	
	g/l	59			57					± 4	
	mmol/l	3.65			3.52					± 0.25	
HT HCT	%	16.3			16.6					± 2.0	
	l/l	0.163			0.166					± 0.020	
VGM MCV	µm <sup>3</sup> ; fl	68			69					± 4	
TGMH MCH	pg	24.5			23.6					± 2.0	
	fmol	1.52			1.47					± 0.12	
CCMH MCHC	g/dl	36.0			34.3					± 3.0	
	g/l	360			343					± 30	
	mmol/l	22.37			21.27					± 1.86	
IDR RDW	%	14.0			14.0					± 3.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	70			75					± 20	
VPM MPV	µm <sup>3</sup> ; fl	9.0			8.2					± 2.0	
LYMPHO	#	1.36			1.25					± 0.40	
	%	64.6			62.5					± 8.0	
MONO	#	0.26			0.26					± 0.20	
	%	12.5			13.0					± 6.0	
GRANULO	#	0.48			0.49					± 0.40	
	%	23.0			24.5					± 7.0	

LOT		MX 075 N		Rev 1		CONTROL N		 (Exp.)		2015-10-05 ( YYYY - MM - DD )	
PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
			CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l		7.5			7.3					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l		4.60			4.60					± 0.18
HB HGB	g/dl		13.4			13.2					± 0.5
	g/l		134			132					± 5
	mmol/l		8.32			8.22					± 0.31
HT HCT	%		36.8			36.8					± 2.5
	l/l		0.368			0.368					± 0.025
VGM MCV	µm <sup>3</sup> ; fl		80			80					± 4
TGMH MCH	pg		29.1			28.8					± 2.0
	fmol		1.81			1.79					± 0.12
CCMH MCHC	g/dl		36.4			36.0					± 3.0
	g/l		364			360					± 30
	mmol/l		22.61			22.33					± 1.86
IDR RDW	%		13.5			13.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l		250			252					± 40
VPM MPV	µm <sup>3</sup> ; fl		8.5			8.0					± 2.0
LYMPHO	#		2.66			2.45					± 0.40
	%		35.5			33.7					± 5.0
MONO	#		0.57			0.54					± 0.30
	%		7.6			7.4					± 4.0
GRANULO	#		4.26			4.29					± 0.60
	%		56.8			59.0					± 6.0

1219930092-A  
 Rel: TEMP-0830 Rev.37



LOT		MX 075 H		CONTROL		H		 (Exp.)		2015-10-05 ( YYYY - MM - DD )	
PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
			CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l		20.1			19.4					± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l		5.70			5.70					± 0.20
HB HGB	g/dl		17.8			17.6					± 0.6
	g/l		178			176					± 6
	mmol/l		11.05			10.94					± 0.37
HT HCT	%		50.2			50.2					± 3.0
	l/l		0.502			0.502					± 0.030
VGM MCV	µm <sup>3</sup> ; fl		88			88					± 4
TGMH MCH	pg		31.2			30.9					± 2.0
	fmol		1.94			1.92					± 0.12
CCMH MCHC	g/dl		35.5			35.1					± 3.0
	g/l		355			351					± 30
	mmol/l		22.04			21.82					± 1.86
IDR RDW	%		13.0			13.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l		520			520					± 45
VPM MPV	µm <sup>3</sup> ; fl		8.0			7.7					± 2.0
LYMPHO	#		3.09			2.80					± 1.00
	%		15.4			14.4					± 5.0
MONO	#		1.19			0.97					± 0.80
	%		5.9			5.0					± 4.0
GRANULO	#		15.76			15.56					± 1.40
	%		78.4			80.2					± 7.0