



LOT

MX 095 L
 Rev 1

CONTROL L

(Exp.)

2015-12-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l		CRP 200			60 8P - 60 18P					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l										± 0.15
HB HGB	g/dl		6.1			5.8					± 0.4
	g/l		61			58					± 4
	mmol/l		3.77			3.58					± 0.25
HT HCT	%		16.7			16.7					± 2.0
	l/l		0.167			0.167					± 0.020
VGM MCV	µm ³ ; fl		68			68					± 4
TGMH MCH	pg		24.8			23.5					± 2.0
	fmol		1.54			1.46					± 0.12
CCMH MCHC	g/dl		36.5			34.6					± 3.0
	g/l		365			346					± 30
	mmol/l		22.65			21.49					± 1.86
IDR RDW	%		14.0			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l		71			75					± 20
VPM MPV	µm ³ ; fl		8.8			8.2					± 2.0
LYMPHO	#		1.32			1.22					± 0.40
	%		62.6			61.0					± 8.0
MONO	#		0.23			0.22					± 0.20
	%		11.0			11.0					± 6.0
GRANULO	#		0.56			0.56					± 0.40
	%		26.5			28.0					± 7.0

LOT

MX 095 N
Rev 1

CONTROL N

(Exp.)

2015-12-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l		CRP 200			60 8P - 60 18P					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l										± 0.18
HB HGB	g/dl		7.7			7.5					± 0.5
	g/l		13.3			13.1					± 5
	mmol/l		8.26			8.15					± 0.31
HT HCT	%		36.5			36.5					± 2.5
	l/l		0.365			0.365					± 0.025
VGM MCV	µm ³ ; fl		81			81					± 4
TGMH MCH	pg		29.6			29.2					± 2.0
	fmol		1.84			1.81					± 0.12
CCMH MCHC	g/dl		36.5			36.0					± 3.0
	g/l		365			360					± 30
	mmol/l		22.66			22.37					± 1.86
IDR RDW	%		13.5			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l		255			255					± 40
VPM MPV	µm ³ ; fl		8.3			8.0					± 2.0
LYMPHO	#		2.73			2.55					± 0.40
	%		35.5			34.2					± 5.0
MONO	#		0.51			0.52					± 0.30
	%		6.7			7.0					± 4.0
GRANULO	#		4.45			4.41					± 0.60
	%		57.8			59.0					± 6.0

1219930092-A
Ref: TEMP-0830 Rev.37



LOT

MX 095 H

CONTROL

H

(Exp.)

2015-12-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l		CRP 200			60 8P - 60 18P					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l										± 0.20
HB HGB	g/dl		20.1			19.6					± 0.6
	g/l		18.0			17.9					± 6
	mmol/l		180			179					± 0.37
HT HCT	%		11.18			11.13					± 3.0
	l/l		49.7			50.2					± 0.030
VGM MCV	µm ³ ; fl		0.497			0.502					± 4
TGMH MCH	pg		88			88					± 2.0
	fmol		31.9			31.4					± 0.12
CCMH MCHC	g/dl		1.98			1.95					± 3.0
	g/l		36.2			35.7					± 30
	mmol/l		362			357					± 1.86
IDR RDW	%		22.48			22.18					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l		13.0			13.0					± 45
VPM MPV	µm ³ ; fl		505			505					± 2.0
LYMPHO	#		7.9			7.5					± 1.00
	%		3.09			2.83					± 5.0
MONO	#		15.4			14.4					± 0.80
	%		1.19			1.08					± 4.0
GRANULO	#		5.9			5.5					± 1.40
	%		15.76			15.62					± 7.0
			78.4			79.7					