



LOT MX 115 L
 Rev 1

CONTROL L

(Exp.)

2016-02-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l		CRP 200			60 8P - 60 18P					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l										± 0.15
HB HGB	g/dl		5.9			5.5					± 0.4
	g/l		59			55					± 4
	mmol/l		3.65			3.41					± 0.25
HT HCT	%		16.1			16.1					± 2.0
	l/l		0.161			0.161					± 0.020
VGM MCV	µm ³ ; fl		68			69					± 4
TGMH MCH	pg		24.8			23.5					± 2.0
	fmol		1.54			1.46					± 0.12
CCMH MCHC	g/dl		36.5			34.1					± 3.0
	g/l		365			341					± 30
	mmol/l		22.66			21.19					± 1.86
IDR RDW	%		13.7			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l		70			71					± 20
VPM MPV	µm ³ ; fl		8.9			8.0					± 2.0
LYMPHO	#		1.34			1.32					± 0.40
	%		63.6			63.0					± 8.0
MONO	#		0.20			0.19					± 0.20
	%		9.5			9.0					± 6.0
GRANULO	#		0.57			0.59					± 0.40
	%		27.0			28.0					± 7.0

LOT

MX 115 N
Rev 1

CONTROL N

(Exp.)

2016-02-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l		8.0			7.7					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l		4.55			4.52					± 0.18
HB HGB	g/dl		13.4			13.1					± 0.5
	g/l		134			131					± 5
	mmol/l		8.32			8.15					± 0.31
HT HCT	%		36.9			37.1					± 2.5
	l/l		0.369			0.371					± 0.025
VGM MCV	µm ³ ; fl		81			82					± 4
TGMH MCH	pg		29.5			29.1					± 2.0
	fmol		1.83			1.80					± 0.12
CCMH MCHC	g/dl		36.4			35.4					± 3.0
	g/l		364			354					± 30
	mmol/l		22.58			22.00					± 1.86
IDR RDW	%		12.9			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l		255			250					± 40
VPM MPV	µm ³ ; fl		8.0			7.5					± 2.0
LYMPHO	#		2.71			2.46					± 0.40
	%		33.9			32.1					± 5.0
MONO	#		0.50			0.46					± 0.30
	%		6.2			6.0					± 4.0
GRANULO	#		4.78			4.75					± 0.60
	%		59.8			62.0					± 6.0

1219930092-A
Ref: TEMP-0830 Rev.37



LOT

MX 115 H

CONTROL

H

(Exp.)

2016-02-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l		CRP 200			60 8P - 60 18P					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l										± 0.20
HB HGB	g/dl		18.3			18.1					± 0.6
	g/l		183			181					± 6
	mmol/l		11.36			11.25					± 0.37
HT HCT	%		51.4			51.3					± 3.0
	l/l		0.514			0.513					± 0.030
VGM MCV	µm ³ ; fl		89			89					± 4
TGMH MCH	pg		31.7			31.5					± 2.0
	fmol		1.97			1.95					± 0.12
CCMH MCHC	g/dl		35.6			35.3					± 3.0
	g/l		356			353					± 30
	mmol/l		22.13			21.95					± 1.86
IDR RDW	%		12.7			13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l		520			510					± 45
VPM MPV	µm ³ ; fl		7.7			7.2					± 2.0
LYMPHO	#		3.23			3.11					± 1.00
	%		15.9			16.0					± 5.0
MONO	#		1.43			1.07					± 0.80
	%		7.0			5.5					± 4.0
GRANULO	#		15.62			15.25					± 1.40
	%		76.9			78.2					± 7.0