

LOT MX 036 L  
Rev 1

CONTROL L

 (Exp.)

2016-06-05  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l		CRP 200			60 8P - 60 18P					± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l										± 0.15
HB HGB	g/dl		6.1			5.8					± 0.4
	g/l		61			58					± 4
	mmol/l		3.77			3.58					± 0.25
HT HCT	%		17.1			16.9					± 2.0
	l/l		0.171			0.169					± 0.020
VGM MCV	µm <sup>3</sup> ; fl		70			70					± 4
TGMH MCH	pg		24.9			23.8					± 2.0
	fmol		1.55			1.48					± 0.12
CCMH MCHC	g/dl		35.6			34.0					± 3.0
	g/l		356			340					± 30
	mmol/l		22.09			21.14					± 1.86
IDR RDW	%		14.0			14.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l		75			77					± 20
VPM MPV	µm <sup>3</sup> ; fl		9.0			8.2					± 2.0
LYMPHO	#		1.36			1.33					± 0.40
	%		64.6			63.5					± 8.0
MONO	#		0.23			0.21					± 0.20
	%		11.0			10.0					± 6.0
GRANULO	#		0.51			0.56					± 0.40
	%		24.5			26.5					± 7.0

LOT		MX 036 N		CONTROL N						(Exp.) <b>2016-06-05</b> (YYYY-MM-DD)	
Rev 1		Milysebio									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
			CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l		<b>7.5</b>			<b>7.3</b>					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l		<b>4.66</b>			<b>4.64</b>					± 0.18
HB HGB	g/dl		<b>13.4</b>			<b>13.2</b>					± 0.5
	g/l		<b>134</b>			<b>132</b>					± 5
	mmol/l		<b>8.32</b>			<b>8.22</b>					± 0.31
HT HCT	%		<b>37.3</b>			<b>37.6</b>					± 2.5
	l/l		<b>0.373</b>			<b>0.376</b>					± 0.025
VGM MCV	µm <sup>3</sup> ; fl		<b>80</b>			<b>81</b>					± 4
TGMH MCH	pg		<b>28.8</b>			<b>28.5</b>					± 2.0
	fmol		<b>1.79</b>			<b>1.77</b>					± 0.12
CCMH MCHC	g/dl		<b>35.9</b>			<b>35.2</b>					± 3.0
	g/l		<b>359</b>			<b>352</b>					± 30
	mmol/l		<b>22.32</b>			<b>21.86</b>					± 1.86
IDR RDW	%		<b>13.5</b>			<b>13.5</b>					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l		<b>250</b>			<b>255</b>					± 40
VPM MPV	µm <sup>3</sup> ; fl		<b>8.0</b>			<b>7.7</b>					± 2.0
LYMPHO	#		<b>2.46</b>			<b>2.37</b>					± 0.40
	%		<b>32.9</b>			<b>32.6</b>					± 5.0
MONO	#		<b>0.53</b>			<b>0.47</b>					± 0.30
	%		<b>7.1</b>			<b>6.5</b>					± 4.0
GRANULO	#		<b>4.48</b>			<b>4.44</b>					± 0.60
	%		<b>59.8</b>			<b>61.0</b>					± 6.0

1219930092-A  
 Ref: TEMP-0830 Rev.37



**LOT**

**MX 036 H**  
Rev 1

**CONTROL H**

(Exp.)

**2016-06-05**  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l		CRP 200			60 8P - 60 18P					± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l										± 0.20
HB HGB	g/dl		17.8			17.7					± 0.6
	g/l		178			177					± 6
	mmol/l		11.05			11.00					± 0.37
HT HCT	%		49.8			50.1					± 3.0
	l/l		0.498			0.501					± 0.030
VGM MCV	µm <sup>3</sup> ; fl		88			89					± 4
TGMH MCH	pg		31.4			31.5					± 2.0
	fmol		1.95			1.95					± 0.12
CCMH MCHC	g/dl		35.7			35.4					± 3.0
	g/l		357			354					± 30
	mmol/l		22.19			21.96					± 1.86
IDR RDW	%		13.0			13.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l		515			510					± 45
VPM MPV	µm <sup>3</sup> ; fl		7.8			7.5					± 2.0
LYMPHO	#		3.13			3.07					± 1.00
	%		15.9			16.0					± 5.0
MONO	#		1.06			0.86					± 0.80
	%		5.4			4.5					± 4.0
GRANULO	#		15.44			15.21					± 1.40
	%		78.4			79.2					± 7.0