



LOT MX 421 L
 Rev 1

CONTROL L



2020-04-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	1.9	1.8			1.9					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.40	2.35			2.38					± 0.15
HB HGB	g/dl	6.0	5.9			5.8					± 0.4
	g/l	60	59			58					± 4
	mmol/l	3.73	3.65			3.60					± 0.25
HT HCT	%	16.6	16.2			16.4					± 2.0
	l/l	0.166	0.162			0.164					± 0.020
VGM MCV	µm ³ ; fl	69	69			69					± 4
TGMH MCH	pg	25.0	25.0			24.4					± 2.0
	fmol	1.55	1.55			1.51					± 0.12
CCMH MCHC	g/dl	36.2	36.3			35.3					± 3.0
	g/l	362	363			353					± 30
	mmol/l	22.50	22.52			21.93					± 1.86
IDR RDW	%	14.5	14.0			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	73	73			76					± 20
VPM MPV	µm ³ ; fl	9.0	8.5			8.9					± 2.0
LYMPHO	#	1.25	1.18			1.25					± 0.40
	%	66.0	65.7			66.0					± 8.0
MONO	#	0.20	0.18			0.20					± 0.20
	%	8.5	10.0			8.0					± 6.0
GRANULO	#	0.48	0.44			0.49					± 0.40
	%	25.5	24.5			26.0					± 7.0

LOT MX 421 N
Rev 1

CONTROL N

 (Exp.)

2020-04-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.8	7.3			7.8					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.59	4.55			4.61					± 0.18
HB HGB	g/dl	13.5	13.4			13.5					± 0.5
	g/l	135	134			135					± 5
	mmol/l	8.38	8.32			8.38					± 0.31
HT HCT	%	37.2	36.9			37.3					± 2.5
	l/l	0.372	0.369			0.373					± 0.025
VGM MCV	µm ³ ; fl	81	81			81					± 4
TGMH MCH	pg	29.4	29.5			29.3					± 2.0
	fmol	1.83	1.83			1.82					± 0.12
CCMH MCHC	g/dl	36.3	36.4			36.2					± 3.0
	g/l	363	364			362					± 30
	mmol/l	22.55	22.58			22.45					± 1.86
IDR RDW	%	14.0	13.5			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	263	260			268					± 40
VPM MPV	µm ³ ; fl	8.4	8.0			8.5					± 2.0
LYMPHO	#	2.81	2.71			2.77					± 0.40
	%	36.0	37.1			35.5					± 5.0
MONO	#	0.47	0.45			0.47					± 0.30
	%	6.0	6.2			6.0					± 4.0
GRANULO	#	4.52	4.15			4.56					± 0.60
	%	58.0	56.8			58.5					± 6.0

Ref: TEMP-0830 Rev.39 FRONT / RECTO 130060626



LOT

MX 421 H

CONTROL

H

(Exp.)

2020-04-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.8	19.6			20.8					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.72	5.62			5.69					± 0.20
HB HGB	g/dl	18.4	17.8			18.4					± 0.6
	g/l	184	178			184					± 6
	mmol/l	11.43	11.05			11.43					± 0.37
HT HCT	%	50.3	49.5			50.1					± 3.0
	l/l	0.503	0.495			0.501					± 0.030
VGM MCV	µm ³ ; fl	88	88			88					± 4
TGMH MCH	pg	32.2	31.7			32.3					± 2.0
	fmol	2.00	1.97			2.01					± 0.12
CCMH MCHC	g/dl	36.6	36.0			36.7					± 3.0
	g/l	366	360			367					± 30
	mmol/l	22.70	22.35			22.82					± 1.86
IDR RDW	%	13.5	13.5			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	565	545			574					± 45
VPM MPV	µm ³ ; fl	8.3	8.0			8.5					± 2.0
LYMPHO	#	3.74	3.64			3.74					± 1.00
	%	18.0	18.6			18.0					± 5.0
MONO	#	0.83	1.27			0.83					± 0.80
	%	4.0	6.5			4.0					± 4.0
GRANULO	#	16.22	14.69			16.22					± 1.40
	%	78.0	75.0			78.0					± 7.0