



LOT

MX 056 L

CONTROL

L

(Exp.)

2016-08-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l		CRP 200			60 8P - 60 18P					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l										± 0.15
HB HGB	g/dl		6.0			5.7					± 0.4
	g/l		60			57					± 4
	mmol/l		3.71			3.52					± 0.25
HT HCT	%		16.5			16.5					± 2.0
	l/l		0.165			0.165					± 0.020
VGM MCV	µm ³ ; fl		69			69					± 4
TGMH MCH	pg		25.0			23.7					± 2.0
	fmol		1.55			1.47					± 0.12
CCMH MCHC	g/dl		36.3			34.4					± 3.0
	g/l		363			344					± 30
	mmol/l		22.51			21.36					± 1.86
IDR RDW	%		14.5			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l		71			73					± 20
VPM MPV	µm ³ ; fl		9.0			8.0					± 2.0
LYMPHO	#		1.28			1.25					± 0.40
	%		64.1			62.5					± 8.0
MONO	#		0.21			0.22					± 0.20
	%		10.5			11.0					± 6.0
GRANULO	#		0.51			0.53					± 0.40
	%		25.5			26.5					± 7.0

LOT

MX 056 N
Rev 1

CONTROL N

(Exp.)

2016-08-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l		CRP 200			60 8P - 60 18P					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l										± 0.18
HB HGB	g/dl		7.8			7.8					± 0.5
	g/l		13.1			12.9					± 5
	mmol/l		8.14			8.03					± 0.31
HT HCT	%		36.7			36.8					± 2.5
	l/l		0.367			0.368					± 0.025
VGM MCV	µm ³ ; fl		80			80					± 4
TGMH MCH	pg		28.5			28.1					± 2.0
	fmol		1.77			1.75					± 0.12
CCMH MCHC	g/dl		35.7			35.2					± 3.0
	g/l		357			352					± 30
	mmol/l		22.15			21.83					± 1.86
IDR RDW	%		13.8			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l		258			255					± 40
VPM MPV	µm ³ ; fl		8.5			7.8					± 2.0
LYMPHO	#		2.60			2.53					± 0.40
	%		33.4			32.6					± 5.0
MONO	#		0.59			0.61					± 0.30
	%		7.6			7.9					± 4.0
GRANULO	#		4.59			4.62					± 0.60
	%		58.8			59.5					± 6.0

1219930092-A
Ref: TEMP-0830 Rev.37



LOT

MX 056 H
Rev 1

CONTROL H

(Exp.)

2016-08-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l		CRP 200			60 8P - 60 18P					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l										± 0.20
HB HGB	g/dl		18.1			18.0					± 0.6
	g/l		181			180					± 6
	mmol/l		11.24			11.19					± 0.37
HT HCT	%		51.6			51.6					± 3.0
	l/l		0.516			0.516					± 0.030
VGM MCV	µm ³ ; fl		89			89					± 4
TGMH MCH	pg		31.2			31.1					± 2.0
	fmol		1.94			1.93					± 0.12
CCMH MCHC	g/dl		35.1			34.9					± 3.0
	g/l		351			349					± 30
	mmol/l		21.77			21.68					± 1.86
IDR RDW	%		13.2			13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l		523			510					± 45
VPM MPV	µm ³ ; fl		8.0			7.4					± 2.0
LYMPHO	#		3.25			3.23					± 1.00
	%		16.4			16.5					± 5.0
MONO	#		1.28			1.08					± 0.80
	%		6.5			5.5					± 4.0
GRANULO	#		15.23			15.23					± 1.40
	%		76.9			77.7					± 7.0