

LOT MX 408 L
Rev 1

CONTROL L

 (Exp.)

2018-02-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0	2.0			2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.35	2.31			2.32					± 0.15
HB HGB	g/dl	5.9	5.9			5.9					± 0.4
	g/l	59	59			59					± 4
	mmol/l	3.66	3.65			3.66					± 0.25
HT HCT	%	16.5	16.2			16.2					± 2.0
	l/l	0.165	0.162			0.162					± 0.020
VGM MCV	µm ³ ; fl	70	70			70					± 4
TGMH MCH	pg	25.1	25.5			25.4					± 2.0
	fmol	1.56	1.58			1.58					± 0.12
CCMH MCHC	g/dl	35.9	36.4			36.3					± 3.0
	g/l	359	364			363					± 30
	mmol/l	22.27	22.58			22.56					± 1.86
IDR RDW	%	14.0	13.5			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	76	73			77					± 20
VPM MPV	µm ³ ; fl	8.4	8.2			8.2					± 2.0
LYMPHO	#	1.31	1.30			1.30					± 0.40
	%	65.5	65.1			65.0					± 8.0
MONO	#	0.20	0.19			0.20					± 0.20
	%	8.5	9.5			8.0					± 6.0
GRANULO	#	0.52	0.51			0.54					± 0.40
	%	26.0	25.5			27.0					± 7.0

LOT

MX 408 N
Rev 1

CONTROL N

(Exp.)

2018-02-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.5	7.6			7.6					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.63	4.59			4.58					± 0.18
HB HGB	g/dl	13.3	13.2			13.3					± 0.5
	g/l	133	132			133					± 5
	mmol/l	8.26	8.20			8.26					± 0.31
HT HCT	%	37.0	36.7			36.6					± 2.5
	l/l	0.370	0.367			0.366					± 0.025
VGM MCV	µm ³ ; fl	80	80			80					± 4
TGMH MCH	pg	28.7	28.8			29.0					± 2.0
	fmol	1.78	1.79			1.80					± 0.12
CCMH MCHC	g/dl	35.9	35.9			36.3					± 3.0
	g/l	359	359			363					± 30
	mmol/l	22.30	22.32			22.54					± 1.86
IDR RDW	%	13.5	13.5			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	252	253			252					± 40
VPM MPV	µm ³ ; fl	8.6	8.3			8.3					± 2.0
LYMPHO	#	2.59	2.74			2.66					± 0.40
	%	34.5	36.0			35.0					± 5.0
MONO	#	0.49	0.47			0.46					± 0.30
	%	6.5	6.2			6.0					± 4.0
GRANULO	#	4.43	4.39			4.48					± 0.60
	%	59.0	57.8			59.0					± 6.0

1300027266-A
Ref: TEMP-0830 Rev.38

LOT MX 408 H
Rev 1

CONTROL H

 (Exp.)

2018-02-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	19.5	19.9			19.9					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.71	5.62			5.62					± 0.20
HB HGB	g/dl	18.1	17.6			18.0					± 0.6
	g/l	181	176			180					± 6
	mmol/l	11.24	10.93			11.18					± 0.37
HT HCT	%	50.8	49.5			49.5					± 3.0
	l/l	0.508	0.495			0.495					± 0.030
VGM MCV	µm ³ ; fl	89	88			88					± 4
TGMH MCH	pg	31.7	31.3			32.0					± 2.0
	fmol	1.97	1.94			1.99					± 0.12
CCMH MCHC	g/dl	35.6	35.6			36.4					± 3.0
	g/l	356	356			364					± 30
	mmol/l	22.12	22.10			22.60					± 1.86
IDR RDW	%	13.5	13.0			13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	500	490			500					± 45
VPM MPV	µm ³ ; fl	8.3	7.9			8.1					± 2.0
LYMPHO	#	2.93	3.06			3.08					± 1.00
	%	15.0	15.4			15.5					± 5.0
MONO	#	0.98	1.18			1.00					± 0.80
	%	5.0	5.9			5.0					± 4.0
GRANULO	#	15.60	15.60			15.82					± 1.40
	%	80.0	78.4			79.5					± 7.0