



LOT

MX 419 L
 Rev 1

CONTROL L



2019-12-05
 (YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.1	2.0			2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.32	2.26			2.29					± 0.15
HB HGB	g/dl	5.8	5.8			5.7					± 0.4
	g/l	58	58			57					± 4
	mmol/l	3.60	3.59			3.54					± 0.25
HT HCT	%	16.0	15.8			15.8					± 2.0
	l/l	0.160	0.158			0.158					± 0.020
VGM MCV	µm ³ ; fl	69	70			69					± 4
TGMH MCH	pg	25.0	25.6			24.9					± 2.0
	fmol	1.55	1.59			1.55					± 0.12
CCMH MCHC	g/dl	36.2	36.5			36.1					± 3.0
	g/l	362	365			361					± 30
	mmol/l	22.50	22.70			22.40					± 1.86
IDR RDW	%	14.0	14.0			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	80	81			80					± 20
VPM MPV	µm ³ ; fl	8.8	8.2			8.8					± 2.0
LYMPHO	#	1.39	1.34			1.31					± 0.40
	%	66.0	67.2			65.5					± 8.0
MONO	#	0.20	0.20			0.20					± 0.20
	%	8.5	10.0			8.5					± 6.0
GRANULO	#	0.54	0.46			0.52					± 0.40
	%	25.5	23.0			26.0					± 7.0

LOT MX 419 N
 Rev 1

CONTROL N

(Exp.)

2019-12-05
 (YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.6	7.0			7.5					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.65	4.55			4.63					± 0.18
HB HGB	g/dl	13.5	13.2			13.3					± 0.5
	g/l	135	132			133					± 5
	mmol/l	8.38	8.20			8.26					± 0.31
HT HCT	%	36.7	35.9			36.6					± 2.5
	l/l	0.367	0.359			0.366					± 0.025
VGM MCV	µm ³ ; fl	79	79			79					± 4
TGMH MCH	pg	29.0	29.0			28.7					± 2.0
	fmol	1.80	1.80			1.78					± 0.12
CCMH MCHC	g/dl	36.7	36.7			36.4					± 3.0
	g/l	367	367			364					± 30
	mmol/l	22.82	22.80			22.58					± 1.86
IDR RDW	%	14.0	13.5			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	263	260			267					± 40
VPM MPV	µm ³ ; fl	7.9	7.5			7.8					± 2.0
LYMPHO	#	2.62	2.49			2.51					± 0.40
	%	34.5	35.5			33.5					± 5.0
MONO	#	0.49	0.50			0.45					± 0.30
	%	6.5	7.1			6.0					± 4.0
GRANULO	#	4.48	4.01			4.54					± 0.60
	%	59.0	57.3			60.5					± 6.0

Ref: TEMP-0830 Rev.39 FRONT / RECTO 1300060626

LOT MX 419 H
Rev 1

CONTROL H

 (Exp.)

2019-12-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.1	19.0			20.2					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.75	5.62			5.71					± 0.20
HB HGB	g/dl	18.3	17.6			18.2					± 0.6
	g/l	183	176			182					± 6
	mmol/l	11.36	10.93			11.30					± 0.37
HT HCT	%	50.6	49.5			49.7					± 3.0
	l/l	0.506	0.495			0.497					± 0.030
VGM MCV	µm ³ ; fl	88	88			87					± 4
TGMH MCH	pg	31.8	31.3			31.9					± 2.0
	fmol	1.98	1.94			1.98					± 0.12
CCMH MCHC	g/dl	36.2	35.6			36.6					± 3.0
	g/l	362	356			366					± 30
	mmol/l	22.46	22.10			22.75					± 1.86
IDR RDW	%	13.5	13.5			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	510	500			515					± 45
VPM MPV	µm ³ ; fl	8.2	7.8			8.2					± 2.0
LYMPHO	#	3.52	3.22			3.33					± 1.00
	%	17.5	17.0			16.5					± 5.0
MONO	#	0.80	1.23			0.81					± 0.80
	%	4.0	6.5			4.0					± 4.0
GRANULO	#	15.78	14.52			16.06					± 1.40
	%	78.5	76.4			79.5					± 7.0