

LOT

MX 400 L
Rev 1

CONTROL L

(Exp.)

2016-10-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l		CRP 200			60 8P - 60 18P					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l										± 0.15
HB HGB	g/dl		6.0			5.7					± 0.4
	g/l		60			57					± 4
	mmol/l		3.71			3.52					± 0.25
HT HCT	%		16.4			16.5					± 2.0
	l/l		0.164			0.165					± 0.020
VGM MCV	µm ³ ; fl		69			70					± 4
TGMH MCH	pg		25.1			24.0					± 2.0
	fmol		1.56			1.49					± 0.12
CCMH MCHC	g/dl		36.4			34.3					± 3.0
	g/l		364			343					± 30
	mmol/l		22.61			21.33					± 1.86
IDR RDW	%		14.0			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l		78			78					± 20
VPM MPV	µm ³ ; fl		8.5			8.2					± 2.0
LYMPHO	#		1.29			1.26					± 0.40
	%		64.6			63.0					± 8.0
MONO	#		0.21			0.22					± 0.20
	%		10.5			11.0					± 6.0
GRANULO	#		0.50			0.52					± 0.40
	%		25.0			26.0					± 7.0



HORIBA ABX SAS

HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée - BP 7290
 34184 MONTPELLIER Cedex 4
 FRANCE
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

ABX Minotrol 16



LOT

MX 400 N

CONTROL N

(Exp.)

2016-10-05
 (YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l		CRP 200			60 8P - 60 18P					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l										± 0.18
HB HGB	g/dl		7.4			7.3					± 0.5
	g/l		13.6			13.3					± 5
	mmol/l		8.45			8.28					± 0.31
HT HCT	%		37.4			37.6					± 2.5
	l/l		0.374			0.376					± 0.025
VGM MCV	µm ³ ; fl		81			81					± 4
TGMH MCH	pg		29.4			28.7					± 2.0
	fmol		1.83			1.78					± 0.12
CCMH MCHC	g/dl		36.3			35.5					± 3.0
	g/l		363			355					± 30
	mmol/l		22.57			22.02					± 1.86
IDR RDW	%		13.5			13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l		253			250					± 40
VPM MPV	µm ³ ; fl		8.0			7.8					± 2.0
LYMPHO	#		2.75			2.63					± 0.40
	%		37.1			36.2					± 5.0
MONO	#		0.49			0.51					± 0.30
	%		6.7			7.0					± 4.0
GRANULO	#		4.17			4.15					± 0.60
	%		56.4			57.0					± 6.0

1219930092-A
 Ref: TEMP-0830 Rev.37



LOT

MX 400 H

CONTROL

H

(Exp.)

2016-10-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l		CRP 200			60 8P - 60 18P					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l										± 0.20
HB HGB	g/dl		18.0			17.9					± 0.6
	g/l		180			179					± 6
	mmol/l		11.18			11.13					± 0.37
HT HCT	%		49.9			51.0					± 3.0
	l/l		0.499			0.510					± 0.030
VGM MCV	µm ³ ; fl		87			88					± 4
TGMH MCH	pg		31.4			30.9					± 2.0
	fmol		1.95			1.92					± 0.12
CCMH MCHC	g/dl		36.0			35.1					± 3.0
	g/l		360			351					± 30
	mmol/l		22.38			21.80					± 1.86
IDR RDW	%		13.0			12.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l		518			513					± 45
VPM MPV	µm ³ ; fl		8.0			7.8					± 2.0
LYMPHO	#		3.25			3.20					± 1.00
	%		16.4			16.5					± 5.0
MONO	#		1.28			1.07					± 0.80
	%		6.5			5.5					± 4.0
GRANULO	#		15.23			15.08					± 1.40
	%		76.9			77.7					± 7.0