

LOT MX 410 L
 Rev 1

CONTROL L

(Exp.)

2018-06-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0	2.0			2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.29	2.25			2.25					± 0.15
HB HGB	g/dl	5.9	5.9			5.8					± 0.4
	g/l	59	59			58					± 4
	mmol/l	3.66	3.65			3.60					± 0.25
HT HCT	%	16.3	16.0			16.0					± 2.0
	l/l	0.163	0.160			0.160					± 0.020
VGM MCV	µm ³ ; fl	71	71			71					± 4
TGMH MCH	pg	25.8	26.1			25.8					± 2.0
	fmol	1.60	1.62			1.60					± 0.12
CCMH MCHC	g/dl	36.3	36.8			36.3					± 3.0
	g/l	363	368			363					± 30
	mmol/l	22.53	22.86			22.55					± 1.86
IDR RDW	%	14.0	13.5			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	82	78			85					± 20
VPM MPV	µm ³ ; fl	8.8	8.5			8.6					± 2.0
LYMPHO	#	1.35	1.33			1.35					± 0.40
	%	67.5	66.7			67.5					± 8.0
MONO	#	0.20	0.18			0.20					± 0.20
	%	7.5	9.0			7.0					± 6.0
GRANULO	#	0.50	0.49			0.51					± 0.40
	%	25.0	24.5			25.5					± 7.0

LOT		MX 410 N	CONTROL N								(Exp.)	2018-06-05 (YYYY-MM-DD)
Rev 1		Minilysebio										
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE	
		ES60	CRP 200			60 8P - 60 18P						
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.5	7.7			7.7					± 0.8	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.60	4.55			4.55					± 0.18	
HB HGB	g/dl	13.5	13.4			13.4					± 0.5	
	g/l	135	134			134					± 5	
	mmol/l	8.38	8.32			8.32					± 0.31	
HT HCT	%	37.3	36.9			36.9					± 2.5	
	l/l	0.373	0.369			0.369					± 0.025	
VGM MCV	µm ³ ; fl	81	81			81					± 4	
TGMH MCH	pg	29.3	29.5			29.5					± 2.0	
	fmol	1.82	1.83			1.83					± 0.12	
CCMH MCHC	g/dl	36.2	36.4			36.4					± 3.0	
	g/l	362	364			364					± 30	
	mmol/l	22.50	22.58			22.58					± 1.86	
IDR RDW	%	14.0	13.5			14.0					± 3.0	
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	257	256			260					± 40	
VPM MPV	µm ³ ; fl	8.5	8.2			8.3					± 2.0	
LYMPHO	#	2.66	2.82			2.77					± 0.40	
	%	35.5	36.6			36.0					± 5.0	
MONO	#	0.49	0.55			0.46					± 0.30	
	%	6.5	7.1			6.0					± 4.0	
GRANULO	#	4.35	4.34			4.47					± 0.60	
	%	58.0	56.4			58.0					± 6.0	

1300027266-A
Ref: TEMP-0830 Rev.36

LOT MX 410 H
Rev 1

CONTROL H

 (Exp.)

2018-06-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	19.8	20.3			20.5					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.68	5.63			5.65					± 0.20
HB HGB	g/dl	18.3	17.9			18.2					± 0.6
	g/l	183	179			182					± 6
	mmol/l	11.36	11.12			11.30					± 0.37
HT HCT	%	50.6	49.5			49.7					± 3.0
	l/l	0.506	0.495			0.497					± 0.030
VGM MCV	µm ³ ; fl	89	88			88					± 4
TGMH MCH	pg	32.2	31.8			32.2					± 2.0
	fmol	2.00	1.97			2.00					± 0.12
CCMH MCHC	g/dl	36.2	36.1			36.6					± 3.0
	g/l	362	361			366					± 30
	mmol/l	22.48	22.44			22.73					± 1.86
IDR RDW	%	13.0	13.0			13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	498	495			493					± 45
VPM MPV	µm ³ ; fl	8.6	8.3			8.5					± 2.0
LYMPHO	#	3.27	3.34			3.49					± 1.00
	%	16.5	16.4			17.0					± 5.0
MONO	#	0.79	0.99			0.82					± 0.80
	%	4.0	4.9			4.0					± 4.0
GRANULO	#	15.74	15.92			16.20					± 1.40
	%	79.5	78.4			79.0					± 7.0