



**LOT** MX 401 L  
 Rev 1

**CONTROL** L



**2016-12-05**  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l		CRP 200			60 8P - 60 18P					± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l										± 0.15
HB HGB	g/dl		5.9			5.6					± 0.4
	g/l		59			56					± 4
	mmol/l		3.65			3.47					± 0.25
HT HCT	%		16.0			16.1					± 2.0
	l/l		0.160			0.161					± 0.020
VGM MCV	µm <sup>3</sup> ; fl		69			70					± 4
TGMH MCH	pg		25.3			24.3					± 2.0
	fmol		1.57			1.51					± 0.12
CCMH MCHC	g/dl		36.7			34.7					± 3.0
	g/l		367			347					± 30
	mmol/l		22.81			21.52					± 1.86
IDR RDW	%		13.9			13.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l		74			75					± 20
VPM MPV	µm <sup>3</sup> ; fl		8.9			8.0					± 2.0
LYMPHO	#		1.29			1.20					± 0.40
	%		64.6			63.0					± 8.0
MONO	#		0.21			0.20					± 0.20
	%		10.5			10.5					± 6.0
GRANULO	#		0.50			0.50					± 0.40
	%		25.0			26.5					± 7.0

<b>LOT</b>		<b>MX 401 N</b>	<b>CONTROL N</b>				(Exp.)			<b>2016-12-05</b> (YYYY-MM-DD)	
Rev 1		<b>Minilysebio</b>									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
			CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l		<b>7.7</b>			<b>7.5</b>					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l		<b>4.50</b>			<b>4.50</b>					± 0.18
HB HGB	g/dl		<b>13.1</b>			<b>12.8</b>					± 0.5
	g/l		<b>131</b>			<b>128</b>					± 5
	mmol/l		<b>8.14</b>			<b>7.97</b>					± 0.31
HT HCT	%		<b>36.0</b>			<b>36.0</b>					± 2.5
	l/l		<b>0.360</b>			<b>0.360</b>					± 0.025
VGM MCV	µm <sup>3</sup> ; fl		<b>80</b>			<b>80</b>					± 4
TGMH MCH	pg		<b>29.1</b>			<b>28.5</b>					± 2.0
	fmol		<b>1.81</b>			<b>1.77</b>					± 0.12
CCMH MCHC	g/dl		<b>36.4</b>			<b>35.7</b>					± 3.0
	g/l		<b>364</b>			<b>357</b>					± 30
	mmol/l		<b>22.60</b>			<b>22.15</b>					± 1.86
IDR RDW	%		<b>13.2</b>			<b>12.5</b>					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l		<b>253</b>			<b>254</b>					± 40
VPM MPV	µm <sup>3</sup> ; fl		<b>8.9</b>			<b>8.2</b>					± 2.0
LYMPHO	#		<b>2.69</b>			<b>2.51</b>					± 0.40
	%		<b>35.0</b>			<b>33.7</b>					± 5.0
MONO	#		<b>0.58</b>			<b>0.56</b>					± 0.30
	%		<b>7.6</b>			<b>7.4</b>					± 4.0
GRANULO	#		<b>4.41</b>			<b>4.41</b>					± 0.60
	%		<b>57.3</b>			<b>59.0</b>					± 6.0

1219930092-A  
Ref: TEMP-0830 Rev.37



**LOT**

**MX 401 H**  
Rev 1

**CONTROL H**

(Exp.)

**2016-12-05**  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l		CRP 200			60 8P - 60 18P					± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l										± 0.20
HB HGB	g/dl		17.8			17.6					± 0.6
	g/l		178			176					± 6
	mmol/l		11.05			10.94					± 0.37
HT HCT	%		49.9			49.8					± 3.0
	l/l		0.499			0.498					± 0.030
VGM MCV	µm <sup>3</sup> ; fl		88			88					± 4
TGMH MCH	pg		31.4			31.1					± 2.0
	fmol		1.95			1.93					± 0.12
CCMH MCHC	g/dl		35.7			35.4					± 3.0
	g/l		357			354					± 30
	mmol/l		22.15			21.97					± 1.86
IDR RDW	%		12.7			12.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l		500			500					± 45
VPM MPV	µm <sup>3</sup> ; fl		8.1			7.7					± 2.0
LYMPHO	#		3.13			2.94					± 1.00
	%		15.9			15.5					± 5.0
MONO	#		1.17			0.95					± 0.80
	%		5.9			5.0					± 4.0
GRANULO	#		15.35			15.06					± 1.40
	%		77.9			79.2					± 7.0