

**LOT** MX 411 L  
 Rev 1

**CONTROL** L

(Exp.)

**2018-08-05**  
 (YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.0	2.0			2.0					± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.25	2.25			2.23					± 0.15
HB HGB	g/dl	5.7	5.8			5.7					± 0.4
	g/l	57	58			57					± 4
	mmol/l	3.54	3.59			3.54					± 0.25
HT HCT	%	16.0	16.0			15.8					± 2.0
	l/l	0.160	0.160			0.158					± 0.020
VGM MCV	µm <sup>3</sup> ; fl	71	71			71					± 4
TGMH MCH	pg	25.3	25.7			25.6					± 2.0
	fmol	1.57	1.60			1.59					± 0.12
CCMH MCHC	g/dl	35.7	36.2			36.0					± 3.0
	g/l	357	362			360					± 30
	mmol/l	22.16	22.48			22.36					± 1.86
IDR RDW	%	14.0	13.5			14.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	83	76			81					± 20
VPM MPV	µm <sup>3</sup> ; fl	9.2	8.9			9.2					± 2.0
LYMPHO	#	1.34	1.33			1.35					± 0.40
	%	67.0	66.7			67.5					± 8.0
MONO	#	0.20	0.18			0.20					± 0.20
	%	8.0	9.0			7.5					± 6.0
GRANULO	#	0.50	0.49			0.50					± 0.40
	%	25.0	24.5			25.0					± 7.0

**LOT**

**MX 411 N**  
 Rev 1

**CONTROL N**

(Exp.)

**2018-08-05**  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.3	7.5			7.7					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.64	4.62			4.60					± 0.18
HB HGB	g/dl	13.4	13.4			13.4					± 0.5
	g/l	134	134			134					± 5
	mmol/l	8.32	8.32			8.32					± 0.31
HT HCT	%	37.1	37.4			37.3					± 2.5
	l/l	0.371	0.374			0.373					± 0.025
VGM MCV	µm <sup>3</sup> ; fl	80	81			81					± 4
TGMH MCH	pg	28.9	29.0			29.1					± 2.0
	fmol	1.79	1.80			1.81					± 0.12
CCMH MCHC	g/dl	36.1	35.8			36.0					± 3.0
	g/l	361	358			360					± 30
	mmol/l	22.42	22.24			22.33					± 1.86
IDR RDW	%	14.0	13.5			14.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	267	264			266					± 40
VPM MPV	µm <sup>3</sup> ; fl	8.9	8.4			8.8					± 2.0
LYMPHO	#	2.52	2.66			2.70					± 0.40
	%	34.5	35.5			35.0					± 5.0
MONO	#	0.47	0.50			0.50					± 0.30
	%	6.5	6.7			6.5					± 4.0
GRANULO	#	4.31	4.34			4.50					± 0.60
	%	59.0	57.8			58.5					± 6.0

1300027266-A  
 Ref: TEMP-0830 Rev.38



**LOT** MX 411 H  
 Rev 1

**CONTROL** H



**2018-08-05**  
 (YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	19.7	20.2			20.6					± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	5.80	5.78			5.78					± 0.20
HB HGB	g/dl	18.5	18.1			18.5					± 0.6
	g/l	185	181			185					± 6
	mmol/l	11.49	11.24			11.49					± 0.37
HT HCT	%	51.0	50.9			50.9					± 3.0
	l/l	0.510	0.509			0.509					± 0.030
VGM MCV	µm <sup>3</sup> ; fl	88	88			88					± 4
TGMH MCH	pg	31.9	31.3			32.0					± 2.0
	fmol	1.98	1.94			1.99					± 0.12
CCMH MCHC	g/dl	36.2	35.6			36.4					± 3.0
	g/l	362	356			364					± 30
	mmol/l	22.51	22.10			22.59					± 1.86
IDR RDW	%	13.5	13.0			13.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	496	495			490					± 45
VPM MPV	µm <sup>3</sup> ; fl	9.1	8.6			9.0					± 2.0
LYMPHO	#	3.35	3.32			3.61					± 1.00
	%	17.0	16.4			17.5					± 5.0
MONO	#	0.89	1.20			0.93					± 0.80
	%	4.5	5.9			4.5					± 4.0
GRANULO	#	15.46	15.64			16.07					± 1.40
	%	78.5	77.4			78.0					± 7.0