



LOT

MX 430 L
Rev 1

CONTROL L

(Exp.)

2021-10-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0	2.0			2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.36	2.30			2.33					± 0.15
HB HGB	g/dl	5.9	5.9			5.6					± 0.4
	g/l	59	59			56					± 4
	mmol/l	3.66	3.65			3.47					± 0.25
HT HCT	%	16.5	16.1			16.5					± 2.0
	l/l	0.165	0.161			0.165					± 0.020
VGM MCV	µm ³ ; fl	70	70			71					± 4
TGMH MCH	pg	25.0	25.6			23.9					± 2.0
	fmol	1.55	1.59			1.49					± 0.12
CCMH MCHC	g/dl	35.7	36.5			33.7					± 3.0
	g/l	357	365			337					± 30
	mmol/l	22.18	22.68			20.95					± 1.86
IDR RDW	%	16.0	14.0			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	74	75			76					± 20
VPM MPV	µm ³ ; fl	9.0	9.3			9.7					± 2.0
LYMPHO	#	1.37	1.42			1.36					± 0.40
	%	68.5	71.2			68.0					± 8.0
MONO	#	0.20	0.15			0.20					± 0.20
	%	6.5	7.5			6.0					± 6.0
GRANULO	#	0.50	0.43			0.52					± 0.40
	%	25.0	21.6			26.0					± 7.0

LOT MX 430 N
 Rev 1

CONTROL N

(Exp.)

2021-10-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.9	7.8			7.8					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.60	4.55			4.60					± 0.18
HB HGB	g/dl	13.6	13.5			13.3					± 0.5
	g/l	136	135			133					± 5
	mmol/l	8.45	8.38			8.28					± 0.31
HT HCT	%	37.3	37.3			37.3					± 2.5
	l/l	0.373	0.373			0.373					± 0.025
VGM MCV	µm ³ ; fl	81	82			81					± 4
TGMH MCH	pg	29.6	29.7			29.0					± 2.0
	fmol	1.84	1.84			1.80					± 0.12
CCMH MCHC	g/dl	36.5	36.2			35.8					± 3.0
	g/l	365	362			358					± 30
	mmol/l	22.67	22.47			22.21					± 1.86
IDR RDW	%	15.0	14.0			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	240	253			260					± 40
VPM MPV	µm ³ ; fl	8.5	8.3			9.0					± 2.0
LYMPHO	#	3.00	3.14			2.93					± 0.40
	%	38.0	40.3			37.5					± 5.0
MONO	#	0.43	0.49			0.39					± 0.30
	%	5.5	6.2			5.0					± 4.0
GRANULO	#	4.46	4.20			4.49					± 0.60
	%	56.5	53.9			57.5					± 6.0

Ref: TEMP-0830 Rev.41 FRONT / RECTO 130060626



HORIBA ABX SAS

HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée - BP 7290
 34184 MONTPELLIER Cedex 4
 FRANCE
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

ABX Minotrol 16



LOT

MX 430 H

CONTROL H



2021-10-05
 (YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.8	20.3			20.4					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.62	5.55			5.60					± 0.20
HB HGB	g/dl	18.2	17.7			17.8					± 0.6
	g/l	182	177			178					± 6
	mmol/l	11.30	10.99			11.07					± 0.37
HT HCT	%	50.0	49.4			49.8					± 3.0
	l/l	0.500	0.494			0.498					± 0.030
VGM MCV	µm ³ ; fl	89	89			89					± 4
TGMH MCH	pg	32.4	31.9			31.8					± 2.0
	fmol	2.01	1.98			1.98					± 0.12
CCMH MCHC	g/dl	36.4	35.8			35.8					± 3.0
	g/l	364	358			358					± 30
	mmol/l	22.60	22.25			22.20					± 1.86
IDR RDW	%	14.0	13.5			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	472	490			500					± 45
VPM MPV	µm ³ ; fl	8.5	8.0			8.6					± 2.0
LYMPHO	#	3.64	3.66			3.47					± 1.00
	%	17.5	18.0			17.0					± 5.0
MONO	#	0.83	1.43			0.82					± 0.80
	%	4.0	7.0			4.0					± 4.0
GRANULO	#	16.33	15.22			16.12					± 1.40
	%	78.5	75.0			79.0					± 7.0