

LOT MX 432 L
 Rev 1

CONTROL L

(Exp.)

2022-02-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0	2.0			2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.32	2.31			2.31					± 0.17
HB HGB	g/dl	6.1	6.0			6.1					± 0.5
	g/l	61	60			61					± 5
	mmol/l	3.79	3.73			3.79					± 0.31
HT HCT	%	16.5	16.2			16.4					± 2.0
	l/l	0.165	0.162			0.164					± 0.020
VGM MCV	µm ³ ; fl	71	70			71					± 4
TGMH MCH	pg	26.3	26.0			26.4					± 2.0
	fmol	1.63	1.61			1.64					± 0.12
CCMH MCHC	g/dl	37.0	37.1			37.2					± 3.0
	g/l	370	371			372					± 30
	mmol/l	23.00	23.04			23.10					± 1.86
IDR RDW	%	15.0	14.0			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	80	80			78					± 22
VPM MPV	µm ³ ; fl	8.3	8.3			7.6					± 2.0
LYMPHO	#	1.25	1.28			1.27					± 0.40
	%	62.5	64.0			63.5					± 8.0
MONO	#	0.22	0.26			0.22					± 0.20
	%	11.0	13.0			11.0					± 6.0
GRANULO	#	0.53	0.46			0.51					± 0.40
	%	26.5	23.0			25.5					± 7.0

LOT

MX 432 N
Rev 1


CONTROL N

(Exp.)

2022-02-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.5	7.3			7.5					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.54	4.53			4.54					± 0.26
HB HGB	g/dl	13.4	13.4			13.5					± 0.7
	g/l	134	134			135					± 7
	mmol/l	8.32	8.32			8.38					± 0.43
HT HCT	%	37.2	37.1			37.2					± 2.7
	l/l	0.372	0.371			0.372					± 0.027
VGM MCV	µm ³ ; fl	82	82			82					± 4
TGMH MCH	pg	29.5	29.6			29.7					± 2.0
	fmol	1.83	1.84			1.85					± 0.12
CCMH MCHC	g/dl	36.0	36.1			36.3					± 3.0
	g/l	360	361			363					± 30
	mmol/l	22.35	22.40			22.52					± 1.86
IDR RDW	%	13.5	13.5			13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	255	250			250					± 40
VPM MPV	µm ³ ; fl	7.6	7.7			7.4					± 2.0
LYMPHO	#	2.66	2.63			2.63					± 0.40
	%	35.5	36.0			35.0					± 5.0
MONO	#	0.41	0.51			0.38					± 0.30
	%	5.5	7.0			5.0					± 4.0
GRANULO	#	4.43	4.16			4.50					± 0.60
	%	59.0	57.0			60.0					± 6.0

Ref: TEMP-0830 Rev.41 FRONT / RECTO 130090307

LOT		MX 432 H	CONTROL H				 (Exp.)	2022-02-05 (YYYY - MM - DD)				
Rev 1		ABX Lyse										
PARAMETRES PARAMETERS	UNITES UNITS	MICROS		MICROS		MICROS		ARGOS - HELIOS		PENTRA		TOLERANCES TOLERANCE
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60						
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.5	19.7			20.2						± 1.7
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.66	5.63			5.65						± 0.32
HB HGB	g/dl	18.1	18.1			18.2						± 0.9
	g/l	181	181			182						± 9
	mmol/l	11.24	11.24			11.30						± 0.56
HT HCT	%	50.9	50.1			50.9						± 3.4
	l/l	0.509	0.501			0.509						± 0.034
VGM MCV	µm ³ ; fl	90	89			90						± 4
TGMH MCH	pg	32.0	32.1			32.2						± 2.0
	fmol	1.99	2.00			2.00						± 0.12
CCMH MCHC	g/dl	35.5	36.1			35.8						± 3.0
	g/l	355	361			358						± 30
	mmol/l	22.07	22.43			22.23						± 1.86
IDR RDW	%	15.0	14.0			13.5						± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	502	485			495						± 65
VPM MPV	µm ³ ; fl	7.7	7.7			7.3						± 2.0
LYMPHO	#	3.08	2.76			2.83						± 1.00
	%	15.0	14.0			14.0						± 5.0
MONO	#	1.23	1.38			1.21						± 0.80
	%	6.0	7.0			6.0						± 4.0
GRANULO	#	16.20	15.56			16.16						± 1.40
	%	79.0	79.0			80.0						± 7.0