



LOT MX 419 L
 Rev 1

CONTROL L

(Exp.)

2019-12-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.1	2.0	2.1	2.1	2.1					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.26	2.26	2.30	2.29	2.28					± 0.15
HB HGB	g/dl	5.9	5.9	5.9	6.0	6.0					± 0.4
	g/l	59	59	59	60	60					± 4
	mmol/l	3.66	3.66	3.66	3.73	3.73					± 0.25
HT HCT	%	15.8	15.8	16.1	16.0	16.0					± 2.0
	l/l	0.158	0.158	0.161	0.160	0.160					± 0.020
VGM MCV	µm ³ ; fl	70	70	70	70	70					± 4
TGMH MCH	pg	26.1	26.1	25.7	26.2	26.3					± 2.0
	fmol	1.62	1.62	1.59	1.63	1.63					± 0.12
CCMH MCHC	g/dl	37.3	37.3	36.6	37.4	37.6					± 3.0
	g/l	373	373	366	374	376					± 30
	mmol/l	23.16	23.16	22.76	23.24	23.35					± 1.86
IDR RDW	%	14.0	14.0	N/A	14.0	13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	81	81	70	78	78					± 20
VPM MPV	µm ³ ; fl	8.5	8.2	N/A	8.6	7.7					± 2.0
LYMPHO	#	1.37	1.33	N/A	1.00	1.34					± 0.40
	%	65.0	66.5	N/A	47.5	64.0					± 8.0
MONO	#	0.20	0.20	N/A	0.42	0.20					± 0.20
	%	9.0	10.0	N/A	20.0	9.5					± 6.0
GRANULO	#	0.55	0.47	N/A	0.68	0.56					± 0.40
	%	26.0	23.5	N/A	32.5	26.5					± 7.0

LOT MX 419 N
 Rev 1

CONTROL N

(Exp.)

2019-12-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.5	7.0	7.2	7.3	7.4					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.59	4.55	4.58	4.53	4.60					± 0.18
HB HGB	g/dl	13.3	13.2	13.3	13.4	13.4					± 0.5
	g/l	133	132	133	134	134					± 5
	mmol/l	8.26	8.20	8.26	8.32	8.32					± 0.31
HT HCT	%	36.3	35.9	36.2	35.8	36.3					± 2.5
	l/l	0.363	0.359	0.362	0.358	0.363					± 0.025
VGM MCV	µm ³ ; fl	79	79	79	79	79					± 4
TGMH MCH	pg	29.0	29.0	29.0	29.6	29.1					± 2.0
	fmol	1.80	1.80	1.80	1.84	1.81					± 0.12
CCMH MCHC	g/dl	36.7	36.7	36.8	37.4	36.9					± 3.0
	g/l	367	367	368	374	369					± 30
	mmol/l	22.78	22.80	22.83	23.25	22.90					± 1.86
IDR RDW	%	14.5	13.5	N/A	14.0	13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	267	260	255	268	265					± 40
	µm ³ ; fl	7.5	7.5	N/A	8.2	7.2					± 2.0
LYMPHO	#	2.51	2.35	N/A	1.97	2.41					± 0.40
	%	33.5	33.5	N/A	27.0	32.5					± 5.0
MONO	#	0.53	0.56	N/A	0.55	0.52					± 0.30
	%	7.0	8.0	N/A	7.5	7.0					± 4.0
GRANULO	#	4.46	4.10	N/A	4.78	4.48					± 0.60
	%	59.5	58.5	N/A	65.5	60.5					± 6.0

Ref: TEMP-0830 Rev.39 FRONT / RECTO 130060513

LOT		MX 419 H		CONTROL H						(Exp.)	2019-12-05 (YYYY-MM-DD)
Rev 1		ABX Lyse									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60 Care ST	CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.0	19.0	19.3	19.8	20.0					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.65	5.62	5.64	5.65	5.68					± 0.20
HB HGB	g/dl	17.8	17.6	17.7	18.0	17.9					± 0.6
	g/l	178	176	177	180	179					± 6
	mmol/l	11.05	10.93	10.99	11.18	11.12					± 0.37
HT HCT	%	49.7	49.5	49.6	49.7	50.0					± 3.0
	l/l	0.497	0.495	0.496	0.497	0.500					± 0.030
VGM MCV	µm ³ ; fl	88	88	88	88	88					± 4
TGMH MCH	pg	31.5	31.3	31.4	31.9	31.5					± 2.0
	fmol	1.96	1.94	1.95	1.98	1.96					± 0.12
CCMH MCHC	g/dl	35.8	35.6	35.7	36.2	35.8					± 3.0
	g/l	358	356	357	362	358					± 30
	mmol/l	22.23	22.10	22.15	22.48	22.24					± 1.86
IDR RDW	%	13.5	13.5	N/A	13.5	12.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	508	500	515	525	510					± 45
VPM MPV	µm ³ ; fl	7.8	7.8	N/A	8.5	7.6					± 2.0
LYMPHO	#	3.30	3.04	N/A	2.57	3.10					± 1.00
	%	16.5	16.0	N/A	13.0	15.5					± 5.0
MONO	#	1.00	1.14	N/A	0.80	0.90					± 0.80
	%	5.0	6.0	N/A	4.0	4.5					± 4.0
GRANULO	#	15.70	14.82	N/A	16.43	16.00					± 1.40
	%	78.5	78.0	N/A	83.0	80.0					± 7.0