



**LOT** MX 429 L  
 Rev 1

**CONTROL** L



**2021-08-05**  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.0	2.0			2.0					± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.27	2.24			2.28					± 0.15
HB HGB	g/dl	6.1	6.0			6.1					± 0.4
	g/l	61	60			61					± 4
	mmol/l	3.79	3.73			3.79					± 0.25
HT HCT	%	16.6	16.4			16.9					± 2.0
	l/l	0.166	0.164			0.169					± 0.020
VGM MCV	µm <sup>3</sup> ; fl	73	73			74					± 4
TGMH MCH	pg	26.9	26.8			26.8					± 2.0
	fmol	1.67	1.66			1.66					± 0.12
CCMH MCHC	g/dl	36.8	36.7			36.2					± 3.0
	g/l	368	367			362					± 30
	mmol/l	22.86	22.79			22.45					± 1.86
IDR RDW	%	15.0	14.0			14.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	72	70			73					± 20
VPM MPV	µm <sup>3</sup> ; fl	8.2	8.1			7.6					± 2.0
LYMPHO	#	1.34	1.38			1.34					± 0.40
	%	67.0	69.0			67.0					± 8.0
MONO	#	0.20	0.20			0.20					± 0.20
	%	8.0	8.0			7.5					± 6.0
GRANULO	#	0.50	0.46			0.51					± 0.40
	%	25.0	23.0			25.5					± 7.0

**LOT** MX 429 N  
 Rev 1

**CONTROL** N

(Exp.)

**2021-08-05**  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.8	7.6			7.8					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.55	4.50			4.55					± 0.18
HB HGB	g/dl	13.4	13.3			13.5					± 0.5
	g/l	134	133			135					± 5
	mmol/l	8.32	8.26			8.38					± 0.31
HT HCT	%	37.8	37.4			37.8					± 2.5
	l/l	0.378	0.374			0.378					± 0.025
VGM MCV	µm <sup>3</sup> ; fl	83	83			83					± 4
TGMH MCH	pg	29.5	29.6			29.7					± 2.0
	fmol	1.83	1.84			1.84					± 0.12
CCMH MCHC	g/dl	35.5	35.6			35.7					± 3.0
	g/l	355	356			357					± 30
	mmol/l	22.03	22.11			22.20					± 1.86
IDR RDW	%	14.0	13.5			13.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	251	243			246					± 40
VPM MPV	µm <sup>3</sup> ; fl	8.1	7.9			7.5					± 2.0
LYMPHO	#	2.73	2.70			2.73					± 0.40
	%	35.0	35.5			35.0					± 5.0
MONO	#	0.62	0.65			0.62					± 0.30
	%	8.0	8.5			8.0					± 4.0
GRANULO	#	4.45	4.26			4.45					± 0.60
	%	57.0	56.0			57.0					± 6.0

Ref: TEMP-0830 Rev.39 FRONT / RECTO 130060513



**LOT**

**MX 429 H**

**CONTROL**

**H**

(Exp.)

**2021-08-05**  
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>20.6</b>	<b>20.2</b>			<b>20.7</b>					± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>5.64</b>	<b>5.55</b>			<b>5.63</b>					± 0.20
HB HGB	g/dl	<b>18.0</b>	<b>17.9</b>			<b>18.2</b>					± 0.6
	g/l	<b>180</b>	<b>179</b>			<b>182</b>					± 6
	mmol/l	<b>11.18</b>	<b>11.12</b>			<b>11.30</b>					± 0.37
HT HCT	%	<b>50.8</b>	<b>50.5</b>			<b>51.2</b>					± 3.0
	l/l	<b>0.508</b>	<b>0.505</b>			<b>0.512</b>					± 0.030
VGM MCV	µm <sup>3</sup> ; fl	<b>90</b>	<b>91</b>			<b>91</b>					± 4
TGMH MCH	pg	<b>31.9</b>	<b>32.3</b>			<b>32.3</b>					± 2.0
	fmol	<b>1.98</b>	<b>2.00</b>			<b>2.01</b>					± 0.12
CCMH MCHC	g/dl	<b>35.5</b>	<b>35.4</b>			<b>35.5</b>					± 3.0
	g/l	<b>355</b>	<b>354</b>			<b>355</b>					± 30
	mmol/l	<b>22.02</b>	<b>22.01</b>			<b>22.06</b>					± 1.86
IDR RDW	%	<b>13.7</b>	<b>13.5</b>			<b>12.5</b>					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>499</b>	<b>480</b>			<b>494</b>					± 45
VPM MPV	µm <sup>3</sup> ; fl	<b>7.8</b>	<b>7.8</b>			<b>7.4</b>					± 2.0
LYMPHO	#	<b>3.19</b>	<b>3.33</b>			<b>3.21</b>					± 1.00
	%	<b>15.5</b>	<b>16.5</b>			<b>15.5</b>					± 5.0
MONO	#	<b>1.03</b>	<b>1.11</b>			<b>0.93</b>					± 0.80
	%	<b>5.0</b>	<b>5.5</b>			<b>4.5</b>					± 4.0
GRANULO	#	<b>16.38</b>	<b>15.76</b>			<b>16.56</b>					± 1.40
	%	<b>79.5</b>	<b>78.0</b>			<b>80.0</b>					± 7.0