

LOT		MX 418 L		CONTROL L						(Exp.)	2019-10-05 (YYYY-MM-DD)
Rev 1		ABX Lyse									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60 Care ST	CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0	1.9	2.1	2.0	2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.32	2.29	2.32	2.30	2.32					± 0.15
HB HGB	g/dl	6.0	6.0	6.0	6.0	6.1					± 0.4
	g/l	60	60	60	60	61					± 4
	mmol/l	3.73	3.73	3.73	3.73	3.79					± 0.25
HT HCT	%	16.2	16.0	16.2	16.3	16.2					± 2.0
	l/l	0.162	0.160	0.162	0.163	0.162					± 0.020
VGM MCV	µm ³ ; fl	70	70	70	71	70					± 4
TGMH MCH	pg	25.9	26.2	25.9	26.1	26.3					± 2.0
	fmol	1.61	1.63	1.61	1.62	1.63					± 0.12
CCMH MCHC	g/dl	36.9	37.4	36.9	36.7	37.6					± 3.0
	g/l	369	374	369	367	376					± 30
	mmol/l	22.94	23.24	22.94	22.82	23.33					± 1.86
IDR RDW	%	14.0	14.0	N/A	14.0	13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	76	78	70	75	78					± 20
VPM MPV	µm ³ ; fl	8.5	8.1	N/A	8.6	7.6					± 2.0
LYMPHO	#	1.29	1.27	N/A	1.01	1.28					± 0.40
	%	64.5	67.0	N/A	50.5	64.0					± 8.0
MONO	#	0.20	0.20	N/A	0.35	0.20					± 0.20
	%	9.0	9.5	N/A	17.5	9.5					± 6.0
GRANULO	#	0.53	0.45	N/A	0.64	0.53					± 0.40
	%	26.5	23.5	N/A	32.0	26.5					± 7.0

LOT
 MX 418 N
 Rev 1

CONTROL N

(Exp.)

2019-10-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.7	7.3	7.6	7.6	7.8					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.60	4.55	4.64	4.56	4.61					± 0.18
HB HGB	g/dl	13.3	13.2	13.3	13.3	13.4					± 0.5
	g/l	133	132	133	133	134					± 5
	mmol/l	8.26	8.20	8.26	8.26	8.32					± 0.31
HT HCT	%	36.8	36.4	37.1	36.5	36.9					± 2.5
	l/l	0.368	0.364	0.371	0.365	0.369					± 0.025
VGM MCV	µm ³ ; fl	80	80	80	80	80					± 4
TGMH MCH	pg	28.9	29.0	28.7	29.2	29.1					± 2.0
	fmol	1.80	1.80	1.78	1.81	1.81					± 0.12
CCMH MCHC	g/dl	36.1	36.3	35.8	36.5	36.3					± 3.0
	g/l	361	363	358	365	363					± 30
	mmol/l	22.44	22.52	22.25	22.64	22.56					± 1.86
IDR RDW	%	14.0	14.0	N/A	14.0	13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	254	253	251	256	253					± 40
VPM MPV	µm ³ ; fl	8.2	7.9	N/A	8.7	7.5					± 2.0
LYMPHO	#	2.77	2.66	N/A	2.17	2.73					± 0.40
	%	36.0	36.5	N/A	28.5	35.0					± 5.0
MONO	#	0.50	0.55	N/A	0.72	0.55					± 0.30
	%	6.5	7.5	N/A	9.5	7.0					± 4.0
GRANULO	#	4.43	4.09	N/A	4.71	4.52					± 0.60
	%	57.5	56.0	N/A	62.0	58.0					± 6.0

Ref: TEMP-0830 Rev.39 FRONT / RECTO 130060513

LOT MX 418 H
 Rev 1

CONTROL H

(Exp.)

2019-10-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.8	20.0	20.2	20.2	20.7					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.62	5.57	5.64	5.54	5.60					± 0.20
HB HGB	g/dl	17.9	17.5	17.8	18.0	17.9					± 0.6
	g/l	179	175	178	180	179					± 6
	mmol/l	11.12	10.87	11.05	11.18	11.12					± 0.37
HT HCT	%	50.0	49.6	50.2	49.3	49.8					± 3.0
	l/l	0.500	0.496	0.502	0.493	0.498					± 0.030
VGM MCV	µm ³ ; fl	89	89	89	89	89					± 4
TGMH MCH	pg	31.9	31.4	31.6	32.5	32.0					± 2.0
	fmol	1.98	1.95	1.96	2.02	1.98					± 0.12
CCMH MCHC	g/dl	35.8	35.3	35.5	36.5	35.9					± 3.0
	g/l	358	353	355	365	359					± 30
	mmol/l	22.22	21.92	22.02	22.67	22.30					± 1.86
IDR RDW	%	13.5	13.5	N/A	13.5	12.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	520	515	520	530	522					± 45
VPM MPV	µm ³ ; fl	8.0	7.8	N/A	8.5	7.4					± 2.0
LYMPHO	#	3.43	3.40	N/A	2.63	3.42					± 1.00
	%	16.5	17.0	N/A	13.0	16.5					± 5.0
MONO	#	1.14	1.20	N/A	0.91	1.04					± 0.80
	%	5.5	6.0	N/A	4.5	5.0					± 4.0
GRANULO	#	16.22	15.40	N/A	16.67	16.25					± 1.40
	%	78.0	77.0	N/A	82.5	78.5					± 7.0