

LOT MX 427 L
 Rev 1

CONTROL L

(Exp.)

2021-04-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0	2.0	2.1	2.0	2.1					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.35	2.32	2.41	2.34	2.35					± 0.15
HB HGB	g/dl	6.0	5.9	6.1	6.0	6.1					± 0.4
	g/l	60	59	61	60	61					± 4
	mmol/l	3.73	3.66	3.79	3.73	3.79					± 0.25
HT HCT	%	16.2	16.0	16.9	16.4	16.5					± 2.0
	l/l	0.162	0.160	0.169	0.164	0.165					± 0.020
VGM MCV	µm ³ ; fl	69	69	70	70	70					± 4
TGMH MCH	pg	25.5	25.4	25.3	25.6	26.0					± 2.0
	fmol	1.59	1.58	1.57	1.59	1.61					± 0.12
CCMH MCHC	g/dl	37.0	36.9	36.2	36.6	37.1					± 3.0
	g/l	370	369	362	366	371					± 30
	mmol/l	22.98	22.89	22.45	22.75	23.03					± 1.86
IDR RDW	%	14.5	14.0	N/A	14.0	13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	83	83	73	79	81					± 20
VPM MPV	µm ³ ; fl	8.3	8.1	N/A	8.7	7.6					± 2.0
LYMPHO	#	1.27	1.34	N/A	0.99	1.34					± 0.40
	%	63.7	67.0	N/A	49.3	63.9					± 8.0
MONO	#	0.20	0.19	N/A	0.37	0.20					± 0.20
	%	9.8	9.5	N/A	18.7	9.5					± 6.0
GRANULO	#	0.53	0.47	N/A	0.64	0.56					± 0.40
	%	26.5	23.5	N/A	32.0	26.6					± 7.0

LOT MX 427 N
 Rev 1

CONTROL N

(Exp.)

2021-04-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.7	7.5	7.7	7.6	7.7					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.69	4.67	4.75	4.70	4.71					± 0.18
HB HGB	g/dl	13.7	13.5	13.8	13.8	13.8					± 0.5
	g/l	137	135	138	138	138					± 5
	mmol/l	8.51	8.38	8.57	8.57	8.57					± 0.31
HT HCT	%	38.0	37.8	38.5	38.1	38.2					± 2.5
	l/l	0.380	0.378	0.385	0.381	0.382					± 0.025
VGM MCV	µm ³ ; fl	81	81	81	81	81					± 4
TGMH MCH	pg	29.2	28.9	29.1	29.4	29.3					± 2.0
	fmol	1.81	1.80	1.80	1.82	1.82					± 0.12
CCMH MCHC	g/dl	36.1	35.7	35.9	36.2	36.2					± 3.0
	g/l	361	357	359	362	362					± 30
	mmol/l	22.40	22.16	22.27	22.51	22.46					± 1.86
IDR RDW	%	13.8	13.5	N/A	13.5	13.2					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	280	274	268	273	275					± 40
VPM MPV	µm ³ ; fl	7.9	8.1	N/A	8.5	7.4					± 2.0
LYMPHO	#	2.66	2.67	N/A	2.09	2.63					± 0.40
	%	34.5	35.6	N/A	27.5	34.1					± 5.0
MONO	#	0.59	0.62	N/A	0.77	0.58					± 0.30
	%	7.6	8.2	N/A	10.1	7.5					± 4.0
GRANULO	#	4.46	4.22	N/A	4.74	4.50					± 0.60
	%	57.9	56.2	N/A	62.4	58.4					± 6.0

Ref: TEMP-0830 Rev.39 FRONT / RECTO 130060513

LOT		MX 427 H		CONTROL H			(Exp.)				2021-04-05 (YYYY - MM - DD)					
Rev 1		ABX Lyse														
PARAMETRES PARAMETERS	UNITES UNITS	MICROS		MICROS		MICROS		MICROS		MICROS		ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60 Care ST	CRP 200	8P	18P	60 8P - 60 18P ADVIA 60										
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	21.1	20.3	20.5	20.7	21.0										± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.72	5.65	5.73	5.65	5.72										± 0.20
HB HGB	g/dl	18.0	17.7	17.9	18.1	18.1										± 0.6
	g/l	180	177	179	181	181										± 6
	mmol/l	11.18	10.99	11.12	11.24	11.24										± 0.37
HT HCT	%	50.3	49.7	50.4	49.7	50.9										± 3.0
	l/l	0.503	0.497	0.504	0.497	0.509										± 0.030
VGM MCV	µm ³ ; fl	88	88	88	88	89										± 4
TGMH MCH	pg	31.5	31.3	31.2	32.0	31.6										± 2.0
	fmol	1.95	1.95	1.94	1.99	1.97										± 0.12
CCMH MCHC	g/dl	35.8	35.6	35.5	36.4	35.6										± 3.0
	g/l	358	356	355	364	356										± 30
	mmol/l	22.21	22.11	22.04	22.61	22.08										± 1.86
IDR RDW	%	12.8	12.8	N/A	12.6	12.0										± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	550	536	545	552	543										± 45
VPM MPV	µm ³ ; fl	7.7	7.8	N/A	8.4	7.2										± 2.0
LYMPHO	#	3.44	3.27	N/A	2.63	3.28										± 1.00
	%	16.3	16.1	N/A	12.7	15.6										± 5.0
MONO	#	0.99	1.08	N/A	0.89	0.99										± 0.80
	%	4.7	5.3	N/A	4.3	4.7										± 4.0
GRANULO	#	16.67	15.96	N/A	17.18	16.74										± 1.40
	%	79.0	78.6	N/A	83.0	79.7										± 7.0