

LOT		MX 406 L		CONTROL L						(Exp.)	2017-10-05 (YYYY-MM-DD)
Rev 1		ABX Lyse									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0	2.0	2.1	2.0	2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.41	2.45	2.43	2.44	2.43					± 0.15
HB HGB	g/dl	6.3	6.3	6.2	6.2	6.3					± 0.4
	g/l	63	63	62	62	63					± 4
	mmol/l	3.91	3.91	3.85	3.85	3.91					± 0.25
HT HCT	%	16.6	17.2	17.0	17.1	17.0					± 2.0
	l/l	0.166	0.172	0.170	0.171	0.170					± 0.020
VGM MCV	µm ³ ; fl	69	70	70	70	70					± 4
TGMH MCH	pg	26.1	25.7	25.5	25.4	25.9					± 2.0
	fmol	1.62	1.60	1.58	1.58	1.61					± 0.12
CCMH MCHC	g/dl	37.9	36.7	36.4	36.3	37.0					± 3.0
	g/l	379	367	364	363	370					± 30
	mmol/l	23.53	22.81	22.63	22.54	23.00					± 1.86
IDR RDW	%	14.5	14.0	N/A	14.5	12.8					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	77	75	70	80	75					± 20
VPM MPV	µm ³ ; fl	8.8	8.8	N/A	8.5	8.0					± 2.0
LYMPHO	#	1.31	1.32	N/A	1.13	1.28					± 0.40
	%	65.5	66.0	N/A	56.5	64.0					± 8.0
MONO	#	0.20	0.20	N/A	0.22	0.20					± 0.20
	%	8.0	8.5	N/A	11.0	9.0					± 6.0
GRANULO	#	0.53	0.51	N/A	0.65	0.54					± 0.40
	%	26.5	25.5	N/A	32.5	27.0					± 7.0

LOT
 MX 406 N
 Rev 1

CONTROL N

(Exp.)

2017-10-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.4	7.4	7.5	7.3	7.5					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.74	4.77	4.75	4.75	4.79					± 0.18
HB HGB	g/dl	13.4	13.4	13.4	13.4	13.5					± 0.5
	g/l	134	134	134	134	135					± 5
	mmol/l	8.32	8.32	8.32	8.32	8.38					± 0.31
HT HCT	%	37.9	38.2	38.0	38.0	38.3					± 2.5
	l/l	0.379	0.382	0.380	0.380	0.383					± 0.025
VGM MCV	µm ³ ; fl	80	80	80	80	80					± 4
TGMH MCH	pg	28.3	28.1	28.2	28.2	28.2					± 2.0
	fmol	1.76	1.74	1.75	1.75	1.75					± 0.12
CCMH MCHC	g/dl	35.3	35.1	35.3	35.3	35.2					± 3.0
	g/l	353	351	353	353	352					± 30
	mmol/l	21.94	21.81	21.90	21.90	21.88					± 1.86
IDR RDW	%	14.0	13.8	N/A	14.2	12.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	251	250	250	253	250					± 40
VPM MPV	µm ³ ; fl	8.3	8.3	N/A	8.5	7.7					± 2.0
LYMPHO	#	2.52	2.44	N/A	2.08	2.48					± 0.40
	%	34.0	33.0	N/A	28.5	33.0					± 5.0
MONO	#	0.63	0.67	N/A	0.58	0.60					± 0.30
	%	8.5	9.0	N/A	8.0	8.0					± 4.0
GRANULO	#	4.26	4.29	N/A	4.64	4.43					± 0.60
	%	57.5	58.0	N/A	63.5	59.0					± 6.0

Ref: TEMP-0830 Rev.36 FRONT / RECTO 1219930084-B

LOT

MX 406 H

CONTROL

H

(Exp.)

2017-10-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.1	20.1	20.1	19.9	20.3					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.78	5.76	5.76	5.78	5.80					± 0.20
HB HGB	g/dl	18.0	18.0	17.8	18.0	18.1					± 0.6
	g/l	180	180	178	180	181					± 6
	mmol/l	11.18	11.18	11.05	11.18	11.24					± 0.37
HT HCT	%	50.9	50.7	50.7	50.9	51.0					± 3.0
	l/l	0.509	0.507	0.507	0.509	0.510					± 0.030
VGM MCV	µm ³ ; fl	88	88	88	88	88					± 4
TGMH MCH	pg	31.1	31.3	30.9	31.1	31.2					± 2.0
	fmol	1.93	1.94	1.92	1.93	1.94					± 0.12
CCMH MCHC	g/dl	35.4	35.5	35.1	35.4	35.5					± 3.0
	g/l	354	355	351	354	355					± 30
	mmol/l	21.98	22.05	21.81	21.98	22.02					± 1.86
IDR RDW	%	13.5	13.5	N/A	13.8	12.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	500	500	500	505	495					± 45
VPM MPV	µm ³ ; fl	7.5	7.4	N/A	7.6	7.0					± 2.0
LYMPHO	#	3.12	3.12	N/A	2.59	3.05					± 1.00
	%	15.5	15.5	N/A	13.0	15.0					± 5.0
MONO	#	1.31	1.31	N/A	0.80	1.12					± 0.80
	%	6.5	6.5	N/A	4.0	5.5					± 4.0
GRANULO	#	15.68	15.68	N/A	16.52	16.14					± 1.40
	%	78.0	78.0	N/A	83.0	79.5					± 7.0