



**LOT**

**MX 434 L**  
 Rev 1

**CONTROL L**

(Exp.)

**2022-06-05**  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>2.0</b>	<b>2.0</b>			<b>2.0</b>					± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>2.23</b>	<b>2.26</b>			<b>2.23</b>					± 0.17
HB HGB	g/dl	<b>6.0</b>	<b>6.1</b>			<b>6.0</b>					± 0.5
	g/l	<b>60</b>	<b>61</b>			<b>60</b>					± 5
	mmol/l	<b>3.73</b>	<b>3.79</b>			<b>3.73</b>					± 0.31
HT HCT	%	<b>16.1</b>	<b>16.3</b>			<b>16.1</b>					± 2.0
	l/l	<b>0.161</b>	<b>0.163</b>			<b>0.161</b>					± 0.020
VGM MCV	µm <sup>3</sup> ; fl	<b>72</b>	<b>72</b>			<b>72</b>					± 4
TGMH MCH	pg	<b>26.9</b>	<b>27.0</b>			<b>26.9</b>					± 2.0
	fmol	<b>1.67</b>	<b>1.68</b>			<b>1.67</b>					± 0.12
CCMH MCHC	g/dl	<b>37.4</b>	<b>37.5</b>			<b>37.4</b>					± 3.0
	g/l	<b>374</b>	<b>375</b>			<b>374</b>					± 30
	mmol/l	<b>23.21</b>	<b>23.28</b>			<b>23.21</b>					± 1.86
IDR RDW	%	<b>15.0</b>	<b>13.5</b>			<b>13.5</b>					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>80</b>	<b>79</b>			<b>78</b>					± 22
VPM MPV	µm <sup>3</sup> ; fl	<b>8.7</b>	<b>8.5</b>			<b>8.3</b>					± 2.0
LYMPHO	#	<b>1.23</b>	<b>1.26</b>			<b>1.26</b>					± 0.40
	%	<b>61.5</b>	<b>63.0</b>			<b>63.0</b>					± 8.0
MONO	#	<b>0.23</b>	<b>0.25</b>			<b>0.22</b>					± 0.20
	%	<b>11.5</b>	<b>12.5</b>			<b>11.0</b>					± 6.0
GRANULO	#	<b>0.54</b>	<b>0.49</b>			<b>0.52</b>					± 0.40
	%	<b>27.0</b>	<b>24.5</b>			<b>26.0</b>					± 7.0

**LOT** MX 434 N  
Rev 1

**CONTROL** N

 (Exp.)

**2022-06-05**  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.8	7.8			7.8					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.55	4.61			4.59					± 0.26
HB HGB	g/dl	13.4	13.6			13.6					± 0.7
	g/l	134	136			136					± 7
	mmol/l	8.32	8.45			8.45					± 0.43
HT HCT	%	37.3	37.3			37.6					± 2.7
	l/l	0.373	0.373			0.376					± 0.027
VGM MCV	µm <sup>3</sup> ; fl	82	81			82					± 4
TGMH MCH	pg	29.5	29.5			29.6					± 2.0
	fmol	1.83	1.83			1.84					± 0.12
CCMH MCHC	g/dl	35.9	36.4			36.1					± 3.0
	g/l	359	364			361					± 30
	mmol/l	22.30	22.62			22.44					± 1.86
IDR RDW	%	14.0	13.0			13.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	259	260			258					± 40
VPM MPV	µm <sup>3</sup> ; fl	8.4	7.9			7.8					± 2.0
LYMPHO	#	2.57	2.65			2.57					± 0.40
	%	33.0	34.0			33.0					± 5.0
MONO	#	0.55	0.62			0.51					± 0.30
	%	7.0	8.0			6.5					± 4.0
GRANULO	#	4.68	4.52			4.72					± 0.60
	%	60.0	58.0			60.5					± 6.0

Ref: TEMP-0830 Rev.41 FRONT / RECTO 130090307

**LOT**

**MX 434 H**  
Rev 1

**CONTROL H**

(Exp.)

**2022-06-05**  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>20.3</b>	<b>20.2</b>			<b>20.5</b>					± 1.7
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>5.64</b>	<b>5.66</b>			<b>5.70</b>					± 0.32
HB HGB	g/dl	<b>18.2</b>	<b>18.3</b>			<b>18.3</b>					± 0.9
	g/l	<b>182</b>	<b>183</b>			<b>183</b>					± 9
	mmol/l	<b>11.30</b>	<b>11.36</b>			<b>11.36</b>					± 0.56
HT HCT	%	<b>50.2</b>	<b>50.9</b>			<b>50.9</b>					± 3.4
	l/l	<b>0.502</b>	<b>0.509</b>			<b>0.509</b>					± 0.034
VGM MCV	µm <sup>3</sup> ; fl	<b>89</b>	<b>90</b>			<b>89</b>					± 4
TGMH MCH	pg	<b>32.3</b>	<b>32.3</b>			<b>32.1</b>					± 2.0
	fmol	<b>2.00</b>	<b>2.01</b>			<b>1.99</b>					± 0.12
CCMH MCHC	g/dl	<b>36.3</b>	<b>35.9</b>			<b>36.1</b>					± 3.0
	g/l	<b>363</b>	<b>359</b>			<b>361</b>					± 30
	mmol/l	<b>22.52</b>	<b>22.31</b>			<b>22.40</b>					± 1.86
IDR RDW	%	<b>13.0</b>	<b>13.0</b>			<b>12.0</b>					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>518</b>	<b>515</b>			<b>515</b>					± 65
VPM MPV	µm <sup>3</sup> ; fl	<b>8.3</b>	<b>7.9</b>			<b>7.8</b>					± 2.0
LYMPHO	#	<b>3.15</b>	<b>3.23</b>			<b>3.08</b>					± 1.00
	%	<b>15.5</b>	<b>16.0</b>			<b>15.0</b>					± 5.0
MONO	#	<b>1.02</b>	<b>1.11</b>			<b>1.03</b>					± 0.80
	%	<b>5.0</b>	<b>5.5</b>			<b>5.0</b>					± 4.0
GRANULO	#	<b>16.14</b>	<b>15.86</b>			<b>16.40</b>					± 1.40
	%	<b>79.5</b>	<b>78.5</b>			<b>80.0</b>					± 7.0