



**LOT**

**MX 035 L**  
 Rev 1

**CONTROL L**

(Exp.)

**2015-06-05**  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.0	2.0	2.1	2.0	2.1					± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.35	2.35	2.37	2.33	2.34					± 0.15
HB HGB	g/dl	6.1	6.1	6.0	6.0	6.1					± 0.4
	g/l	61	61	60	60	61					± 4
	mmol/l	3.79	3.79	3.73	3.73	3.79					± 0.25
HT HCT	%	16.2	16.2	16.6	16.3	16.1					± 2.0
	l/l	0.162	0.162	0.166	0.163	0.161					± 0.020
VGM MCV	µm <sup>3</sup> ; fl	69	69	70	70	69					± 4
TGMH MCH	pg	26.0	26.0	25.3	25.8	26.1					± 2.0
	fmol	1.61	1.61	1.57	1.60	1.62					± 0.12
CCMH MCHC	g/dl	37.6	37.6	36.2	36.8	37.8					± 3.0
	g/l	376	376	362	368	378					± 30
	mmol/l	23.36	23.36	22.46	22.84	23.46					± 1.86
IDR RDW	%	14.0	13.5	N/A	14.0	13.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	78	74	70	75	75					± 20
VPM MPV	µm <sup>3</sup> ; fl	9.3	9.0	N/A	9.2	8.5					± 2.0
LYMPHO	#	1.24	1.24	N/A	1.01	1.28					± 0.40
	%	62.0	62.0	N/A	50.5	61.0					± 8.0
MONO	#	0.22	0.22	N/A	0.28	0.23					± 0.20
	%	11.0	11.0	N/A	14.0	11.0					± 6.0
GRANULO	#	0.54	0.54	N/A	0.71	0.59					± 0.40
	%	27.0	27.0	N/A	35.5	28.0					± 7.0

LOT MX 035 N  
Rev 1

CONTROL N

(Exp.)

2015-06-05  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.9	7.9	7.7	7.7	7.9					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.55	4.55	4.58	4.51	4.55					± 0.18
HB HGB	g/dl	13.5	13.5	13.5	13.5	13.5					± 0.5
	g/l	135	135	135	135	135					± 5
	mmol/l	8.38	8.38	8.38	8.38	8.38					± 0.31
HT HCT	%	36.9	36.9	37.1	36.5	36.9					± 2.5
	l/l	0.369	0.369	0.371	0.365	0.369					± 0.025
VGM MCV	µm <sup>3</sup> ; fl	81	81	81	81	81					± 4
TGMH MCH	pg	29.7	29.7	29.5	29.9	29.7					± 2.0
	fmol	1.84	1.84	1.83	1.86	1.84					± 0.12
CCMH MCHC	g/dl	36.6	36.6	36.4	37.0	36.6					± 3.0
	g/l	366	366	364	370	366					± 30
	mmol/l	22.75	22.75	22.60	22.95	22.75					± 1.86
IDR RDW	%	13.5	13.0	N/A	13.0	13.3					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	258	259	261	263	262					± 40
VPM MPV	µm <sup>3</sup> ; fl	8.9	8.3	N/A	9.1	8.1					± 2.0
LYMPHO	#	2.53	2.45	N/A	2.00	2.45					± 0.40
	%	32.0	31.0	N/A	26.0	31.0					± 5.0
MONO	#	0.55	0.59	N/A	0.54	0.55					± 0.30
	%	7.0	7.5	N/A	7.0	7.0					± 4.0
GRANULO	#	4.82	4.86	N/A	5.16	4.90					± 0.60
	%	61.0	61.5	N/A	67.0	62.0					± 6.0

Ret: TEMP-0830 Rev.36 FRONT / RECTO 9950084-B

LOT

MX 035 H  
Rev 1

CONTROL H

 (Exp.)

2015-06-05  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	20.9	20.8	20.0	20.4	20.8					± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	5.93	5.92	5.95	5.91	5.94					± 0.20
HB HGB	g/dl	18.4	18.2	18.0	18.5	18.3					± 0.6
	g/l	184	182	180	185	183					± 6
	mmol/l	11.43	11.30	11.18	11.49	11.36					± 0.37
HT HCT	%	51.0	51.5	51.2	51.4	51.7					± 3.0
	l/l	0.510	0.515	0.512	0.514	0.517					± 0.030
VGM MCV	µm <sup>3</sup> ; fl	86	87	86	87	87					± 4
TGMH MCH	pg	31.0	30.7	30.3	31.3	30.8					± 2.0
	fmol	1.93	1.91	1.88	1.94	1.91					± 0.12
CCMH MCHC	g/dl	36.1	35.3	35.2	36.0	35.4					± 3.0
	g/l	361	353	352	360	354					± 30
	mmol/l	22.41	21.94	21.84	22.34	21.99					± 1.86
IDR RDW	%	13.0	12.5	N/A	12.8	13.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	500	500	525	520	510					± 45
VPM MPV	µm <sup>3</sup> ; fl	8.8	8.2	N/A	9.0	8.0					± 2.0
LYMPHO	#	3.55	3.33	N/A	2.75	3.33					± 1.00
	%	17.0	16.0	N/A	13.5	16.0					± 5.0
MONO	#	1.05	1.25	N/A	0.82	1.14					± 0.80
	%	5.0	6.0	N/A	4.0	5.5					± 4.0
GRANULO	#	16.30	16.22	N/A	16.83	16.33					± 1.40
	%	78.0	78.0	N/A	82.5	78.5					± 7.0