

**LOT** MX 095 L  
 Rev 1

**CONTROL** L

(Exp.)

**2015-12-05**  
 (YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.0	2.1	2.0	2.0	2.0					± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.45	2.45	2.45	2.45	2.45					± 0.15
HB HGB	g/dl	6.2	6.2	6.2	6.1	6.2					± 0.4
	g/l	62	62	62	61	62					± 4
	mmol/l	3.85	3.85	3.85	3.79	3.85					± 0.25
HT HCT	%	16.7	16.7	16.9	16.9	16.7					± 2.0
	l/l	0.167	0.167	0.169	0.169	0.167					± 0.020
VGM MCV	µm <sup>3</sup> ; fl	68	68	69	69	68					± 4
TGMH MCH	pg	25.3	25.3	25.3	24.9	25.3					± 2.0
	fmol	1.57	1.57	1.57	1.55	1.57					± 0.12
CCMH MCHC	g/dl	37.2	37.2	36.7	36.1	37.2					± 3.0
	g/l	372	372	367	361	372					± 30
	mmol/l	23.11	23.11	22.78	22.41	23.11					± 1.86
IDR RDW	%	14.5	14.0	N/A	14.5	14.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	78	71	67	72	75					± 20
VPM MPV	µm <sup>3</sup> ; fl	9.1	8.8	N/A	8.9	8.2					± 2.0
LYMPHO	#	1.22	1.30	N/A	0.97	1.22					± 0.40
	%	61.0	62.0	N/A	48.5	61.0					± 8.0
MONO	#	0.21	0.23	N/A	0.32	0.22					± 0.20
	%	10.5	11.0	N/A	16.0	11.0					± 6.0
GRANULO	#	0.57	0.57	N/A	0.71	0.56					± 0.40
	%	28.5	27.0	N/A	35.5	28.0					± 7.0

LOT MX 095 N  
Rev 1

CONTROL N

 (Exp.)

2015-12-05  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.7	7.7	7.6	7.6	7.7					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.50	4.50	4.50	4.50	4.50					± 0.18
HB HGB	g/dl	13.4	13.3	13.3	13.4	13.4					± 0.5
	g/l	134	133	133	134	134					± 5
	mmol/l	8.32	8.26	8.26	8.32	8.32					± 0.31
HT HCT	%	36.5	36.5	36.5	36.5	36.5					± 2.5
	l/l	0.365	0.365	0.365	0.365	0.365					± 0.025
VGM MCV	µm <sup>3</sup> ; fl	81	81	81	81	81					± 4
TGMH MCH	pg	29.8	29.6	29.6	29.8	29.8					± 2.0
	fmol	1.85	1.84	1.84	1.85	1.85					± 0.12
CCMH MCHC	g/dl	36.8	36.5	36.5	36.8	36.8					± 3.0
	g/l	368	365	365	368	368					± 30
	mmol/l	22.83	22.66	22.66	22.83	22.83					± 1.86
IDR RDW	%	13.8	13.5	N/A	13.5	13.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	265	255	250	255	255					± 40
VPM MPV	µm <sup>3</sup> ; fl	8.7	8.3	N/A	8.9	8.0					± 2.0
LYMPHO	#	2.62	2.58	N/A	2.09	2.58					± 0.40
	%	34.0	33.5	N/A	27.5	33.5					± 5.0
MONO	#	0.54	0.58	N/A	0.65	0.58					± 0.30
	%	7.0	7.5	N/A	8.5	7.5					± 4.0
GRANULO	#	4.54	4.54	N/A	4.86	4.54					± 0.60
	%	59.0	59.0	N/A	64.0	59.0					± 6.0

Ref: TEMP-0830 Rev.37 FRONT / RECTO 1219930084-B

<b>LOT</b>		<b>MX 095 H</b>	<b>CONTROL H</b>				(Exp.)				<b>2015-12-05</b> (YYYY-MM-DD)
Rev 1		<b>ABX Lyse</b>									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>19.9</b>	<b>20.1</b>	<b>19.8</b>	<b>19.9</b>	<b>20.2</b>					± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>5.70</b>	<b>5.65</b>	<b>5.67</b>	<b>5.68</b>	<b>5.70</b>					± 0.20
HB HGB	g/dl	<b>18.2</b>	<b>18.0</b>	<b>18.0</b>	<b>18.2</b>	<b>18.1</b>					± 0.6
	g/l	<b>182</b>	<b>180</b>	<b>180</b>	<b>182</b>	<b>181</b>					± 6
	mmol/l	<b>11.30</b>	<b>11.18</b>	<b>11.18</b>	<b>11.30</b>	<b>11.24</b>					± 0.37
HT HCT	%	<b>50.2</b>	<b>49.7</b>	<b>49.9</b>	<b>50.0</b>	<b>50.2</b>					± 3.0
	l/l	<b>0.502</b>	<b>0.497</b>	<b>0.499</b>	<b>0.500</b>	<b>0.502</b>					± 0.030
VGM MCV	µm <sup>3</sup> ; fl	<b>88</b>	<b>88</b>	<b>88</b>	<b>88</b>	<b>88</b>					± 4
TGMH MCH	pg	<b>31.9</b>	<b>31.9</b>	<b>31.7</b>	<b>32.0</b>	<b>31.8</b>					± 2.0
	fmol	<b>1.98</b>	<b>1.98</b>	<b>1.97</b>	<b>1.99</b>	<b>1.97</b>					± 0.12
CCMH MCHC	g/dl	<b>36.3</b>	<b>36.2</b>	<b>36.1</b>	<b>36.4</b>	<b>36.1</b>					± 3.0
	g/l	<b>363</b>	<b>362</b>	<b>361</b>	<b>364</b>	<b>361</b>					± 30
	mmol/l	<b>22.53</b>	<b>22.48</b>	<b>22.40</b>	<b>22.61</b>	<b>22.41</b>					± 1.86
IDR RDW	%	<b>13.0</b>	<b>13.0</b>	N/A	<b>13.0</b>	<b>13.0</b>					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>515</b>	<b>505</b>	<b>520</b>	<b>510</b>	<b>505</b>					± 45
VPM MPV	µm <sup>3</sup> ; fl	<b>8.4</b>	<b>7.9</b>	N/A	<b>8.5</b>	<b>7.5</b>					± 2.0
LYMPHO	#	<b>2.89</b>	<b>2.91</b>	N/A	<b>2.39</b>	<b>2.83</b>					± 1.00
	%	<b>14.5</b>	<b>14.5</b>	N/A	<b>12.0</b>	<b>14.0</b>					± 5.0
MONO	#	<b>1.09</b>	<b>1.11</b>	N/A	<b>0.80</b>	<b>1.11</b>					± 0.80
	%	<b>5.5</b>	<b>5.5</b>	N/A	<b>4.0</b>	<b>5.5</b>					± 4.0
GRANULO	#	<b>15.92</b>	<b>16.08</b>	N/A	<b>16.72</b>	<b>16.26</b>					± 1.40
	%	<b>80.0</b>	<b>80.0</b>	N/A	<b>84.0</b>	<b>80.5</b>					± 7.0