



LOT MX 075 L
 Rev 1

CONTROL L



2015-10-05
 (YYYY-MM-DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lyse | | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|--|-----------------|----------------|--------|--------|----------------------------|----------------|--------|--------|--------|-------------------------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | ARGOS - HELIOS | PENTRA | PENTRA | PENTRA | |
| | | ES60 Care ST | CRP CRP 200 | 8P | 18P | 60 8P - 60 18P ADVIA 60 | | | | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 2.0 | 2.1 | 2.1 | 2.0 | 2.0 | | | | | ± 0.4 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 2.41 | 2.40 | 2.44 | 2.37 | 2.40 | | | | | ± 0.15 |
| HB HGB | g/dl | 6.1 | 6.0 | 6.1 | 6.0 | 6.1 | | | | | ± 0.4 |
| | g/l | 61 | 60 | 61 | 60 | 61 | | | | | ± 4 |
| | mmol/l | 3.79 | 3.73 | 3.79 | 3.73 | 3.79 | | | | | ± 0.25 |
| HT HCT | % | 16.6 | 16.3 | 17.1 | 16.6 | 16.6 | | | | | ± 2.0 |
| | l/l | 0.166 | 0.163 | 0.171 | 0.166 | 0.166 | | | | | ± 0.020 |
| VGM MCV | µm ³ ; fl | 69 | 68 | 70 | 70 | 69 | | | | | ± 4 |
| TGMH MCH | pg | 25.3 | 25.0 | 25.0 | 25.3 | 25.4 | | | | | ± 2.0 |
| | fmol | 1.57 | 1.55 | 1.55 | 1.57 | 1.58 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | 36.7 | 36.8 | 35.7 | 36.2 | 36.8 | | | | | ± 3.0 |
| | g/l | 367 | 368 | 357 | 362 | 368 | | | | | ± 30 |
| | mmol/l | 22.78 | 22.83 | 22.18 | 22.46 | 22.88 | | | | | ± 1.86 |
| IDR RDW | % | 14.4 | 14.0 | N/A | 14.0 | 14.0 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 79 | 70 | 73 | 68 | 75 | | | | | ± 20 |
| VPM MPV | µm ³ ; fl | 9.5 | 9.0 | N/A | 9.1 | 8.2 | | | | | ± 2.0 |
| LYMPHO | # | 1.26 | 1.34 | N/A | 1.02 | 1.25 | | | | | ± 0.40 |
| | % | 63.0 | 64.0 | N/A | 51.0 | 62.5 | | | | | ± 8.0 |
| MONO | # | 0.25 | 0.26 | N/A | 0.34 | 0.26 | | | | | ± 0.20 |
| | % | 12.5 | 12.5 | N/A | 17.0 | 13.0 | | | | | ± 6.0 |
| GRANULO | # | 0.49 | 0.49 | N/A | 0.64 | 0.49 | | | | | ± 0.40 |
| | % | 24.5 | 23.5 | N/A | 32.0 | 24.5 | | | | | ± 7.0 |

LOT MX 075 N
Rev 1

CONTROL N

 (Exp.)

2015-10-05
(YYYY-MM-DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lyse | | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|--|-----------------|----------------|--------|--------|----------------------------|----------------|--------|--------|--------|-------------------------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | ARGOS - HELIOS | PENTRA | PENTRA | PENTRA | |
| | | ES60 Care ST | CRP CRP 200 | 8P | 18P | 60 8P - 60 18P ADVIA 60 | | | | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 7.5 | 7.5 | 7.4 | 7.4 | 7.5 | | | | | ± 0.8 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 4.60 | 4.60 | 4.60 | 4.54 | 4.60 | | | | | ± 0.18 |
| HB HGB | g/dl | 13.5 | 13.4 | 13.5 | 13.4 | 13.5 | | | | | ± 0.5 |
| | g/l | 135 | 134 | 135 | 134 | 135 | | | | | ± 5 |
| | mmol/l | 8.38 | 8.32 | 8.38 | 8.32 | 8.38 | | | | | ± 0.31 |
| HT HCT | % | 36.8 | 36.8 | 36.8 | 36.8 | 36.8 | | | | | ± 2.5 |
| | l/l | 0.368 | 0.368 | 0.368 | 0.368 | 0.368 | | | | | ± 0.025 |
| VGM MCV | µm ³ ; fl | 80 | 80 | 80 | 81 | 80 | | | | | ± 4 |
| TGMH MCH | pg | 29.3 | 29.1 | 29.3 | 29.5 | 29.3 | | | | | ± 2.0 |
| | fmol | 1.82 | 1.81 | 1.82 | 1.83 | 1.82 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | 36.7 | 36.4 | 36.7 | 36.4 | 36.7 | | | | | ± 3.0 |
| | g/l | 367 | 364 | 367 | 364 | 367 | | | | | ± 30 |
| | mmol/l | 22.78 | 22.61 | 22.78 | 22.63 | 22.78 | | | | | ± 1.86 |
| IDR RDW | % | 13.8 | 13.5 | N/A | 13.7 | 13.5 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 260 | 250 | 258 | 245 | 252 | | | | | ± 40 |
| VPM MPV | µm ³ ; fl | 8.9 | 8.5 | N/A | 8.8 | 8.0 | | | | | ± 2.0 |
| LYMPHO | # | 2.51 | 2.51 | N/A | 2.00 | 2.48 | | | | | ± 0.40 |
| | % | 33.5 | 33.5 | N/A | 27.0 | 33.0 | | | | | ± 5.0 |
| MONO | # | 0.60 | 0.64 | N/A | 0.63 | 0.60 | | | | | ± 0.30 |
| | % | 8.0 | 8.5 | N/A | 8.5 | 8.0 | | | | | ± 4.0 |
| GRANULO | # | 4.39 | 4.35 | N/A | 4.77 | 4.43 | | | | | ± 0.60 |
| | % | 58.5 | 58.0 | N/A | 64.5 | 59.0 | | | | | ± 6.0 |

Ref: TEMP-0830 Rev.37 FRONT / RECTO 1219930084-B

LOT MX 075 H
Rev 1

CONTROL H

 (Exp.)

2015-10-05
(YYYY-MM-DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lyse | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|--|-----------------|----------------|--------|--------|----------------------------|----------------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | ARGOS - HELIOS | PENTRA | PENTRA | | PENTRA |
| | | ES60 Care ST | CRP CRP 200 | 8P | 18P | 60 8P - 60 18P ADVIA 60 | | | | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 19.8 | 20.1 | 19.8 | 19.9 | 20.0 | | | | | ± 1.6 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 5.74 | 5.70 | 5.70 | 5.65 | 5.70 | | | | | ± 0.20 |
| HB HGB | g/dl | 17.9 | 17.8 | 17.8 | 18.0 | 17.8 | | | | | ± 0.6 |
| | g/l | 179 | 178 | 178 | 180 | 178 | | | | | ± 6 |
| | mmol/l | 11.12 | 11.05 | 11.05 | 11.18 | 11.05 | | | | | ± 0.37 |
| HT HCT | % | 50.5 | 50.2 | 50.2 | 49.7 | 50.2 | | | | | ± 3.0 |
| | l/l | 0.505 | 0.502 | 0.502 | 0.497 | 0.502 | | | | | ± 0.030 |
| VGM MCV | µm ³ ; fl | 88 | 88 | 88 | 88 | 88 | | | | | ± 4 |
| TGMH MCH | pg | 31.2 | 31.2 | 31.2 | 31.9 | 31.2 | | | | | ± 2.0 |
| | fmol | 1.94 | 1.94 | 1.94 | 1.98 | 1.94 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | 35.4 | 35.5 | 35.5 | 36.2 | 35.5 | | | | | ± 3.0 |
| | g/l | 354 | 355 | 355 | 362 | 355 | | | | | ± 30 |
| | mmol/l | 22.01 | 22.04 | 22.04 | 22.48 | 22.04 | | | | | ± 1.86 |
| IDR RDW | % | 13.2 | 13.0 | N/A | 13.2 | 13.0 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 530 | 520 | 542 | 515 | 520 | | | | | ± 45 |
| VPM MPV | µm ³ ; fl | 8.5 | 8.0 | N/A | 8.4 | 7.7 | | | | | ± 2.0 |
| LYMPHO | # | 2.97 | 2.91 | N/A | 2.39 | 2.80 | | | | | ± 1.00 |
| | % | 15.0 | 14.5 | N/A | 12.0 | 14.0 | | | | | ± 5.0 |
| MONO | # | 0.99 | 1.11 | N/A | 0.80 | 1.00 | | | | | ± 0.80 |
| | % | 5.0 | 5.5 | N/A | 4.0 | 5.0 | | | | | ± 4.0 |
| GRANULO | # | 15.84 | 16.08 | N/A | 16.72 | 16.20 | | | | | ± 1.40 |
| | % | 80.0 | 80.0 | N/A | 84.0 | 81.0 | | | | | ± 7.0 |