



**LOT**      **MX 430**    **L**  
 Rev 1

**CONTROL**    **L**



**2021-10-05**  
 (YYYY - MM - DD)

| PARAMETRES<br>PARAMETERS | UNITES<br>UNITS  | ABX Lyse        |              |        |        |                            |                |        |        | TOLERANCES<br>TOLERANCE |         |
|--------------------------|--|-----------------|--------------|--------|--------|----------------------------|----------------|--------|--------|-------------------------|---------|
|                          |  | MICROS          | MICROS       | MICROS | MICROS | MICROS                     | ARGOS - HELIOS | PENTRA | PENTRA |                         | PENTRA  |
|                          |  | ES60<br>Care ST | CRP 200      |        |        | 60 8P - 60 18P<br>ADVIA 60 |                |        |        |                         |         |
| GB WBC                   | 10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l  | <b>2.0</b>      | <b>2.0</b>   |        |        | <b>2.0</b>                 |                |        |        |                         | ± 0.4   |
| GR RBC                   | 10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l | <b>2.35</b>     | <b>2.30</b>  |        |        | <b>2.35</b>                |                |        |        |                         | ± 0.15  |
| HB HGB                   | g/dl   | <b>6.0</b>      | <b>6.0</b>   |        |        | <b>6.0</b>                 |                |        |        |                         | ± 0.4   |
|                          | g/l  | <b>60</b>       | <b>60</b>    |        |        | <b>60</b>                  |                |        |        |                         | ± 4     |
|                          | mmol/l   | <b>3.73</b>     | <b>3.73</b>  |        |        | <b>3.73</b>                |                |        |        |                         | ± 0.25  |
| HT HCT                   | %  | <b>16.5</b>     | <b>16.1</b>  |        |        | <b>16.5</b>                |                |        |        |                         | ± 2.0   |
|                          | l/l  | <b>0.165</b>    | <b>0.161</b> |        |        | <b>0.165</b>               |                |        |        |                         | ± 0.020 |
| VGM MCV                  | µm <sup>3</sup> ; fl                                   | <b>70</b>       | <b>70</b>    |        |        | <b>70</b>                  |                |        |        |                         | ± 4     |
| TGMH MCH                 | pg   | <b>25.5</b>     | <b>26.1</b>  |        |        | <b>25.5</b>                |                |        |        |                         | ± 2.0   |
|                          | fmol   | <b>1.59</b>     | <b>1.62</b>  |        |        | <b>1.59</b>                |                |        |        |                         | ± 0.12  |
| CCMH MCHC                | g/dl   | <b>36.5</b>     | <b>37.3</b>  |        |        | <b>36.5</b>                |                |        |        |                         | ± 3.0   |
|                          | g/l  | <b>365</b>      | <b>373</b>   |        |        | <b>365</b>                 |                |        |        |                         | ± 30    |
|                          | mmol/l   | <b>22.65</b>    | <b>23.14</b> |        |        | <b>22.65</b>               |                |        |        |                         | ± 1.86  |
| IDR RDW                  | %  | <b>14.5</b>     | <b>14.0</b>  |        |        | <b>13.5</b>                |                |        |        |                         | ± 3.0   |
| PLAQ. PLTS               | 10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l  | <b>77</b>       | <b>75</b>    |        |        | <b>78</b>                  |                |        |        |                         | ± 20    |
| VPM MPV                  | µm <sup>3</sup> ; fl                                   | <b>9.3</b>      | <b>9.3</b>   |        |        | <b>8.8</b>                 |                |        |        |                         | ± 2.0   |
| LYMPHO                   | #  | <b>1.36</b>     | <b>1.41</b>  |        |        | <b>1.35</b>                |                |        |        |                         | ± 0.40  |
|                          | %  | <b>68.0</b>     | <b>70.5</b>  |        |        | <b>67.5</b>                |                |        |        |                         | ± 8.0   |
| MONO                     | #  | <b>0.20</b>     | <b>0.20</b>  |        |        | <b>0.20</b>                |                |        |        |                         | ± 0.20  |
|                          | %  | <b>7.0</b>      | <b>7.5</b>   |        |        | <b>7.5</b>                 |                |        |        |                         | ± 6.0   |
| GRANULO                  | #  | <b>0.50</b>     | <b>0.44</b>  |        |        | <b>0.50</b>                |                |        |        |                         | ± 0.40  |
|                          | %  | <b>25.0</b>     | <b>22.0</b>  |        |        | <b>25.0</b>                |                |        |        |                         | ± 7.0   |

**LOT** MX 430 N  
 Rev 1

**CONTROL** N

(Exp.)

**2021-10-05**  
 (YYYY - MM - DD)

| PARAMETRES<br>PARAMETERS | UNITES<br>UNITS  | ABX Lyse        |         |        |        |                            |                |        |        | TOLERANCES<br>TOLERANCE |         |
|--------------------------|--|-----------------|---------|--------|--------|----------------------------|----------------|--------|--------|-------------------------|---------|
|                          |  | MICROS          | MICROS  | MICROS | MICROS | MICROS                     | ARGOS - HELIOS | PENTRA | PENTRA |                         | PENTRA  |
|                          |  | ES60<br>Care ST | CRP 200 |        |        | 60 8P - 60 18P<br>ADVIA 60 |                |        |        |                         |         |
| GB WBC                   | 10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l  | 7.9             | 7.8     |        |        | 8.0                        |                |        |        |                         | ± 0.8   |
| GR RBC                   | 10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l | 4.60            | 4.55    |        |        | 4.62                       |                |        |        |                         | ± 0.18  |
| HB HGB                   | g/dl   | 13.5            | 13.5    |        |        | 13.6                       |                |        |        |                         | ± 0.5   |
|                          | g/l  | 135             | 135     |        |        | 136                        |                |        |        |                         | ± 5     |
|                          | mmol/l   | 8.38            | 8.38    |        |        | 8.45                       |                |        |        |                         | ± 0.31  |
| HT HCT                   | %  | 37.7            | 37.3    |        |        | 37.4                       |                |        |        |                         | ± 2.5   |
|                          | l/l  | 0.377           | 0.373   |        |        | 0.374                      |                |        |        |                         | ± 0.025 |
| VGM MCV                  | µm <sup>3</sup> ; fl                                   | 82              | 82      |        |        | 81                         |                |        |        |                         | ± 4     |
| TGMH MCH                 | pg   | 29.3            | 29.7    |        |        | 29.4                       |                |        |        |                         | ± 2.0   |
|                          | fmol   | 1.82            | 1.84    |        |        | 1.83                       |                |        |        |                         | ± 0.12  |
| CCMH MCHC                | g/dl   | 35.8            | 36.2    |        |        | 36.3                       |                |        |        |                         | ± 3.0   |
|                          | g/l  | 358             | 362     |        |        | 363                        |                |        |        |                         | ± 30    |
|                          | mmol/l   | 22.23           | 22.47   |        |        | 22.57                      |                |        |        |                         | ± 1.86  |
| IDR RDW                  | %  | 14.5            | 14.0    |        |        | 13.5                       |                |        |        |                         | ± 3.0   |
| PLAQ. PLTS               | 10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l  | 258             | 253     |        |        | 260                        |                |        |        |                         | ± 40    |
| VPM MPV                  | µm <sup>3</sup> ; fl                                   | 8.4             | 8.3     |        |        | 7.8                        |                |        |        |                         | ± 2.0   |
| LYMPHO                   | #  | 2.92            | 2.96    |        |        | 2.92                       |                |        |        |                         | ± 0.40  |
|                          | %  | 37.0            | 38.0    |        |        | 36.5                       |                |        |        |                         | ± 5.0   |
| MONO                     | #  | 0.47            | 0.55    |        |        | 0.48                       |                |        |        |                         | ± 0.30  |
|                          | %  | 6.0             | 7.0     |        |        | 6.0                        |                |        |        |                         | ± 4.0   |
| GRANULO                  | #  | 4.50            | 4.29    |        |        | 4.60                       |                |        |        |                         | ± 0.60  |
|                          | %  | 57.0            | 55.0    |        |        | 57.5                       |                |        |        |                         | ± 6.0   |

Ref: TEMP-0830 Rev.41 FRONT / RECTO 130090307



**HORIBA ABX SAS**

HORIBA ABX SAS  
 Parc Euromédecine  
 Rue du Caducée - BP 7290  
 34184 MONTPELLIER Cedex 4  
 FRANCE  
 Tél.: 33 (0) 4 67 14 15 16  
 Fax.: 33 (0) 4 67 14 15 17

## ABX Minotrol 16



**LOT**

**MX 430 H**

**CONTROL**

**H**



**2021-10-05**  
(YYYY - MM - DD)

Rev 1

| PARAMETRES<br>PARAMETERS | UNITES<br>UNITS  | ABX Lyse        |              |        |        |                            |                |        |        | TOLERANCES<br>TOLERANCE |         |
|--------------------------|--|-----------------|--------------|--------|--------|----------------------------|----------------|--------|--------|-------------------------|---------|
|                          |  | MICROS          | MICROS       | MICROS | MICROS | MICROS                     | ARGOS - HELIOS | PENTRA | PENTRA |                         | PENTRA  |
|                          |  | ES60<br>Care ST | CRP 200      |        |        | 60 8P - 60 18P<br>ADVIA 60 |                |        |        |                         |         |
| GB WBC                   | 10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l  | <b>20.7</b>     | <b>20.3</b>  |        |        | <b>21.0</b>                |                |        |        |                         | ± 1.6   |
| GR RBC                   | 10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l | <b>5.61</b>     | <b>5.55</b>  |        |        | <b>5.64</b>                |                |        |        |                         | ± 0.20  |
| HB HGB                   | g/dl   | <b>17.8</b>     | <b>17.7</b>  |        |        | <b>18.0</b>                |                |        |        |                         | ± 0.6   |
|                          | g/l  | <b>178</b>      | <b>177</b>   |        |        | <b>180</b>                 |                |        |        |                         | ± 6     |
|                          | mmol/l   | <b>11.05</b>    | <b>10.99</b> |        |        | <b>11.18</b>               |                |        |        |                         | ± 0.37  |
| HT HCT                   | %  | <b>49.9</b>     | <b>49.4</b>  |        |        | <b>50.2</b>                |                |        |        |                         | ± 3.0   |
|                          | l/l  | <b>0.499</b>    | <b>0.494</b> |        |        | <b>0.502</b>               |                |        |        |                         | ± 0.030 |
| VGM MCV                  | µm <sup>3</sup> ; fl                                   | <b>89</b>       | <b>89</b>    |        |        | <b>89</b>                  |                |        |        |                         | ± 4     |
| TGMH MCH                 | pg   | <b>31.7</b>     | <b>31.9</b>  |        |        | <b>31.9</b>                |                |        |        |                         | ± 2.0   |
|                          | fmol   | <b>1.97</b>     | <b>1.98</b>  |        |        | <b>1.98</b>                |                |        |        |                         | ± 0.12  |
| CCMH MCHC                | g/dl   | <b>35.7</b>     | <b>35.8</b>  |        |        | <b>35.9</b>                |                |        |        |                         | ± 3.0   |
|                          | g/l  | <b>357</b>      | <b>358</b>   |        |        | <b>359</b>                 |                |        |        |                         | ± 30    |
|                          | mmol/l   | <b>22.14</b>    | <b>22.25</b> |        |        | <b>22.27</b>               |                |        |        |                         | ± 1.86  |
| IDR RDW                  | %  | <b>13.5</b>     | <b>13.5</b>  |        |        | <b>13.0</b>                |                |        |        |                         | ± 3.0   |
| PLAQ. PLTS               | 10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l  | <b>501</b>      | <b>490</b>   |        |        | <b>500</b>                 |                |        |        |                         | ± 45    |
| VPM MPV                  | µm <sup>3</sup> ; fl                                   | <b>8.0</b>      | <b>8.0</b>   |        |        | <b>7.5</b>                 |                |        |        |                         | ± 2.0   |
| LYMPHO                   | #  | <b>3.52</b>     | <b>3.45</b>  |        |        | <b>3.36</b>                |                |        |        |                         | ± 1.00  |
|                          | %  | <b>17.0</b>     | <b>17.0</b>  |        |        | <b>16.0</b>                |                |        |        |                         | ± 5.0   |
| MONO                     | #  | <b>1.04</b>     | <b>1.32</b>  |        |        | <b>0.95</b>                |                |        |        |                         | ± 0.80  |
|                          | %  | <b>5.0</b>      | <b>6.5</b>   |        |        | <b>4.5</b>                 |                |        |        |                         | ± 4.0   |
| GRANULO                  | #  | <b>16.15</b>    | <b>15.53</b> |        |        | <b>16.70</b>               |                |        |        |                         | ± 1.40  |
|                          | %  | <b>78.0</b>     | <b>76.5</b>  |        |        | <b>79.5</b>                |                |        |        |                         | ± 7.0   |