

**LOT**

**MX 422 L**  
 Rev 1

**CONTROL L**

(Exp.)

**2020-06-05**  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.0	2.0	2.0	2.0	2.0					± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.35	2.35	2.35	2.34	2.34					± 0.15
HB HGB	g/dl	6.1	5.9	6.0	6.1	6.1					± 0.4
	g/l	61	59	60	61	61					± 4
	mmol/l	3.79	3.66	3.73	3.79	3.79					± 0.25
HT HCT	%	16.5	16.5	16.5	16.6	16.4					± 2.0
	l/l	0.165	0.165	0.165	0.166	0.164					± 0.020
VGM MCV	µm <sup>3</sup> ; fl	70	70	70	71	70					± 4
TGMH MCH	pg	26.0	25.1	25.5	26.1	26.1					± 2.0
	fmol	1.61	1.56	1.59	1.62	1.62					± 0.12
CCMH MCHC	g/dl	37.1	35.9	36.5	36.7	37.2					± 3.0
	g/l	371	359	365	367	372					± 30
	mmol/l	23.03	22.27	22.65	22.80	23.13					± 1.86
IDR RDW	%	14.0	14.0	N/A	14.5	13.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	76	73	69	75	76					± 20
VPM MPV	µm <sup>3</sup> ; fl	8.4	8.7	N/A	8.4	8.0					± 2.0
LYMPHO	#	1.29	1.30	N/A	0.97	1.28					± 0.40
	%	64.5	65.0	N/A	48.5	64.0					± 8.0
MONO	#	0.20	0.22	N/A	0.37	0.21					± 0.20
	%	10.0	11.0	N/A	18.5	10.5					± 6.0
GRANULO	#	0.51	0.48	N/A	0.66	0.51					± 0.40
	%	25.5	24.0	N/A	33.0	25.5					± 7.0

**LOT** MX 422 N  
 Rev 1

**CONTROL** N

(Exp.)

**2020-06-05**  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.6	7.5	7.5	7.7	7.8					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.65	4.68	4.67	4.65	4.68					± 0.18
HB HGB	g/dl	13.3	13.2	13.2	13.4	13.4					± 0.5
	g/l	133	132	132	134	134					± 5
	mmol/l	8.26	8.20	8.20	8.32	8.32					± 0.31
HT HCT	%	36.7	37.0	36.9	36.7	37.0					± 2.5
	l/l	0.367	0.370	0.369	0.367	0.370					± 0.025
VGM MCV	µm <sup>3</sup> ; fl	79	79	79	79	79					± 4
TGMH MCH	pg	28.6	28.2	28.3	28.8	28.6					± 2.0
	fmol	1.78	1.75	1.76	1.79	1.78					± 0.12
CCMH MCHC	g/dl	36.2	35.7	35.8	36.5	36.2					± 3.0
	g/l	362	357	358	365	362					± 30
	mmol/l	22.48	22.17	22.22	22.65	22.51					± 1.86
IDR RDW	%	14.0	13.5	N/A	14.0	13.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	265	260	260	260	265					± 40
VPM MPV	µm <sup>3</sup> ; fl	7.7	7.8	N/A	8.2	7.4					± 2.0
LYMPHO	#	2.62	2.66	N/A	2.08	2.65					± 0.40
	%	34.5	35.5	N/A	27.0	34.0					± 5.0
MONO	#	0.46	0.49	N/A	0.73	0.51					± 0.30
	%	6.0	6.5	N/A	9.5	6.5					± 4.0
GRANULO	#	4.52	4.35	N/A	4.89	4.64					± 0.60
	%	59.5	58.0	N/A	63.5	59.5					± 6.0

Ref: TEMP-0830 Rev.39 FRONT / RECTO 130060513



**HORIBA ABX SAS**

HORIBA ABX SAS  
 Parc Euromédecine  
 Rue du Caducée - BP 7290  
 34184 MONTPELLIER Cedex 4  
 FRANCE  
 Tél.: 33 (0) 4 67 14 15 16  
 Fax.: 33 (0) 4 67 14 15 17

## ABX Minotrol 16



**LOT**

**MX 422 H**  
 Rev 1

**CONTROL H**

(Exp.)

**2020-06-05**  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>20.8</b>	<b>20.3</b>	<b>20.1</b>	<b>20.5</b>	<b>21.0</b>					± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>5.72</b>	<b>5.67</b>	<b>5.67</b>	<b>5.70</b>	<b>5.69</b>					± 0.20
HB HGB	g/dl	<b>18.1</b>	<b>17.9</b>	<b>17.9</b>	<b>18.2</b>	<b>18.2</b>					± 0.6
	g/l	<b>181</b>	<b>179</b>	<b>179</b>	<b>182</b>	<b>182</b>					± 6
	mmol/l	<b>11.24</b>	<b>11.12</b>	<b>11.12</b>	<b>11.30</b>	<b>11.30</b>					± 0.37
HT HCT	%	<b>50.3</b>	<b>49.9</b>	<b>49.9</b>	<b>50.2</b>	<b>50.6</b>					± 3.0
	l/l	<b>0.503</b>	<b>0.499</b>	<b>0.499</b>	<b>0.502</b>	<b>0.506</b>					± 0.030
VGM MCV	µm <sup>3</sup> ; fl	<b>88</b>	<b>88</b>	<b>88</b>	<b>88</b>	<b>89</b>					± 4
TGMH MCH	pg	<b>31.6</b>	<b>31.6</b>	<b>31.6</b>	<b>31.9</b>	<b>32.0</b>					± 2.0
	fmol	<b>1.97</b>	<b>1.96</b>	<b>1.96</b>	<b>1.98</b>	<b>1.99</b>					± 0.12
CCMH MCHC	g/dl	<b>36.0</b>	<b>35.9</b>	<b>35.9</b>	<b>36.3</b>	<b>35.9</b>					± 3.0
	g/l	<b>360</b>	<b>359</b>	<b>359</b>	<b>363</b>	<b>359</b>					± 30
	mmol/l	<b>22.33</b>	<b>22.28</b>	<b>22.28</b>	<b>22.53</b>	<b>22.32</b>					± 1.86
IDR RDW	%	<b>13.0</b>	<b>12.5</b>	N/A	<b>12.5</b>	<b>12.0</b>					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>530</b>	<b>510</b>	<b>530</b>	<b>530</b>	<b>525</b>					± 45
VPM MPV	µm <sup>3</sup> ; fl	<b>8.0</b>	<b>8.2</b>	N/A	<b>8.7</b>	<b>7.6</b>					± 2.0
LYMPHO	#	<b>3.33</b>	<b>3.35</b>	N/A	<b>2.67</b>	<b>3.26</b>					± 1.00
	%	<b>16.0</b>	<b>16.5</b>	N/A	<b>13.0</b>	<b>15.5</b>					± 5.0
MONO	#	<b>0.83</b>	<b>0.91</b>	N/A	<b>0.82</b>	<b>0.84</b>					± 0.80
	%	<b>4.0</b>	<b>4.5</b>	N/A	<b>4.0</b>	<b>4.0</b>					± 4.0
GRANULO	#	<b>16.64</b>	<b>16.04</b>	N/A	<b>17.02</b>	<b>16.91</b>					± 1.40
	%	<b>80.0</b>	<b>79.0</b>	N/A	<b>83.0</b>	<b>80.5</b>					± 7.0