

LOT		MX 404 L		CONTROL L						(Exp.)	2017-06-05 (YYYY-MM-DD)
Rev 1		ABX Lyse									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0	2.1	2.1	2.0	2.1					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.40	2.40	2.42	2.41	2.42					± 0.15
HB HGB	g/dl	6.2	6.3	6.2	6.1	6.2					± 0.4
	g/l	62	63	62	61	62					± 4
	mmol/l	3.85	3.91	3.85	3.79	3.85					± 0.25
HT HCT	%	16.6	16.6	16.9	16.6	16.7					± 2.0
	l/l	0.166	0.166	0.169	0.166	0.167					± 0.020
VGM MCV	µm ³ ; fl	69	69	70	69	69					± 4
TGMH MCH	pg	25.8	26.3	25.6	25.3	25.6					± 2.0
	fmol	1.60	1.63	1.59	1.57	1.59					± 0.12
CCMH MCHC	g/dl	37.4	38.0	36.6	36.7	37.1					± 3.0
	g/l	374	380	366	367	371					± 30
	mmol/l	23.25	23.63	22.73	22.78	23.06					± 1.86
IDR RDW	%	13.9	13.9	N/A	14.5	12.8					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	80	77	71	77	77					± 20
VPM MPV	µm ³ ; fl	8.8	9.0	N/A	8.6	8.3					± 2.0
LYMPHO	#	1.31	1.38	N/A	1.11	1.34					± 0.40
	%	65.5	65.5	N/A	55.5	64.0					± 8.0
MONO	#	0.20	0.20	N/A	0.24	0.20					± 0.20
	%	8.5	9.0	N/A	12.0	9.5					± 6.0
GRANULO	#	0.52	0.54	N/A	0.65	0.56					± 0.40
	%	26.0	25.5	N/A	32.5	26.5					± 7.0

LOT

MX 404 N
 Rev 1

CONTROL N

(Exp.)

2017-06-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.5	7.5	7.5	7.4	7.6					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.75	4.73	4.73	4.77	4.77					± 0.18
HB HGB	g/dl	13.4	13.4	13.3	13.4	13.4					± 0.5
	g/l	134	134	133	134	134					± 5
	mmol/l	8.32	8.32	8.26	8.32	8.32					± 0.31
HT HCT	%	36.6	36.4	36.9	36.7	36.7					± 2.5
	l/l	0.366	0.364	0.369	0.367	0.367					± 0.025
VGM MCV	µm ³ ; fl	77	77	78	77	77					± 4
TGMH MCH	pg	28.2	28.3	28.1	28.1	28.1					± 2.0
	fmol	1.75	1.76	1.75	1.74	1.74					± 0.12
CCMH MCHC	g/dl	36.6	36.8	36.0	36.5	36.5					± 3.0
	g/l	366	368	360	365	365					± 30
	mmol/l	22.75	22.85	22.39	22.66	22.66					± 1.86
IDR RDW	%	14.0	13.6	N/A	13.8	13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	260	263	260	268	263					± 40
VPM MPV	µm ³ ; fl	8.7	8.8	N/A	8.8	8.0					± 2.0
LYMPHO	#	2.63	2.63	N/A	2.35	2.66					± 0.40
	%	35.0	35.0	N/A	31.7	35.0					± 5.0
MONO	#	0.53	0.53	N/A	0.41	0.52					± 0.30
	%	7.0	7.0	N/A	5.5	6.8					± 4.0
GRANULO	#	4.35	4.35	N/A	4.65	4.42					± 0.60
	%	58.0	58.0	N/A	62.8	58.2					± 6.0

Ref: TEMP-0830 Rev.36 FRONT / RECTO 1219930084-B

LOT		MX 404 H		CONTROL H						(Exp.)	2017-06-05 (YYYY-MM-DD)
Rev 1		ABX Lyse									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.3	20.2	19.8	19.8	20.3					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.67	5.63	5.57	5.64	5.68					± 0.20
HB HGB	g/dl	17.8	17.8	17.5	17.8	17.8					± 0.6
	g/l	178	178	175	178	178					± 6
	mmol/l	11.05	11.05	10.87	11.05	11.05					± 0.37
HT HCT	%	49.9	49.5	49.6	49.6	50.0					± 3.0
	l/l	0.499	0.495	0.496	0.496	0.500					± 0.030
VGM MCV	µm ³ ; fl	88	88	89	88	88					± 4
TGMH MCH	pg	31.4	31.6	31.4	31.6	31.3					± 2.0
	fmol	1.95	1.96	1.95	1.96	1.95					± 0.12
CCMH MCHC	g/dl	35.7	35.9	35.3	35.9	35.6					± 3.0
	g/l	357	359	353	359	356					± 30
	mmol/l	22.15	22.31	21.92	22.27	22.11					± 1.86
IDR RDW	%	13.4	13.2	N/A	13.0	12.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	510	515	524	525	518					± 45
VPM MPV	µm ³ ; fl	8.0	8.0	N/A	8.2	7.5					± 2.0
LYMPHO	#	3.45	3.23	N/A	2.77	3.21					± 1.00
	%	17.0	16.0	N/A	14.0	15.8					± 5.0
MONO	#	1.12	1.11	N/A	0.80	1.02					± 0.80
	%	5.5	5.5	N/A	4.0	5.0					± 4.0
GRANULO	#	15.73	15.86	N/A	16.24	16.08					± 1.40
	%	77.5	78.5	N/A	82.0	79.2					± 7.0