

LOT		MX 428 L		CONTROL L						(Exp.) 2021-06-05 (YYYY - MM - DD)	
Rev 1		ABX Lyse									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.1	2.0			2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.38	2.34			2.38					± 0.15
HB HGB	g/dl	6.0	5.9			6.1					± 0.4
	g/l	69	59			61					± 4
	mmol/l	3.73	3.66			3.79					± 0.25
HT HCT	%	16.4	15.9			16.4					± 2.0
	l/l	0.164	0.159			0.164					± 0.020
VGM MCV	µm ³ ; fl	69	68			69					± 4
TGMH MCH	pg	25.2	25.2			25.6					± 2.0
	fmol	1.57	1.57			1.59					± 0.12
CCMH MCHC	g/dl	36.5	37.1			37.1					± 3.0
	g/l	365	371			371					± 30
	mmol/l	22.69	23.03			23.07					± 1.86
IDR RDW	%	14.5	14.5			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	77	81			77					± 20
VPM MPV	µm ³ ; fl	9.1	8.6			8.1					± 2.0
LYMPHO	#	1.40	1.40			1.34					± 0.40
	%	66.5	70.0			67.0					± 8.0
MONO	#	0.20	0.20			0.20					± 0.20
	%	8.5	9.0			8.5					± 6.0
GRANULO	#	0.53	0.42			0.49					± 0.40
	%	25.0	21.0			24.5					± 7.0

LOT MX 428 N
 Rev 1

CONTROL N

(Exp.)

2021-06-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.7	7.5			7.7					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.65	4.60			4.61					± 0.18
HB HGB	g/dl	13.4	13.2			13.5					± 0.5
	g/l	134	132			135					± 5
	mmol/l	8.32	8.20			8.38					± 0.31
HT HCT	%	37.2	36.8			37.3					± 2.5
	l/l	0.372	0.368			0.373					± 0.025
VGM MCV	µm ³ ; fl	80	80			81					± 4
TGMH MCH	pg	28.8	28.7			29.3					± 2.0
	fmol	1.79	1.78			1.82					± 0.12
CCMH MCHC	g/dl	36.0	35.9			36.2					± 3.0
	g/l	360	359			362					± 30
	mmol/l	22.37	22.28			22.45					± 1.86
IDR RDW	%	13.7	13.8			13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	252	248			250					± 40
VPM MPV	µm ³ ; fl	8.2	8.1			7.5					± 2.0
LYMPHO	#	2.66	2.63			2.62					± 0.40
	%	34.5	35.0			34.0					± 5.0
MONO	#	0.54	0.60			0.58					± 0.30
	%	7.0	8.0			7.5					± 4.0
GRANULO	#	4.50	4.28			4.50					± 0.60
	%	58.5	57.0			58.5					± 6.0

Ref: TEMP-0830 Rev.39 FRONT / RECTO 130060513



LOT

MX 428 H
 Rev 1

CONTROL H

(Exp.)

2021-06-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.8	19.8			20.5					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.68	5.55			5.63					± 0.20
HB HGB	g/dl	17.8	17.5			17.9					± 0.6
	g/l	178	175			179					± 6
	mmol/l	11.05	10.87			11.12					± 0.37
HT HCT	%	50.0	48.8			50.1					± 3.0
	l/l	0.500	0.488			0.501					± 0.030
VGM MCV	µm ³ ; fl	88	88			89					± 4
TGMH MCH	pg	31.3	31.5			31.8					± 2.0
	fmol	1.95	1.96			1.97					± 0.12
CCMH MCHC	g/dl	35.6	35.8			35.7					± 3.0
	g/l	356	358			357					± 30
	mmol/l	22.11	22.25			22.18					± 1.86
IDR RDW	%	13.2	13.2			12.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	498	490			493					± 45
VPM MPV	µm ³ ; fl	8.1	8.1			7.5					± 2.0
LYMPHO	#	3.43	3.17			3.28					± 1.00
	%	16.5	16.0			16.0					± 5.0
MONO	#	0.94	1.09			0.92					± 0.80
	%	4.5	5.5			4.5					± 4.0
GRANULO	#	16.43	15.54			16.30					± 1.40
	%	79.0	78.5			79.5					± 7.0