

**LOT** MX 036 L  
Rev 1

**CONTROL** L

 (Exp.)

2016-06-05  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.0	2.1	2.1	2.0	2.1					± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.44	2.44	2.46	2.43	2.42					± 0.15
HB HGB	g/dl	6.3	6.2	6.2	6.1	6.2					± 0.4
	g/l	63	62	62	61	62					± 4
	mmol/l	3.91	3.85	3.85	3.79	3.85					± 0.25
HT HCT	%	17.1	17.1	17.2	17.0	16.9					± 2.0
	l/l	0.171	0.171	0.172	0.170	0.169					± 0.020
VGM MCV	µm <sup>3</sup> ; fl	70	70	70	70	70					± 4
TGMH MCH	pg	25.8	25.4	25.2	25.1	25.6					± 2.0
	fmol	1.60	1.58	1.57	1.56	1.59					± 0.12
CCMH MCHC	g/dl	36.9	36.3	36.0	35.9	36.6					± 3.0
	g/l	369	363	360	359	366					± 30
	mmol/l	22.91	22.54	22.36	22.27	22.73					± 1.86
IDR RDW	%	14.5	14.0	N/A	14.4	14.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	80	75	70	77	77					± 20
VPM MPV	µm <sup>3</sup> ; fl	9.2	9.0	N/A	9.0	8.2					± 2.0
LYMPHO	#	1.26	1.34	N/A	1.04	1.33					± 0.40
	%	63.0	64.0	N/A	52.0	63.5					± 8.0
MONO	#	0.21	0.23	N/A	0.31	0.21					± 0.20
	%	10.5	11.0	N/A	15.5	10.0					± 6.0
GRANULO	#	0.53	0.53	N/A	0.65	0.56					± 0.40
	%	26.5	25.0	N/A	32.5	26.5					± 7.0

**LOT** MX 036 N  
Rev 1

**CONTROL** N

 (Exp.)

**2016-06-05**  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.5	7.5	7.3	7.2	7.5					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.66	4.66	4.67	4.64	4.64					± 0.18
HB HGB	g/dl	13.6	13.4	13.5	13.5	13.5					± 0.5
	g/l	136	134	135	135	135					± 5
	mmol/l	8.45	8.32	8.38	8.38	8.38					± 0.31
HT HCT	%	37.3	37.3	37.8	37.6	37.6					± 2.5
	l/l	0.373	0.373	0.378	0.376	0.376					± 0.025
VGM MCV	µm <sup>3</sup> ; fl	80	80	81	81	81					± 4
TGMH MCH	pg	29.2	28.8	28.9	29.1	29.1					± 2.0
	fmol	1.81	1.79	1.80	1.81	1.81					± 0.12
CCMH MCHC	g/dl	36.5	35.9	35.7	35.9	35.9					± 3.0
	g/l	365	359	357	359	359					± 30
	mmol/l	22.65	22.32	22.16	22.31	22.31					± 1.86
IDR RDW	%	14.0	13.5	N/A	14.0	13.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	261	250	250	250	255					± 40
VPM MPV	µm <sup>3</sup> ; fl	8.4	8.0	N/A	8.3	7.7					± 2.0
LYMPHO	#	2.44	2.33	N/A	1.93	2.40					± 0.40
	%	32.5	31.0	N/A	26.8	32.0					± 5.0
MONO	#	0.56	0.60	N/A	0.56	0.53					± 0.30
	%	7.5	8.0	N/A	7.8	7.0					± 4.0
GRANULO	#	4.50	4.58	N/A	4.71	4.58					± 0.60
	%	60.0	61.0	N/A	65.4	61.0					± 6.0

Ref: TEMP-0830 Rev.37 FRONT / RECTO 1219930084-B

LOT		MX 036 H		CONTROL H						(Exp.)	2016-06-05 (YYYY-MM-DD)
Rev 1		ABX Lyse									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	19.6	19.7	19.3	19.3	19.8					± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	5.66	5.66	5.66	5.64	5.63					± 0.20
HB HGB	g/dl	18.0	17.8	17.9	18.0	17.9					± 0.6
	g/l	180	178	179	180	179					± 6
	mmol/l	11.18	11.05	11.12	11.18	11.12					± 0.37
HT HCT	%	49.8	49.8	50.4	50.2	50.1					± 3.0
	l/l	0.498	0.498	0.504	0.502	0.501					± 0.030
VGM MCV	µm <sup>3</sup> ; fl	88	88	89	89	89					± 4
TGMH MCH	pg	31.8	31.4	31.6	31.9	31.8					± 2.0
	fmol	1.97	1.95	1.96	1.98	1.97					± 0.12
CCMH MCHC	g/dl	36.1	35.7	35.5	35.9	35.7					± 3.0
	g/l	361	357	355	359	357					± 30
	mmol/l	22.44	22.19	22.07	22.27	22.18					± 1.86
IDR RDW	%	13.5	13.0	N/A	13.3	13.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	524	515	525	515	510					± 45
VPM MPV	µm <sup>3</sup> ; fl	8.0	7.8	N/A	7.8	7.5					± 2.0
LYMPHO	#	3.04	2.96	N/A	2.41	3.07					± 1.00
	%	15.5	15.0	N/A	12.5	15.5					± 5.0
MONO	#	0.98	0.99	N/A	0.77	0.89					± 0.80
	%	5.0	5.0	N/A	4.0	4.5					± 4.0
GRANULO	#	15.58	15.76	N/A	16.12	15.84					± 1.40
	%	79.5	80.0	N/A	83.5	80.0					± 7.0