






LOT		MX 403 L		CONTROL		L		 (Exp.)		2017-04-05 (YYYY - MM - DD)			
		Rev 1		LyseVet									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS		MICROS							TOLERANCES TOLERANCE		
		Scil Vet ABC	Scil Vet ABC Plus ESV 60										
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	1.6	1.6								± 0.4		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.41	2.40								± 0.15		
HB HGB	g/dl	6.1	6.1								± 0.4		
	g/l	61	61								± 4		
	mmol/l	3.79	3.79								± 0.25		
HT HCT	%	16.9	16.6								± 2.0		
	l/l	0.169	0.166								± 0.020		
VGM MCV	µm ³ ; fl	70	69								± 4		
TGMH MCH	pg	25.3	25.4								± 2.0		
	fmol	1.57	1.58								± 0.12		
CCMH MCHC	g/dl	36.2	36.8								± 3.0		
	g/l	362	368								± 30		
	mmol/l	22.45	22.88								± 1.86		
IDR RDW	%	N/A	14.0								± 3.0		
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	63	75								± 20		
VPM MPV	µm ³ ; fl	N/A	8.5								± 2.0		
LYMPHO	#	N/A	0.80								± 0.40		
	%	N/A	50.0								± 9.0		
MONO	#	N/A	0.20								± 0.20		
	%	N/A	11.0								± 6.0		
GRANULO	#	N/A	0.62								± 0.40		
	%	N/A	39.0								± 8.0		



LOT		MX 403 N		CONTROL		N		 (Exp.)				2017-04-05 (YYYY - MM - DD)			
		Rev 1		LyseVet											
PARAMETRES PARAMETERS		UNITES UNITS		MICROS		MICROS								TOLERANCES TOLERANCE	
				Scil Vet ABC	Scil Vet ABC Plus ESV 60										
GB	WBC	10 ³ /mm ³ ; 10 ⁹ /l	6.7	6.8										± 0.8	
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.66	4.68										± 0.18	
HB	HGB	g/dl	13.6	13.6										± 0.5	
		g/l	136	136										± 5	
		mmol/l	8.45	8.45										± 0.31	
HT	HCT	%	37.3	37.4										± 2.5	
		l/l	0.373	0.374										± 0.025	
VGM	MCV	µm ³ ; fl	80	80										± 4	
TGMH	MCH	pg	29.2	29.1										± 2.0	
		fmol	1.81	1.80										± 0.12	
CCMH	MCHC	g/dl	36.5	36.3										± 3.0	
		g/l	365	363										± 30	
		mmol/l	22.65	22.56										± 1.86	
IDR	RDW	%	N/A	13.5										± 3.0	
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	240	250										± 40	
VPM	MPV	µm ³ ; fl	N/A	8.6										± 2.0	
LYMPHO	#	N/A	1.73											± 0.50	
	%	N/A	25.5											± 6.0	
MONO	#	N/A	0.54											± 0.30	
	%	N/A	8.0											± 4.0	
GRANULO	#	N/A	4.52											± 0.60	
	%	N/A	66.5											± 7.0	

1300010912-A
 Ref: TEMP-0830 Rev.38



LOT		MX 403 H		CONTROL H		 (Exp.)		2017-04-05 (YYYY - MM - DD)		
Rev 1										
PARAMETRES PARAMETERS	UNITES UNITS	LyseVet								TOLERANCES TOLERANCE
		MICROS		MICROS						
		Scil Vet ABC	Scil Vet ABC Plus ESV 60							
GB WBC	10 ³ /mm ³ ; 10 ⁹ /l	18.7	19.0							± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.58	5.62							± 0.20
HB HGB	g/dl	17.9	18.0							± 0.6
	g/l	179	180							± 6
	mmol/l	11.12	11.18							± 0.37
HT HCT	%	49.1	49.5							± 3.0
	l/l	0.491	0.495							± 0.030
VGM MCV	µm ³ ; fl	88	88							± 4
TGMH MCH	pg	32.1	32.0							± 2.0
	fmol	1.99	1.99							± 0.12
CCMH MCHC	g/dl	36.5	36.4							± 3.0
	g/l	365	364							± 30
	mmol/l	22.64	22.60							± 1.86
IDR RDW	%	N/A	13.0							± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	500	500							± 45
VPM MPV	µm ³ ; fl	N/A	8.6							± 2.0
LYMPHO	#	N/A	2.00							± 1.20
	%	N/A	10.5							± 6.0
MONO	#	N/A	0.86							± 0.80
	%	N/A	4.5							± 4.0
GRANULO	#	N/A	16.15							± 1.60
	%	N/A	85.0							± 8.0