



<b>LOT</b>		<b>MX 422 L</b>	<b>CONTROL</b>		<b>L</b>					 (Exp.)	<b>2020-06-05</b> (YYYY - MM - DD)	
		Rev 1	<b>LyseVet</b>									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS									TOLERANCES TOLERANCE	
		Sci Vet ABC	Sci Vet ABC Plus ESV 60									
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	1.8	1.9								± 0.4	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.38	2.37								± 0.15	
HB HGB	g/dl	6.0	6.0								± 0.4	
	g/l	60	60								± 4	
	mmol/l	3.73	3.73								± 0.25	
HT HCT	%	16.7	16.6								± 2.0	
	l/l	0.167	0.166								± 0.020	
VGM MCV	µm <sup>3</sup> ; fl	70	70								± 4	
TGMH MCH	pg	25.2	25.3								± 2.0	
	fmol	1.57	1.57								± 0.12	
CCMH MCHC	g/dl	36.0	36.2								± 3.0	
	g/l	360	362								± 30	
	mmol/l	22.36	22.46								± 1.86	
IDR RDW	%	N/A	15.0								± 3.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	67	76								± 20	
VPM MPV	µm <sup>3</sup> ; fl	N/A	8.9								± 2.0	
LYMPHO	#	N/A	1.05								± 0.40	
	%	N/A	55.0								± 8.0	
MONO	#	N/A	0.24								± 0.20	
	%	N/A	12.5								± 6.0	
GRANULO	#	N/A	0.62								± 0.40	
	%	N/A	32.5								± 7.0	

<b>LOT</b>		<b>MX 422 N</b>	<b>CONTROL N</b>							 (Exp.)	<b>2020-06-05</b> (YYYY-MM-DD)
		Rev 1	<b>LyseVet</b>								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS								TOLERANCES TOLERANCE	
		Scil Vet ABC	Scil Vet ABC Plus ESV 60								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.3	7.4							± 0.8	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.68	4.72							± 0.18	
HB HGB	g/dl	13.2	13.3							± 0.5	
	g/l	132	133							± 5	
	mmol/l	8.20	8.26							± 0.31	
HT HCT	%	36.5	37.3							± 2.5	
	l/l	0.365	0.373							± 0.025	
VGM MCV	µm <sup>3</sup> ; fl	78	79							± 4	
TGMH MCH	pg	28.2	28.2							± 2.0	
	fmol	1.75	1.75							± 0.12	
CCMH MCHC	g/dl	36.2	35.7							± 3.0	
	g/l	362	357							± 30	
	mmol/l	22.46	22.15							± 1.86	
IDR RDW	%	N/A	14.5							± 3.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	248	265							± 40	
VPM MPV	µm <sup>3</sup> ; fl	N/A	8.6							± 2.0	
LYMPHO	#	N/A	2.00							± 0.40	
	%	N/A	27.0							± 5.0	
MONO	#	N/A	0.67							± 0.30	
	%	N/A	9.0							± 4.0	
GRANULO	#	N/A	4.74							± 0.60	
	%	N/A	64.0							± 6.0	

Ref: TEMP-0830 Rev.39 BACK / VERSO 1300060626

LOT

MX 422 H  
Rev 1

CONTROL

H

 (Exp.)

2020-06-05  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	LyseVet								TOLERANCES TOLERANCE
		MICROS		MICROS						
		Scil Vet ABC	Scil Vet ABC Plus ESV 60							
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	20.1	20.4							± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	5.69	5.75							± 0.20
HB HGB	g/dl	18.2	18.2							± 0.6
	g/l	182	182							± 6
	mmol/l	11.30	11.30							± 0.37
HT HCT	%	50.1	50.6							± 3.0
	l/l	0.501	0.506							± 0.030
VGM MCV	µm <sup>3</sup> ; fl	88	88							± 4
TGMH MCH	pg	32.0	31.7							± 2.0
	fmol	1.99	1.97							± 0.12
CCMH MCHC	g/dl	36.3	36.0							± 3.0
	g/l	363	360							± 30
	mmol/l	22.57	22.34							± 1.86
IDR RDW	%	N/A	13.0							± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	505	537							± 45
VPM MPV	µm <sup>3</sup> ; fl	N/A	8.9							± 2.0
LYMPHO	#	N/A	2.35							± 1.00
	%	N/A	11.5							± 5.0
MONO	#	N/A	0.92							± 0.80
	%	N/A	4.5							± 4.0
GRANULO	#	N/A	17.14							± 1.40
	%	N/A	84.0							± 7.0