




<b>LOT</b>		<b>MX 411 L</b>	<b>CONTROL L</b>							 (Exp.)	<b>2018-08-05</b> (YYYY - MM - DD)
		Rev 1	<b>LyseVet</b>								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS									TOLERANCES TOLERANCE
		Scil Vet ABC	Scil Vet ABC Plus ESV 60								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	1.8	1.8								± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.25	2.25								± 0.15
HB HGB	g/dl	5.8	5.8								± 0.4
	g/l	58	58								± 4
	mmol/l	3.60	3.60								± 0.25
HT HCT	%	16.0	15.8								± 2.0
	l/l	0.160	0.158								± 0.020
VGM MCV	µm <sup>3</sup> ; fl	71	70								± 4
TGMH MCH	pg	25.8	25.8								± 2.0
	fmol	1.60	1.60								± 0.12
CCMH MCHC	g/dl	36.3	36.8								± 3.0
	g/l	363	368								± 30
	mmol/l	22.55	22.87								± 1.86
IDR RDW	%	N/A	13.5								± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	69	84								± 20
VPM MPV	µm <sup>3</sup> ; fl	N/A	8.4								± 2.0
LYMPHO	#	N/A	1.01								± 0.40
	%	N/A	56.0								± 8.0
MONO	#	N/A	0.20								± 0.20
	%	N/A	10.0								± 6.0
GRANULO	#	N/A	0.61								± 0.40
	%	N/A	34.0								± 7.0

<b>LOT</b>		<b>MX 411 N</b>	<b>CONTROL N</b>							 (Exp.)	<b>2018-08-05</b> (YYYY-MM-DD)
		Rev 1	<b>LyseVet</b>								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS								TOLERANCES TOLERANCE	
		Soil Vet ABC	Soil Vet ABC Plus ESV 60								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.1	7.2							± 0.8	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.64	4.66							± 0.18	
HB HGB	g/dl	13.3	13.3							± 0.5	
	g/l	133	133							± 5	
	mmol/l	8.26	8.26							± 0.31	
HT HCT	%	37.6	37.3							± 2.5	
	l/l	0.376	0.373							± 0.025	
VGM MCV	µm <sup>3</sup> ; fl	81	80							± 4	
TGMH MCH	pg	28.7	28.5							± 2.0	
	fmol	1.78	1.77							± 0.12	
CCMH MCHC	g/dl	35.4	35.7							± 3.0	
	g/l	354	357							± 30	
	mmol/l	21.98	22.15							± 1.86	
IDR RDW	%	N/A	13.5							± 3.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	250	258							± 40	
VPM MPV	µm <sup>3</sup> ; fl	N/A	8.5							± 2.0	
LYMPHO	#	N/A	1.94							± 0.40	
	%	N/A	27.0							± 5.0	
MONO	#	N/A	0.65							± 0.30	
	%	N/A	9.0							± 4.0	
GRANULO	#	N/A	4.61							± 0.60	
	%	N/A	64.0							± 6.0	

1300010912-A  
Ref: TEMP-0830 Rev.38

<b>LOT</b>		<b>MX 411 H</b>	<b>CONTROL H</b>							 (Exp.)	<b>2018-08-05</b> (YYYY-MM-DD)
		Rev 1	<b>LyseVet</b>								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS									TOLERANCES TOLERANCE
		Scil Vet ABC	Scil Vet ABC Plus ESV 60								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>19.6</b>	<b>19.5</b>								± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>5.81</b>	<b>5.83</b>								± 0.20
HB HGB	g/dl	<b>18.2</b>	<b>18.2</b>								± 0.6
	g/l	<b>182</b>	<b>182</b>								± 6
	mmol/l	<b>11.30</b>	<b>11.30</b>								± 0.37
HT HCT	%	<b>51.1</b>	<b>50.7</b>								± 3.0
	l/l	<b>0.511</b>	<b>0.507</b>								± 0.030
VGM MCV	µm <sup>3</sup> ; fl	<b>88</b>	<b>87</b>								± 4
TGMH MCH	pg	<b>31.3</b>	<b>31.2</b>								± 2.0
	fmol	<b>1.95</b>	<b>1.94</b>								± 0.12
CCMH MCHC	g/dl	<b>35.6</b>	<b>35.9</b>								± 3.0
	g/l	<b>356</b>	<b>359</b>								± 30
	mmol/l	<b>22.11</b>	<b>22.28</b>								± 1.86
IDR RDW	%	N/A	<b>13.0</b>								± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>478</b>	<b>485</b>								± 45
VPM MPV	µm <sup>3</sup> ; fl	N/A	<b>8.7</b>								± 2.0
LYMPHO	#	N/A	<b>2.34</b>								± 1.00
	%	N/A	<b>12.0</b>								± 5.0
MONO	#	N/A	<b>0.98</b>								± 0.80
	%	N/A	<b>5.0</b>								± 4.0
GRANULO	#	N/A	<b>16.19</b>								± 1.40
	%	N/A	<b>83.0</b>								± 7.0