




<b>LOT</b>		<b>MX 420 L</b>	<b>CONTROL</b>		<b>L</b>		 <b>(Exp.)</b>				<b>2020-02-05</b> (YYYY - MM - DD)	
		Rev 1	<b>LyseVet</b>									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS									TOLERANCES TOLERANCE	
		Scil Vet ABC	Scil Vet ABC Plus ESV 60									
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>1.8</b>	<b>1.8</b>								± 0.4	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>2.36</b>	<b>2.36</b>								± 0.15	
HB HGB	g/dl	<b>5.9</b>	<b>5.9</b>								± 0.4	
	g/l	<b>59</b>	<b>59</b>								± 4	
	mmol/l	<b>3.66</b>	<b>3.66</b>								± 0.25	
HT HCT	%	<b>16.5</b>	<b>16.3</b>								± 2.0	
	l/l	<b>0.165</b>	<b>0.163</b>								± 0.020	
VGM MCV	µm <sup>3</sup> ; fl	<b>70</b>	<b>69</b>								± 4	
TGMH MCH	pg	<b>25.0</b>	<b>25.0</b>								± 2.0	
	fmol	<b>1.55</b>	<b>1.55</b>								± 0.12	
CCMH MCHC	g/dl	<b>35.7</b>	<b>36.2</b>								± 3.0	
	g/l	<b>357</b>	<b>362</b>								± 30	
	mmol/l	<b>22.18</b>	<b>22.50</b>								± 1.86	
IDR RDW	%	N/A	<b>14.5</b>								± 3.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>66</b>	<b>80</b>								± 20	
VPM MPV	µm <sup>3</sup> ; fl	N/A	<b>9.0</b>								± 2.0	
LYMPHO	#	N/A	<b>0.97</b>								± 0.40	
	%	N/A	<b>54.0</b>								± 8.0	
MONO	#	N/A	<b>0.20</b>								± 0.20	
	%	N/A	<b>10.5</b>								± 6.0	
GRANULO	#	N/A	<b>0.64</b>								± 0.40	
	%	N/A	<b>35.5</b>								± 7.0	

<b>LOT</b>		<b>MX 420 N</b>	<b>CONTROL N</b>							 (Exp.)	<b>2020-02-05</b> (YYYY-MM-DD)
		Rev 1	<b>LyseVet</b>								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS								TOLERANCES TOLERANCE	
		Scil Vet ABC	Scil Vet ABC Plus ESV 60								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>7.0</b>	<b>7.2</b>							± 0.8	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>4.66</b>	<b>4.68</b>							± 0.18	
HB HGB	g/dl	<b>13.5</b>	<b>13.4</b>							± 0.5	
	g/l	<b>135</b>	<b>134</b>							± 5	
	mmol/l	<b>8.38</b>	<b>8.32</b>							± 0.31	
HT HCT	%	<b>37.3</b>	<b>37.0</b>							± 2.5	
	l/l	<b>0.373</b>	<b>0.370</b>							± 0.025	
VGM MCV	µm <sup>3</sup> ; fl	<b>80</b>	<b>79</b>							± 4	
TGMH MCH	pg	<b>29.0</b>	<b>28.6</b>							± 2.0	
	fmol	<b>1.80</b>	<b>1.78</b>							± 0.12	
CCMH MCHC	g/dl	<b>36.2</b>	<b>36.2</b>							± 3.0	
	g/l	<b>362</b>	<b>362</b>							± 30	
	mmol/l	<b>22.49</b>	<b>22.51</b>							± 1.86	
IDR RDW	%	N/A	<b>14.0</b>							± 3.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>240</b>	<b>262</b>							± 40	
VPM MPV	µm <sup>3</sup> ; fl	N/A	<b>8.5</b>							± 2.0	
LYMPHO	#	N/A	<b>1.94</b>							± 0.40	
	%	N/A	<b>27.0</b>							± 5.0	
MONO	#	N/A	<b>0.58</b>							± 0.30	
	%	N/A	<b>8.0</b>							± 4.0	
GRANULO	#	N/A	<b>4.68</b>							± 0.60	
	%	N/A	<b>65.0</b>							± 6.0	

Ref: TEMP-0830 Rev.39 BACK / VERSO 1300060626

<b>LOT</b>		<b>MX 420 H</b>	<b>CONTROL H</b>							 (Exp.)	<b>2020-02-05</b> (YYYY - MM - DD)
		Rev 1	<b>LyseVet</b>								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS								TOLERANCES TOLERANCE	
		Scil Vet ABC	Scil Vet ABC Plus ESV 60								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>19.3</b>	<b>19.9</b>							± 1.6	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>5.69</b>	<b>5.69</b>							± 0.20	
HB HGB	g/dl	<b>18.1</b>	<b>18.2</b>							± 0.6	
	g/l	<b>181</b>	<b>182</b>							± 6	
	mmol/l	<b>11.24</b>	<b>11.30</b>							± 0.37	
HT HCT	%	<b>50.6</b>	<b>50.6</b>							± 3.0	
	l/l	<b>0.506</b>	<b>0.506</b>							± 0.030	
VGM MCV	µm <sup>3</sup> ; fl	<b>89</b>	<b>89</b>							± 4	
TGMH MCH	pg	<b>31.8</b>	<b>32.0</b>							± 2.0	
	fmol	<b>1.98</b>	<b>1.99</b>							± 0.12	
CCMH MCHC	g/dl	<b>35.7</b>	<b>35.9</b>							± 3.0	
	g/l	<b>357</b>	<b>359</b>							± 30	
	mmol/l	<b>22.20</b>	<b>22.32</b>							± 1.86	
IDR RDW	%	N/A	<b>13.5</b>							± 3.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>485</b>	<b>515</b>							± 45	
VPM MPV	µm <sup>3</sup> ; fl	N/A	<b>8.3</b>							± 2.0	
LYMPHO	#	N/A	<b>2.49</b>							± 1.00	
	%	N/A	<b>12.5</b>							± 5.0	
MONO	#	N/A	<b>1.00</b>							± 0.80	
	%	N/A	<b>5.0</b>							± 4.0	
GRANULO	#	N/A	<b>16.42</b>							± 1.40	
	%	N/A	<b>82.5</b>							± 7.0	