




<b>LOT</b>		<b>MX 406</b>	<b>L</b>	<b>CONTROL</b>		<b>L</b>	 (Exp.)				<b>2017-10-05</b> (YYYY - MM - DD)	
		Rev 1		<b>LyseVet</b>								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS									TOLERANCES TOLERANCE	
		Sci Vet ABC	Sci Vet ABC Plus ESV 60									
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	1.6	1.6								± 0.4	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.44	2.42								± 0.15	
HB HGB	g/dl	6.1	6.1								± 0.4	
	g/l	61	61								± 4	
	mmol/l	3.79	3.79								± 0.25	
HT HCT	%	17.1	16.7								± 2.0	
	l/l	0.171	0.167								± 0.020	
VGM MCV	µm <sup>3</sup> ; fl	70	69								± 4	
TGMH MCH	pg	25.0	25.2								± 2.0	
	fmol	1.55	1.57								± 0.12	
CCMH MCHC	g/dl	35.7	36.5								± 3.0	
	g/l	357	365								± 30	
	mmol/l	22.18	22.69								± 1.86	
IDR RDW	%	N/A	14.0								± 3.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	70	80								± 20	
VPM MPV	µm <sup>3</sup> ; fl	N/A	8.2								± 2.0	
LYMPHO	#	N/A	0.82								± 0.40	
	%	N/A	51.0								± 8.0	
MONO	#	N/A	0.20								± 0.20	
	%	N/A	10.5								± 6.0	
GRANULO	#	N/A	0.62								± 0.40	
	%	N/A	38.5								± 7.0	

<b>LOT</b>		<b>MX 406 N</b>	<b>CONTROL N</b>							 (Exp.)	<b>2017-10-05</b> (YYYY-MM-DD)
		Rev 1	<b>LyseVet</b>								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS								TOLERANCES TOLERANCE	
		Scil Vet ABC	Scil Vet ABC Plus ESV 60								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>6.9</b>	<b>6.8</b>							± 0.8	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>4.78</b>	<b>4.75</b>							± 0.18	
HB HGB	g/dl	<b>13.4</b>	<b>13.4</b>							± 0.5	
	g/l	<b>134</b>	<b>134</b>							± 5	
	mmol/l	<b>8.32</b>	<b>8.32</b>							± 0.31	
HT HCT	%	<b>37.8</b>	<b>37.5</b>							± 2.5	
	l/l	<b>0.378</b>	<b>0.375</b>							± 0.025	
VGM MCV	µm <sup>3</sup> ; fl	<b>79</b>	<b>79</b>							± 4	
TGMH MCH	pg	<b>28.0</b>	<b>28.2</b>							± 2.0	
	fmol	<b>1.74</b>	<b>1.75</b>							± 0.12	
CCMH MCHC	g/dl	<b>35.5</b>	<b>35.7</b>							± 3.0	
	g/l	<b>355</b>	<b>357</b>							± 30	
	mmol/l	<b>22.04</b>	<b>22.18</b>							± 1.86	
IDR RDW	%	N/A	<b>14.0</b>							± 3.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>240</b>	<b>247</b>							± 40	
VPM MPV	µm <sup>3</sup> ; fl	N/A	<b>8.1</b>							± 2.0	
LYMPHO	#	N/A	<b>1.90</b>							± 0.40	
	%	N/A	<b>28.0</b>							± 5.0	
MONO	#	N/A	<b>0.54</b>							± 0.30	
	%	N/A	<b>8.0</b>							± 4.0	
GRANULO	#	N/A	<b>4.35</b>							± 0.60	
	%	N/A	<b>64.0</b>							± 6.0	

1300010912-A  
Ref: TEMP-0830 Rev.38

<b>LOT</b>		<b>MX 406 H</b>	<b>CONTROL H</b>							 (Exp.)	<b>2017-10-05</b> (YYYY-MM-DD)
		Rev 1	<b>LyseVet</b>								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS									TOLERANCES TOLERANCE
		Scil Vet ABC	Scil Vet ABC Plus ESV 60								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>19.3</b>	<b>19.3</b>								± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>5.78</b>	<b>5.78</b>								± 0.20
HB HGB	g/dl	<b>18.1</b>	<b>18.3</b>								± 0.6
	g/l	<b>181</b>	<b>183</b>								± 6
	mmol/l	<b>11.24</b>	<b>11.36</b>								± 0.37
HT HCT	%	<b>50.9</b>	<b>50.9</b>								± 3.0
	l/l	<b>0.509</b>	<b>0.509</b>								± 0.030
VGM MCV	µm <sup>3</sup> ; fl	<b>88</b>	<b>88</b>								± 4
TGMH MCH	pg	<b>31.3</b>	<b>31.7</b>								± 2.0
	fmol	<b>1.94</b>	<b>1.97</b>								± 0.12
CCMH MCHC	g/dl	<b>35.6</b>	<b>36.0</b>								± 3.0
	g/l	<b>356</b>	<b>360</b>								± 30
	mmol/l	<b>22.10</b>	<b>22.34</b>								± 1.86
IDR RDW	%	N/A	<b>13.5</b>								± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>485</b>	<b>482</b>								± 45
VPM MPV	µm <sup>3</sup> ; fl	N/A	<b>7.4</b>								± 2.0
LYMPHO	#	N/A	<b>2.22</b>								± 1.00
	%	N/A	<b>11.5</b>								± 5.0
MONO	#	N/A	<b>0.97</b>								± 0.80
	%	N/A	<b>5.0</b>								± 4.0
GRANULO	#	N/A	<b>16.12</b>								± 1.40
	%	N/A	<b>83.5</b>								± 7.0