

**HORIBA ABX SAS**  
 HORIBA ABX SAS  
 Parc Euromédecine  
 Rue du Caducée - BP 7290  
 34184 MONTPELLIER Cedex 4  
 FRANCE  
 Tél.: 33 (0) 4 67 14 15 16  
 Fax.: 33 (0) 4 67 14 15 17

**LOT** MX 072 L  
 Rev 1

**CONTROL** L

(Exp.) 2012-10-05  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX VETPACK								TOLERANCES TOLERANCE
		MICROS		MICROS						
		Scil Vet ABC	Scil Vet ABC Plus ESV 60							
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>1.8</b>	<b>2.1</b>							± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>2.43</b>	<b>2.41</b>							± 0.15
HB HGB	g/dl	<b>6.1</b>	<b>5.9</b>							± 0.4
	g/l	<b>61</b>	<b>59</b>							± 4
	mmol/l	<b>3.79</b>	<b>3.66</b>							± 0.25
HT HCT	%	<b>17.0</b>	<b>16.9</b>							± 2.0
	l/l	<b>0.170</b>	<b>0.169</b>							± 0.020
VGM MCV	µm <sup>3</sup> ; fl	<b>70</b>	<b>70</b>							± 4
TGMH MCH	pg	<b>25.1</b>	<b>24.5</b>							± 2.0
	fmoi	<b>1.56</b>	<b>1.52</b>							± 0.12
CCMH MCHC	g/dl	<b>35.9</b>	<b>35.0</b>							± 3.0
	g/l	<b>359</b>	<b>350</b>							± 30
	mmol/l	<b>22.27</b>	<b>21.72</b>							± 1.86
IDR RDW	%	N/A	<b>13.8</b>							± 4.0
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>65</b>	<b>63</b>							± 20
VPM MPV	µm <sup>3</sup> ; fl	N/A	<b>8.0</b>							± 2.0
LYMPHO	#	N/A	<b>1.1</b>							± 0.40
	%	N/A	<b>52.0</b>							± 9.0
MONO	#	N/A	<b>0.3</b>							± 0.20
	%	N/A	<b>13.5</b>							± 7.0
GRANULO	#	N/A	<b>0.7</b>							± 0.40
	%	N/A	<b>34.5</b>							± 8.0


9930041-E

Ref: TEMP-0830 Rev.17

**HORIBA ABX SAS**  
 HORIBA ABX SAS  
 Parc Euromédecine  
 Rue du Caducée - BP 7290  
 34184 MONTPELLIER Cedex 4  
 FRANCE  
 Tél.: 33 (0) 4 67 14 15 16  
 Fax.: 33 (0) 4 67 14 15 17

**LOT** MX 072 N  
 Rev 1

**CONTROL** N

 (Exp.) 2012-10-05  
 (YYYY - MM - DD)


PARAMETRES PARAMETERS	UNITES UNITS	ABX VETPACK								TOLERANCES TOLERANCE
		MICROS		MICROS						
		Scil Vet ABC	Scil Vet ABC Plus ESV 60							
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>7.0</b>	<b>8.1</b>							± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>4.72</b>	<b>4.72</b>							± 0.18
HB HGB	g/dl	<b>13.7</b>	<b>13.3</b>							± 0.5
	g/l	<b>137</b>	<b>133</b>							± 5
	mmol/l	<b>8.51</b>	<b>8.26</b>							± 0.31
HT HCT	%	<b>37.8</b>	<b>37.8</b>							± 2.5
	l/l	<b>0.378</b>	<b>0.378</b>							± 0.025
VGM MCV	µm <sup>3</sup> ; fl	<b>80</b>	<b>80</b>							± 4
TGMH MCH	pg	<b>29.0</b>	<b>28.2</b>							± 2.0
	fmoi	<b>1.80</b>	<b>1.75</b>							± 0.12
CCMH MCHC	g/dl	<b>36.3</b>	<b>35.2</b>							± 3.0
	g/l	<b>363</b>	<b>352</b>							± 30
	mmol/l	<b>22.53</b>	<b>21.87</b>							± 1.86
IDR RDW	%	N/A	<b>13.3</b>							± 4.0
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>250</b>	<b>220</b>							± 40
VPM MPV	µm <sup>3</sup> ; fl	N/A	<b>7.7</b>							± 2.0
LYMPHO	#	N/A	<b>2.1</b>							± 0.50
	%	N/A	<b>26.0</b>							± 6.0
MONO	#	N/A	<b>0.7</b>							± 0.40
	%	N/A	<b>9.0</b>							± 5.0
GRANULO	#	N/A	<b>5.3</b>							± 0.60
	%	N/A	<b>65.0</b>							± 7.0

9930042-E

Ref: TEMP-0830 Rev.17

**LOT** MX 072 H  
 Rev 1

**CONTROL** H

 (Exp.) 2012-10-05  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX VETPACK								TOLERANCES TOLERANCE
		MICROS		MICROS						
		Scil Vet ABC	Scil Vet ABC Plus ESV 60							
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>19.3</b>	<b>22.5</b>							± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>5.70</b>	<b>5.70</b>							± 0.20
HB HGB	g/dl	<b>18.1</b>	<b>17.8</b>							± 0.6
	g/l	<b>181</b>	<b>178</b>							± 6
	mmol/l	<b>11.24</b>	<b>11.05</b>							± 0.37
HT HCT	%	<b>50.2</b>	<b>50.7</b>							± 3.0
	l/l	<b>0.502</b>	<b>0.507</b>							± 0.030
VGM MCV	µm <sup>3</sup> ; fl	<b>88</b>	<b>89</b>							± 4
TGMH MCH	pg	<b>31.8</b>	<b>31.2</b>							± 2.0
	fmoi	<b>1.97</b>	<b>1.94</b>							± 0.12
CCMH MCHC	g/dl	<b>36.1</b>	<b>35.1</b>							± 3.0
	g/l	<b>361</b>	<b>351</b>							± 30
	mmol/l	<b>22.41</b>	<b>21.79</b>							± 1.86
IDR RDW	%	N/A	<b>12.3</b>							± 4.0
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>525</b>	<b>450</b>							± 45
VPM MPV	µm <sup>3</sup> ; fl	N/A	<b>7.5</b>							± 2.0
LYMPHO	#	N/A	<b>2.6</b>							± 1.00
	%	N/A	<b>11.5</b>							± 5.0
MONO	#	N/A	<b>1.2</b>							± 0.80
	%	N/A	<b>5.5</b>							± 4.0
GRANULO	#	N/A	<b>18.7</b>							± 1.40
	%	N/A	<b>83.0</b>							± 7.0

9930043-E

Ref: TEMP-0830 Rev.17