



<b>LOT</b>		<b>MX 036 L</b>	<b>CONTROL L</b>							 <b>(Exp.)</b>	<b>2016-06-05</b> (YYYY-MM-DD)
		Rev 1	<b>LyseVet</b>								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS								TOLERANCES TOLERANCE	
		Scil Vet ABC	Scil Vet ABC Plus ESV 60								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	1.8	1.9							± 0.4	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.46	2.41							± 0.15	
HB HGB	g/dl	6.1	6.2							± 0.4	
	g/l	61	62							± 4	
	mmol/l	3.79	3.85							± 0.25	
HT HCT	%	17.2	16.9							± 2.0	
	l/l	0.172	0.169							± 0.020	
VGM MCV	µm <sup>3</sup> ; fl	70	70							± 4	
TGMH MCH	pg	24.8	25.7							± 2.0	
	fmol	1.54	1.60							± 0.12	
CCMH MCHC	g/dl	35.4	36.8							± 3.0	
	g/l	354	368							± 30	
	mmol/l	22.00	22.82							± 1.86	
IDR RDW	%	N/A	14.5							± 3.0	
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	70	72							± 20	
VPM MPV	µm <sup>3</sup> ; fl	N/A	8.7							± 2.0	
LYMPHO	#	N/A	1.05							± 0.40	
	%	N/A	55.0							± 8.0	
MONO	#	N/A	0.22							± 0.20	
	%	N/A	11.5							± 6.0	
GRANULO	#	N/A	0.64							± 0.40	
	%	N/A	33.5							± 7.0	

LOT		MX 036 N	CONTROL N							(Exp.)	2016-06-05 (YYYY-MM-DD)	
Rev 1		LyseVet										
PARAMETRES PARAMETERS	UNITES UNITS	MICROS										TOLERANCES TOLERANCE
		Scil Vet ABC	Scil Vet ABC Plus ESV 60									
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	6.7	6.9									± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.61	4.62									± 0.18
HB HGB	g/dl	13.4	13.3									± 0.5
	g/l	134	133									± 5
	mmol/l	8.32	8.26									± 0.31
HT HCT	%	37.3	37.0									± 2.5
	l/l	0.373	0.370									± 0.025
VGM MCV	µm <sup>3</sup> ; fl	81	80									± 4
TGMH MCH	pg	29.1	28.8									± 2.0
	fmol	1.81	1.79									± 0.12
CCMH MCHC	g/dl	35.9	36.0									± 3.0
	g/l	359	360									± 30
	mmol/l	22.28	22.35									± 1.86
IDR RDW	%	N/A	13.5									± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	245	256									± 40
VPM MPV	µm <sup>3</sup> ; fl	N/A	8.2									± 2.0
LYMPHO	#	N/A	1.79									± 0.40
	%	N/A	26.0									± 5.0
MONO	#	N/A	0.66									± 0.30
	%	N/A	9.5									± 4.0
GRANULO	#	N/A	4.45									± 0.60
	%	N/A	64.5									± 6.0

1300010912-A  
Ref: TEMP-0830 Rev.37

<b>LOT</b>		<b>MX 036 H</b>	<b>CONTROL H</b>							 (Exp.)	<b>2016-06-05</b> (YYYY-MM-DD)
		Rev 1	<b>LyseVet</b>								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS									TOLERANCES TOLERANCE
		Scil Vet ABC	Scil Vet ABC Plus ESV 60								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>18.4</b>	<b>18.7</b>								± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>5.60</b>	<b>5.65</b>								± 0.20
HB HGB	g/dl	<b>17.9</b>	<b>17.9</b>								± 0.6
	g/l	<b>179</b>	<b>179</b>								± 6
	mmol/l	<b>11.12</b>	<b>11.12</b>								± 0.37
HT HCT	%	<b>49.8</b>	<b>49.7</b>								± 3.0
	l/l	<b>0.498</b>	<b>0.497</b>								± 0.030
VGM MCV	µm <sup>3</sup> ; fl	<b>89</b>	<b>88</b>								± 4
TGMH MCH	pg	<b>32.0</b>	<b>31.7</b>								± 2.0
	fmol	<b>1.98</b>	<b>1.97</b>								± 0.12
CCMH MCHC	g/dl	<b>35.9</b>	<b>36.0</b>								± 3.0
	g/l	<b>359</b>	<b>360</b>								± 30
	mmol/l	<b>22.30</b>	<b>22.36</b>								± 1.86
IDR RDW	%	N/A	<b>12.8</b>								± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>495</b>	<b>517</b>								± 45
VPM MPV	µm <sup>3</sup> ; fl	N/A	<b>7.9</b>								± 2.0
LYMPHO	#	N/A	<b>2.24</b>								± 1.00
	%	N/A	<b>12.0</b>								± 5.0
MONO	#	N/A	<b>0.94</b>								± 0.80
	%	N/A	<b>5.0</b>								± 4.0
GRANULO	#	N/A	<b>15.52</b>								± 1.40
	%	N/A	<b>83.0</b>								± 7.0