



<b>LOT</b>		<b>MX 095</b>	<b>L</b>	<b>CONTROL</b>		<b>L</b>	 <b>(Exp.)</b>				<b>2015-12-05</b> (YYYY - MM - DD)	
		Rev 1		<b>LyseVet</b>								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS									TOLERANCES TOLERANCE	
		Sci Vet ABC	Sci Vet ABC Plus ESV 60									
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>1.8</b>	<b>1.8</b>								± 0.4	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>2.48</b>	<b>2.44</b>								± 0.15	
HB HGB	g/dl	<b>6.2</b>	<b>6.1</b>								± 0.4	
	g/l	<b>62</b>	<b>61</b>								± 4	
	mmol/l	<b>3.85</b>	<b>3.79</b>								± 0.25	
HT HCT	%	<b>16.9</b>	<b>16.6</b>								± 2.0	
	l/l	<b>0.169</b>	<b>0.166</b>								± 0.020	
VGM MCV	µm <sup>3</sup> ; fl	<b>68</b>	<b>68</b>								± 4	
TGMH MCH	pg	<b>25.0</b>	<b>25.0</b>								± 2.0	
	fmol	<b>1.55</b>	<b>1.55</b>								± 0.12	
CCMH MCHC	g/dl	<b>36.8</b>	<b>36.8</b>								± 3.0	
	g/l	<b>368</b>	<b>368</b>								± 30	
	mmol/l	<b>22.83</b>	<b>22.83</b>								± 1.86	
IDR RDW	%	N/A	<b>14.2</b>								± 3.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>65</b>	<b>72</b>								± 20	
VPM MPV	µm <sup>3</sup> ; fl	N/A	<b>8.4</b>								± 2.0	
LYMPHO	#	N/A	<b>0.95</b>								± 0.40	
	%	N/A	<b>53.0</b>								± 8.0	
MONO	#	N/A	<b>0.21</b>								± 0.20	
	%	N/A	<b>11.5</b>								± 6.0	
GRANULO	#	N/A	<b>0.64</b>								± 0.40	
	%	N/A	<b>35.5</b>								± 7.0	

LOT		MX 095 N Rev 1		CONTROL N		(Exp.)				2015-12-05 (YYYY-MM-DD)	
PARAMETRES PARAMETERS	UNITES UNITS	LyseVet								TOLERANCES TOLERANCE	
		MICROS		MICROS							
		Scil Vet ABC	Scil Vet ABC Plus ESV 60								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.2	7.1								± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.51	4.50								± 0.18
HB HGB	g/dl	13.3	13.2								± 0.5
	g/l	133	132								± 5
	mmol/l	8.26	8.20								± 0.31
HT HCT	%	36.5	36.5								± 2.5
	l/l	0.365	0.365								± 0.025
VGM MCV	µm <sup>3</sup> ; fl	81	81								± 4
TGMH MCH	pg	29.5	29.3								± 2.0
	fmol	1.83	1.82								± 0.12
CCMH MCHC	g/dl	36.4	36.2								± 3.0
	g/l	364	362								± 30
	mmol/l	22.61	22.49								± 1.86
IDR RDW	%	N/A	13.5								± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	250	260								± 40
VPM MPV	µm <sup>3</sup> ; fl	N/A	8.6								± 2.0
LYMPHO	#	N/A	2.06								± 0.40
	%	N/A	29.0								± 5.0
MONO	#	N/A	0.57								± 0.30
	%	N/A	8.0								± 4.0
GRANULO	#	N/A	4.47								± 0.60
	%	N/A	63.0								± 6.0

1300010912-A  
Ref: TEMP-0830 Rev.37

<b>LOT</b>		<b>MX 095 H</b>	<b>CONTROL H</b>							 <b>(Exp.)</b>	<b>2015-12-05</b> (YYYY-MM-DD)
		Rev 1	<b>LyseVet</b>								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS									TOLERANCES TOLERANCE
		Scil Vet ABC	Scil Vet ABC Plus ESV 60								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>19.7</b>	<b>19.5</b>								± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>5.68</b>	<b>5.68</b>								± 0.20
HB HGB	g/dl	<b>18.1</b>	<b>18.1</b>								± 0.6
	g/l	<b>181</b>	<b>181</b>								± 6
	mmol/l	<b>11.24</b>	<b>11.24</b>								± 0.37
HT HCT	%	<b>50.0</b>	<b>50.0</b>								± 3.0
	l/l	<b>0.500</b>	<b>0.500</b>								± 0.030
VGM MCV	µm <sup>3</sup> ; fl	<b>88</b>	<b>88</b>								± 4
TGMH MCH	pg	<b>31.9</b>	<b>31.9</b>								± 2.0
	fmol	<b>1.98</b>	<b>1.98</b>								± 0.12
CCMH MCHC	g/dl	<b>36.2</b>	<b>36.2</b>								± 3.0
	g/l	<b>362</b>	<b>362</b>								± 30
	mmol/l	<b>22.49</b>	<b>22.49</b>								± 1.86
IDR RDW	%	N/A	<b>12.7</b>								± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>510</b>	<b>510</b>								± 45
VPM MPV	µm <sup>3</sup> ; fl	N/A	<b>8.3</b>								± 2.0
LYMPHO	#	N/A	<b>2.15</b>								± 1.00
	%	N/A	<b>11.0</b>								± 5.0
MONO	#	N/A	<b>1.07</b>								± 0.80
	%	N/A	<b>5.5</b>								± 4.0
GRANULO	#	N/A	<b>16.28</b>								± 1.40
	%	N/A	<b>83.5</b>								± 7.0