

LOT MX 401 L
 Rev 1

CONTROL L

(Exp.)

2016-12-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		PENTRA	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	
		DX NEXUS DF NEXUS		EMI CRP			XLR	60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0		2.0			1.9	1.9	1.9	2.0	± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.34		2.33			2.26	2.34	2.26	2.34	± 0.15
HB HGB	g/dl	5.9		6.0			5.8	5.8	5.8	5.9	± 0.4
	g/l	59		60			58	58	58	59	± 4
	mmol/l	3.66		3.73			3.60	3.60	3.60	3.66	± 0.25
HT HCT	%	16.6		16.1			16.0	16.1	16.0	16.6	± 2.0
	l/l	0.166		0.161			0.160	0.161	0.160	0.166	± 0.020
VGM MCV	µm ³ ; fl	71		69			71	69	71	71	± 4
TGMH MCH	pg	25.2		25.8			25.7	24.8	25.7	25.2	± 2.0
	fmol	1.57		1.60			1.59	1.54	1.59	1.57	± 0.12
CCMH MCHC	g/dl	35.5		37.3			36.1	35.9	36.1	35.5	± 3.0
	g/l	355		373			361	359	361	355	± 30
	mmol/l	22.05		23.18			22.45	22.31	22.45	22.05	± 1.86
IDR RDW	%	18.0		13.0			15.0	15.5	15.0	18.0	± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	74		74			* 69	* 75	* 69	74	± 20 *± 22
VPM MPV	µm ³ ; fl	10.1		8.0			10.1	10.2	10.1	10.1	± 2.0
LYMPHO	#	N/A		1.3			N/A	N/A	N/A	N/A	± 0.40
	%	N/A		63.0			N/A	N/A	N/A	N/A	± 8.0
MONO	#	N/A		0.2			N/A	N/A	N/A	N/A	± 0.20
	%	N/A		8.5			N/A	N/A	N/A	N/A	± 6.0
GRANULO	#	N/A		0.6			N/A	N/A	N/A	N/A	± 0.40
	%	N/A		28.5			N/A	N/A	N/A	N/A	± 7.0

* For ABX Pentra 60 Pentra 80 instruments only

LOT		MX 401 N		CONTROL N						2016-12-05 (YYYY-MM-DD)			
Rev 1													
		ABX Lysebio											
PARAMETRES PARAMETERS	UNITES UNITS	PENTRA	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE		
		DX NEXUS DF NEXUS		EMI CRP			XLR	60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120			
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	8.1		8.0			7.6	7.6	7.6	8.1	± 0.8		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.53		4.47			4.50	4.52	4.50	4.53	± 0.18		
HB HGB	g/dl	13.0		13.0			13.1	13.1	13.1	13.0	± 0.5		
	g/l	130		130			131	131	131	130	± 5		
	mmol/l	8.07		8.07			8.14	8.14	8.14	8.07	± 0.31		
HT HCT	%	35.3		35.3			36.0	36.2	36.0	35.3	± 2.5		
	l/l	0.353		0.353			0.360	0.362	0.360	0.353	± 0.025		
VGM MCV	µm ³ ; fl	78		79			80	80	80	78	± 4		
TGMH MCH	pg	28.7		29.1			29.1	29.0	29.1	28.7	± 2.0		
	fmol	1.78		1.81			1.81	1.80	1.81	1.78	± 0.12		
CCMH MCHC	g/dl	36.8		36.8			36.4	36.2	36.4	36.8	± 3.0		
	g/l	368		368			364	362	364	368	± 30		
	mmol/l	22.85		22.86			22.60	22.50	22.60	22.85	± 1.86		
IDR RDW	%	17.0		12.9			14.7	14.0	14.7	17.0	± 3.0		
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	260		255			* 249	* 253	* 249	260	± 40 *± 40		
VPM MPV	µm ³ ; fl	10.1		8.9			10.2	9.8	10.2	10.1	± 2.0		
LYMPHO	#	N/A		2.8			N/A	N/A	N/A	N/A	± 0.40		
	%	N/A		34.5			N/A	N/A	N/A	N/A	± 5.0		
MONO	#	N/A		0.6			N/A	N/A	N/A	N/A	± 0.30		
	%	N/A		7.5			N/A	N/A	N/A	N/A	± 4.0		
GRANULO	#	N/A		4.6			N/A	N/A	N/A	N/A	± 0.60		
	%	N/A		58.0			N/A	N/A	N/A	N/A	± 6.0		

* For ABX Pentra 60 Pentra 80 instruments only

1219930084-B
BACK / VERSO
Rev.37

LOT MX 401 H
Rev 1

CONTROL H

 (Exp.)

2016-12-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE	
		PENTRA		MICROS		MICROS		PENTRA		PENTRA		
		DX NEXUS DF NEXUS		EMI CRP		XLR	60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120			
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.7		20.0		19.8	20.2	19.8	20.7	± 1.6		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.70		5.58		5.65	5.63	5.65	5.70	± 0.20		
HB HGB	g/dl	17.9		17.5		18.0	18.0	18.0	17.9	± 0.6		
	g/l	179		175		180	180	180	179	± 6		
	mmol/l	11.12		10.87		11.18	11.18	11.18	11.12	± 0.37		
HT HCT	%	48.5		48.5		48.0	49.0	48.0	48.5	± 3.0		
	l/l	0.485		0.485		0.480	0.490	0.480	0.485	± 0.030		
VGM MCV	µm ³ ; fl	85		87		85	87	85	85	± 4		
TGMH MCH	pg	31.4		31.4		31.9	32.0	31.9	31.4	± 2.0		
	fmol	1.95		1.95		1.98	1.99	1.98	1.95	± 0.12		
CCMH MCHC	g/dl	36.9		36.0		37.5	36.7	37.5	36.9	± 3.0		
	g/l	369		360		375	367	375	369	± 30		
	mmol/l	22.94		22.39		23.28	22.82	23.28	22.94	± 1.86		
IDR RDW	%	16.0		12.3		13.2	13.0	13.2	16.0	± 3.0		
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	515		490		* 515	* 515	* 515	515	± 45 *± 55		
VPM MPV	µm ³ ; fl	9.3		8.4		9.4	9.1	9.4	9.3	± 2.0		
LYMPHO	#	N/A		3.2		N/A	N/A	N/A	N/A	± 1.00		
	%	N/A		16.0		N/A	N/A	N/A	N/A	± 5.0		
MONO	#	N/A		1.1		N/A	N/A	N/A	N/A	± 0.80		
	%	N/A		5.5		N/A	N/A	N/A	N/A	± 4.0		
GRANULO	#	N/A		15.7		N/A	N/A	N/A	N/A	± 1.40		
	%	N/A		78.5		N/A	N/A	N/A	N/A	± 7.0		

* For ABX Pentra 60 Pentra 80 instruments only