

**LOT** MX 412 L  
Rev 1

**CONTROL** L

 (Exp.)

**2018-10-05**  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE	
		PENTRA	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA		PENTRA
		DX NEXUS DF NEXUS		EMI CRP			XLR	60 60 C+ ES60	80 XL80		120 - 120 RETIC DF120 - DX120
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.1		2.1			2.1	2.1	2.1	2.1	± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.30		2.29			2.24	2.30	2.24	2.30	± 0.15
HB HGB	g/dl	5.9		6.1			5.8	5.8	5.8	5.9	± 0.4
	g/l	59		61			58	58	58	59	± 4
	mmol/l	3.66		3.79			3.60	3.60	3.60	3.66	± 0.25
HT HCT	%	16.3		16.0			15.9	15.9	15.9	16.3	± 2.0
	l/l	0.163		0.160			0.159	0.159	0.159	0.163	± 0.020
VGM MCV	µm <sup>3</sup> ; fl	71		70			71	69	71	71	± 4
TGMH MCH	pg	25.7		26.6			25.9	25.2	25.9	25.7	± 2.0
	fmol	1.59		1.65			1.61	1.57	1.61	1.59	± 0.12
CCMH MCHC	g/dl	36.1		38.1			36.5	36.5	36.5	36.1	± 3.0
	g/l	361		381			365	365	365	361	± 30
	mmol/l	22.44		23.63			22.65	22.70	22.65	22.44	± 1.86
IDR RDW	%	18.5		13.5			12.0	14.5	14.0	18.5	± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	79		79			* 73	* 79	* 73	79	± 20   *± 22
VPM MPV	µm <sup>3</sup> ; fl	9.5		7.8			10.0	9.9	10.0	9.5	± 2.0
LYMPHO	#	N/A		1.4			N/A	N/A	N/A	N/A	± 0.40
	%	N/A		65.0			N/A	N/A	N/A	N/A	± 8.0
MONO	#	N/A		0.2			N/A	N/A	N/A	N/A	± 0.20
	%	N/A		7.0			N/A	N/A	N/A	N/A	± 6.0
GRANULO	#	N/A		0.6			N/A	N/A	N/A	N/A	± 0.40
	%	N/A		28.0			N/A	N/A	N/A	N/A	± 7.0

\* For ABX Pentra 60 Pentra 80 instruments only

LOT		MX 412 N		CONTROL N						2018-10-05 (YYYY-MM-DD)	
Rev 1											
		ABX Lysebio									
PARAMETRES PARAMETERS	UNITES UNITS	PENTRA	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		DX NEXUS DF NEXUS		EMI CRP			XLR	60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	8.0		7.7			7.8	7.7	7.8	8.0	± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.63		4.58			4.56	4.56	4.56	4.63	± 0.18
HB HGB	g/dl	13.4		13.4			13.4	13.4	13.4	13.4	± 0.5
	g/l	134		134			134	134	134	134	± 5
	mmol/l	8.32		8.32			8.32	8.32	8.32	8.32	± 0.31
HT HCT	%	36.6		36.6			36.0	36.0	36.0	36.6	± 2.5
	l/l	0.366		0.366			0.360	0.360	0.360	0.366	± 0.025
VGM MCV	µm <sup>3</sup> ; fl	79		80			79	79	79	79	± 4
TGMH MCH	pg	28.9		29.3			29.4	29.4	29.4	28.9	± 2.0
	fmol	1.80		1.82			1.82	1.82	1.82	1.80	± 0.12
CCMH MCHC	g/dl	36.6		36.6			37.2	37.2	37.2	36.6	± 3.0
	g/l	366		366			372	372	372	366	± 30
	mmol/l	22.75		22.71			23.10	23.10	23.10	22.75	± 1.86
IDR RDW	%	18.0		13.0			13.0	13.5	14.0	18.0	± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	275		262			* 264	* 265	* 264	275	± 40   *± 40
VPM MPV	µm <sup>3</sup> ; fl	8.5		8.2			9.2	9.0	8.8	8.5	± 2.0
LYMPHO	#	N/A		2.6			N/A	N/A	N/A	N/A	± 0.40
	%	N/A		34.0			N/A	N/A	N/A	N/A	± 5.0
MONO	#	N/A		0.5			N/A	N/A	N/A	N/A	± 0.30
	%	N/A		7.0			N/A	N/A	N/A	N/A	± 4.0
GRANULO	#	N/A		4.5			N/A	N/A	N/A	N/A	± 0.60
	%	N/A		59.0			N/A	N/A	N/A	N/A	± 6.0

\* For ABX Pentra 60 Pentra 80 instruments only

1219930084-B  
BACK / VERSO  
Ret: TEMP-0830 Rev.36

**LOT** MX 412 H  
Rev 1

**CONTROL** H

 (Exp.)

**2018-10-05**  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		PENTRA	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	
		DX NEXUS DF NEXUS		EMI CRP			XLR	60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	22.1		20.8			21.2	21.7	21.2	22.1	± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	5.68		5.58			5.60	5.54	5.60	5.68	± 0.20
HB HGB	g/dl	18.0		17.8			18.2	18.2	18.2	18.0	± 0.6
	g/l	180		178			182	182	182	180	± 6
	mmol/l	11.18		11.05			11.30	11.30	11.30	11.18	± 0.37
HT HCT	%	48.3		49.1			47.6	48.2	47.6	48.3	± 3.0
	l/l	0.483		0.491			0.476	0.482	0.476	0.483	± 0.030
VGM MCV	µm <sup>3</sup> ; fl	85		88			85	87	85	85	± 4
TGMH MCH	pg	31.7		31.9			32.5	32.9	32.5	31.7	± 2.0
	fmol	1.97		1.98			2.02	2.04	2.02	1.97	± 0.12
CCMH MCHC	g/dl	37.3		36.2			38.2	37.8	38.2	37.3	± 3.0
	g/l	373		362			382	378	382	373	± 30
	mmol/l	23.15		22.51			23.74	23.45	23.74	23.15	± 1.86
IDR RDW	%	16.5		12.5			12.0	13.0	13.5	16.5	± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	540		505			* 538	* 530	* 538	540	± 45   *± 55
VPM MPV	µm <sup>3</sup> ; fl	8.9		8.3			9.3	9.0	8.9	8.9	± 2.0
LYMPHO	#	N/A		3.3			N/A	N/A	N/A	N/A	± 1.00
	%	N/A		16.0			N/A	N/A	N/A	N/A	± 5.0
MONO	#	N/A		1.2			N/A	N/A	N/A	N/A	± 0.80
	%	N/A		6.0			N/A	N/A	N/A	N/A	± 4.0
GRANULO	#	N/A		16.2			N/A	N/A	N/A	N/A	± 1.40
	%	N/A		78.0			N/A	N/A	N/A	N/A	± 7.0

\* For ABX Pentra 60 Pentra 80 instruments only