

LOT MX 056 L
Rev 1

CONTROL L

 (Exp.)

2016-08-05
(YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lysebio | | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|--|----------------------|--------|---------|--------|--------|--------|---------------------|------------|----------------------------------|-------------------------|
| | | PENTRA | MICROS | MICROS | MICROS | MICROS | PENTRA | PENTRA | PENTRA | PENTRA | |
| | | DX NEXUS DF NEXUS | | EMI CRP | | | XLR | 60 60 C+ ES60 | 80 XL80 | 120 - 120 RETIC DF120 - DX120 | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 2.0 | | 2.0 | | | 2.0 | 2.0 | 2.0 | 2.0 | ± 0.4 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 2.41 | | 2.42 | | | 2.35 | 2.41 | 2.35 | 2.41 | ± 0.15 |
| HB HGB | g/dl | 6.0 | | 6.1 | | | 5.9 | 5.9 | 5.9 | 6.0 | ± 0.4 |
| | g/l | 60 | | 61 | | | 59 | 59 | 59 | 60 | ± 4 |
| | mmol/l | 3.73 | | 3.79 | | | 3.66 | 3.66 | 3.66 | 3.73 | ± 0.25 |
| HT HCT | % | 16.9 | | 16.7 | | | 16.7 | 16.6 | 16.7 | 16.9 | ± 2.0 |
| | l/l | 0.169 | | 0.167 | | | 0.167 | 0.166 | 0.167 | 0.169 | ± 0.020 |
| VGM MCV | µm ³ ; fl | 70 | | 69 | | | 71 | 69 | 71 | 70 | ± 4 |
| TGMH MCH | pg | 24.9 | | 25.2 | | | 25.1 | 24.5 | 25.1 | 24.9 | ± 2.0 |
| | fmol | 1.55 | | 1.57 | | | 1.56 | 1.52 | 1.56 | 1.55 | ± 0.12 |
| CCMH MCHC | g/dl | 35.6 | | 36.5 | | | 35.4 | 35.5 | 35.4 | 35.6 | ± 3.0 |
| | g/l | 356 | | 365 | | | 354 | 355 | 354 | 356 | ± 30 |
| | mmol/l | 22.09 | | 22.69 | | | 21.96 | 22.03 | 21.96 | 22.09 | ± 1.86 |
| IDR RDW | % | 19.0 | | 13.5 | | | 15.5 | 16.0 | 15.5 | 19.0 | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 66 | | 72 | | | * 67 | * 68 | * 67 | 66 | ± 20 *± 22 |
| VPM MPV | µm ³ ; fl | 9.6 | | 8.3 | | | 10.2 | 9.5 | 10.2 | 9.6 | ± 2.0 |
| LYMPHO | # | N/A | | 1.3 | | | N/A | N/A | N/A | N/A | ± 0.40 |
| | % | N/A | | 62.5 | | | N/A | N/A | N/A | N/A | ± 8.0 |
| MONO | # | N/A | | 0.2 | | | N/A | N/A | N/A | N/A | ± 0.20 |
| | % | N/A | | 10.0 | | | N/A | N/A | N/A | N/A | ± 6.0 |
| GRANULO | # | N/A | | 0.6 | | | N/A | N/A | N/A | N/A | ± 0.40 |
| | % | N/A | | 27.5 | | | N/A | N/A | N/A | N/A | ± 7.0 |

* For ABX Pentra 60 Pentra 80 instruments only

| LOT | | MX 056 | N | CONTROL | | N | (Exp.) | | | | 2016-08-05 (YYYY - MM - DD) | |
|--------------------------|--|----------------------|----------|----------------|--|----------|---------------------|--------------|----------------------------------|--------------|---------------------------------------|-------------------------|
| Rev 1 | | ABX Lysebio | | | | | | | | | | |
| PARAMETRES PARAMETERS | UNITES UNITS | PENTRA | | MICROS | | MICROS | | PENTRA | | PENTRA | | TOLERANCES TOLERANCE |
| | | DX NEXUS DF NEXUS | | EMI CRP | | XLR | 60 60 C+ ES60 | 80 XL80 | 120 - 120 RETIC DF120 - DX120 | | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 8.5 | | 8.2 | | | | 8.1 | 8.3 | 8.1 | 8.5 | ± 0.8 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 4.60 | | 4.60 | | | | 4.60 | 4.62 | 4.60 | 4.60 | ± 0.18 |
| HB HGB | g/dl | 13.2 | | 13.1 | | | | 13.2 | 13.1 | 13.2 | 13.2 | ± 0.5 |
| | g/l | 132 | | 131 | | | | 132 | 131 | 132 | 132 | ± 5 |
| | mmol/l | 8.20 | | 8.14 | | | | 8.20 | 8.14 | 8.20 | 8.20 | ± 0.31 |
| HT HCT | % | 36.8 | | 36.8 | | | | 36.8 | 37.0 | 36.8 | 36.8 | ± 2.5 |
| | l/l | 0.368 | | 0.368 | | | | 0.368 | 0.370 | 0.368 | 0.368 | ± 0.025 |
| VGM MCV | µm ³ ; fl | 80 | | 80 | | | | 80 | 80 | 80 | 80 | ± 4 |
| TGMH MCH | pg | 28.7 | | 28.5 | | | | 28.7 | 28.4 | 28.7 | 28.7 | ± 2.0 |
| | fmol | 1.78 | | 1.77 | | | | 1.78 | 1.76 | 1.78 | 1.78 | ± 0.12 |
| CCMH MCHC | g/dl | 35.9 | | 35.6 | | | | 35.9 | 35.4 | 35.9 | 35.9 | ± 3.0 |
| | g/l | 359 | | 356 | | | | 359 | 354 | 359 | 359 | ± 30 |
| | mmol/l | 22.28 | | 22.11 | | | | 22.28 | 22.01 | 22.28 | 22.28 | ± 1.86 |
| IDR RDW | % | 16.5 | | 13.0 | | | | 14.5 | 14.5 | 14.5 | 16.5 | ± 3.0 |
| PLAQ. PLTS | 10 ⁹ /mm ³ ; 10 ⁹ /l | 261 | | 253 | | | | * 255 | * 260 | * 255 | 261 | ± 40 *± 40 |
| VPM MPV | µm ³ ; fl | 9.4 | | 8.3 | | | | 9.3 | 9.2 | 9.3 | 9.4 | ± 2.0 |
| LYMPHO | # | N/A | | 2.7 | | | | N/A | N/A | N/A | N/A | ± 0.40 |
| | % | N/A | | 33.0 | | | | N/A | N/A | N/A | N/A | ± 5.0 |
| MONO | # | N/A | | 0.7 | | | | N/A | N/A | N/A | N/A | ± 0.30 |
| | % | N/A | | 8.0 | | | | N/A | N/A | N/A | N/A | ± 4.0 |
| GRANULO | # | N/A | | 4.8 | | | | N/A | N/A | N/A | N/A | ± 0.60 |
| | % | N/A | | 59.0 | | | | N/A | N/A | N/A | N/A | ± 6.0 |

* For ABX Pentra 60 Pentra 80 instruments only

1219930084-B
BACK / VERSO
Ret: TEMP-0830 Rev.37

LOT MX 056 H
Rev 1

CONTROL H

 (Exp.)

2016-08-05
(YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lysebio | | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|--|----------------------|--------|---------|--------|--------|--------|---------------------|------------|----------------------------------|-------------------------|
| | | PENTRA | MICROS | MICROS | MICROS | MICROS | PENTRA | PENTRA | PENTRA | PENTRA | |
| | | DX NEXUS DF NEXUS | | EMI CRP | | | XLR | 60 60 C+ ES60 | 80 XL80 | 120 - 120 RETIC DF120 - DX120 | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 21.5 | | 20.5 | | | 21.0 | 21.5 | 21.0 | 21.5 | ± 1.6 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 5.83 | | 5.76 | | | 5.82 | 5.77 | 5.82 | 5.83 | ± 0.20 |
| HB HGB | g/dl | 18.3 | | 18.0 | | | 18.6 | 18.5 | 18.6 | 18.3 | ± 0.6 |
| | g/l | 183 | | 180 | | | 186 | 185 | 186 | 183 | ± 6 |
| | mmol/l | 11.36 | | 11.18 | | | 11.55 | 11.49 | 11.55 | 11.36 | ± 0.37 |
| HT HCT | % | 50.1 | | 50.7 | | | 50.6 | 50.8 | 50.6 | 50.1 | ± 3.0 |
| | l/l | 0.501 | | 0.507 | | | 0.506 | 0.508 | 0.506 | 0.501 | ± 0.030 |
| VGM MCV | µm ³ ; fl | 86 | | 88 | | | 87 | 88 | 87 | 86 | ± 4 |
| TGMH MCH | pg | 31.4 | | 31.3 | | | 32.0 | 32.1 | 32.0 | 31.4 | ± 2.0 |
| | fmol | 1.95 | | 1.94 | | | 1.98 | 1.99 | 1.98 | 1.95 | ± 0.12 |
| CCMH MCHC | g/dl | 36.5 | | 35.5 | | | 36.7 | 36.4 | 36.7 | 36.5 | ± 3.0 |
| | g/l | 365 | | 355 | | | 367 | 364 | 367 | 365 | ± 30 |
| | mmol/l | 22.67 | | 22.05 | | | 22.81 | 22.63 | 22.81 | 22.67 | ± 1.86 |
| IDR RDW | % | 16.0 | | 12.0 | | | 13.5 | 13.0 | 13.5 | 16.0 | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 529 | | 505 | | | * 532 | * 530 | * 532 | 529 | ± 45 *± 55 |
| VPM MPV | µm ³ ; fl | 8.8 | | 8.1 | | | 9.0 | 8.9 | 9.0 | 8.8 | ± 2.0 |
| LYMPHO | # | N/A | | 3.6 | | | N/A | N/A | N/A | N/A | ± 1.00 |
| | % | N/A | | 17.5 | | | N/A | N/A | N/A | N/A | ± 5.0 |
| MONO | # | N/A | | 1.2 | | | N/A | N/A | N/A | N/A | ± 0.80 |
| | % | N/A | | 6.0 | | | N/A | N/A | N/A | N/A | ± 4.0 |
| GRANULO | # | N/A | | 15.7 | | | N/A | N/A | N/A | N/A | ± 1.40 |
| | % | N/A | | 76.5 | | | N/A | N/A | N/A | N/A | ± 7.0 |

* For ABX Pentra 60 Pentra 80 instruments only