

ABX Minotrol 16



LOT MX 071 L

CONTROL L

(Exp.) **2011-10-05**
 (YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l							2.0	2.0	2.0	± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l							2.41	2.34	2.38	± 0.15
HB HGB	g/dl							5.9	5.9	6.0	± 0.4
	g/l							59	59	60	± 4
	mmol/l							3.66	3.66	3.73	± 0.25
HT HCT	%							16.6	17.1	16.9	± 2.0
	l/l							0.166	0.171	0.169	± 0.020
VGM MCV	µm ³ ; fl							69	73	71	± 4
TGMH MCH	pg							24.5	25.2	25.2	± 2.0
	fmol							1.52	1.57	1.57	± 0.12
CCMH MCHC	g/dl							35.5	34.5	35.5	± 3.0
	g/l							355	345	355	± 30
	mmol/l							22.03	21.45	22.05	± 1.86
IDR RDW	%							15.0	14.5	16.0	± 3.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l							* 77	* 70	73	± 20 *± 22
VPM MPV	µm ³ ; fl							9.5	9.5	9.0	± 2.0
LYMPHO	#							N/A	N/A	N/A	± 0.40
	%							N/A	N/A	N/A	± 7.0
MONO	#							N/A	N/A	N/A	± 0.20
	%							N/A	N/A	N/A	± 6.0
GRANULO	#							N/A	N/A	N/A	± 0.40
	%							N/A	N/A	N/A	± 7.0

* For ABX Pentra 60 Pentra 80 instruments only

Ref: FORM-0830 Rev.15 BACK / VERSO 9930046-A

ABX Minotrol 16



LOT MX 071 N

CONTROL N

(Exp.) **2011-10-05**
 (YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l							7.6	7.4	7.7	± 0.8
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l							4.68	4.65	4.68	± 0.18
HB HGB	g/dl							13.6	13.7	13.5	± 0.5
	g/l							136	137	135	± 5
	mmol/l							8.45	8.51	8.38	± 0.31
HT HCT	%							37.9	38.1	37.9	± 2.5
	l/l							0.379	0.381	0.379	± 0.025
VGM MCV	µm ³ ; fl							81	82	81	± 4
TGMH MCH	pg							29.1	29.5	28.8	± 2.0
	fmol							1.80	1.83	1.79	± 0.12
CCMH MCHC	g/dl							35.9	35.9	35.6	± 3.0
	g/l							359	359	356	± 30
	mmol/l							22.28	22.31	22.12	± 1.86
IDR RDW	%							13.5	14.0	15.5	± 3.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l							* 264	* 260	262	± 40 *± 40
VPM MPV	µm ³ ; fl							8.2	8.3	8.1	± 2.0
LYMPHO	#							N/A	N/A	N/A	± 0.50
	%							N/A	N/A	N/A	± 6.0
MONO	#							N/A	N/A	N/A	± 0.40
	%							N/A	N/A	N/A	± 5.0
GRANULO	#							N/A	N/A	N/A	± 0.60
	%							N/A	N/A	N/A	± 7.0

* For ABX Pentra 60 Pentra 80 instruments only

Ref: FORM-0830 Rev.15 BACK / VERSO 9930047-A

ABX Minotrol 16



LOT MX 071 H
Rev 1

CONTROL H

(Exp.) 2011-10-05
(YYYY - MM - DD)

		ABX Lysebio									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l							20.8	20.2	21.0	± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l							5.69	5.68	5.75	± 0.20
HB HGB	g/dl							18.2	18.2	17.8	± 0.6
	g/l							182	182	178	± 6
	mmol/l							11.30	11.30	11.05	± 0.37
HT HCT	%							50.1	50.0	49.5	± 3.0
	l/l							0.501	0.500	0.495	± 0.030
VGM MCV	µm ³ ; fl							88	88	86	± 4
TGMH MCH	pg							32.0	32.0	31.0	± 2.0
	fmol							1.99	1.99	1.92	± 0.12
CCMH MCHC	g/dl							36.3	36.4	36.0	± 3.0
	g/l							363	364	360	± 30
	mmol/l							22.57	22.61	22.35	± 1.86
IDR RDW	%							12.5	13.5	14.5	± 3.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l							* 522	* 530	520	± 45 *± 55
VPM MPV	µm ³ ; fl							8.5	8.6	8.5	± 2.0
LYMPHO	#							N/A	N/A	N/A	± 1.00
	%							N/A	N/A	N/A	± 5.0
MONO	#							N/A	N/A	N/A	± 0.80
	%							N/A	N/A	N/A	± 4.0
GRANULO	#							N/A	N/A	N/A	± 1.40
	%							N/A	N/A	N/A	± 7.0

* For ABX Pentra 60 Pentra 80 instruments only

Ref: FORM-0830 Rev.15 BACK / VERSO 9930048-A