



ABX Minotrol 16



LOT MX 052 L
 Rev 1

CONTROL L

(Exp.) 2012-08-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		PENTRA	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		DX NEXUS DF NEXUS	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX 120	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0						2.0	2.0	2.0	± 0.4
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.38						2.37	2.33	2.38	± 0.15
HB HGB	g/dl	5.9						5.8	5.9	5.9	± 0.4
	g/l	59						58	59	59	± 4
	mmol/l	3.66						3.60	3.66	3.66	± 0.25
HT HCT	%	17.1						16.6	17.2	17.1	± 2.0
	l/l	0.171						0.166	0.172	0.171	± 0.020
VGM MCV	µm ³ ; fl	72						70	74	72	± 4
TGMH MCH	pg	24.8						24.5	25.3	24.8	± 2.0
	fmol	1.54						1.52	1.57	1.54	± 0.12
CCMH MCHC	g/dl	34.4						35.0	34.2	34.4	± 3.0
	g/l	344						350	342	344	± 30
	mmol/l	21.38						21.71	21.25	21.38	± 1.86
IDR RDW	%	18.0						14.5	14.5	18.0	± 4.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	70						* 71	* 68	70	± 20 *± 22
VPM MPV	µm ³ ; fl	9.0						9.4	9.1	9.0	± 2.0
LYMPHO	#	N/A						N/A	N/A	N/A	± 0.40
	%	N/A						N/A	N/A	N/A	± 9.0
MONO	#	N/A						N/A	N/A	N/A	± 0.20
	%	N/A						N/A	N/A	N/A	± 7.0
GRANULO	#	N/A						N/A	N/A	N/A	± 0.40
	%	N/A						N/A	N/A	N/A	± 8.0

* For ABX Pentra 60 Pentra 80 instruments only

Ref: TEMP-0630 Rev.17 BACK / VERSO 9930067-A

ABX Minotrol 16



LOT MX 052 N
 Rev 1

CONTROL N

(Exp.) 2012-08-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		PENTRA	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		DX NEXUS DF NEXUS	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX 120	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	8.1						7.6	7.5	8.1	± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.70						4.68	4.73	4.70	± 0.18
HB HGB	g/dl	13.6						13.6	13.7	13.6	± 0.5
	g/l	136						136	137	136	± 5
	mmol/l	8.45						8.45	8.51	8.45	± 0.31
HT HCT	%	38.1						37.9	38.3	38.1	± 2.5
	l/l	0.381						0.379	0.383	0.381	± 0.025
VGM MCV	µm ³ ; fl	81						81	81	81	± 4
TGMH MCH	pg	28.9						29.1	29.0	28.9	± 2.0
	fmol	1.80						1.80	1.80	1.80	± 0.12
CCMH MCHC	g/dl	35.7						35.9	35.8	35.7	± 3.0
	g/l	357						359	358	357	± 30
	mmol/l	22.18						22.28	22.21	22.18	± 1.86
IDR RDW	%	18.0						13.5	14.0	18.0	± 4.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	256						* 250	* 255	256	± 40 *± 40
VPM MPV	µm ³ ; fl	8.0						8.3	8.3	8.0	± 2.0
LYMPHO	#	N/A						N/A	N/A	N/A	± 0.50
	%	N/A						N/A	N/A	N/A	± 6.0
MONO	#	N/A						N/A	N/A	N/A	± 0.40
	%	N/A						N/A	N/A	N/A	± 5.0
GRANULO	#	N/A						N/A	N/A	N/A	± 0.60
	%	N/A						N/A	N/A	N/A	± 7.0

Ref: TEMP-0830 Rev.17 BACK / VERSO 9930068-A

* For ABX Pentra 60 Pentra 80 instruments only



ABX Minotrol 16



LOT MX 052 H
 Rev 1

CONTROL H

(Exp.) 2012-08-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		PENTRA	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		DX NEXUS DF NEXUS	CRP CRP 200	8P	18P	60 8P - 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX 120	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	21.8						20.8	20.5	21.8	± 1.6
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	5.65						5.54	5.62	5.65	± 0.20
HB HGB	g/dl	17.9						18.2	18.2	17.9	± 0.6
	g/l	179						182	182	179	± 6
	mmol/l	11.12						11.30	11.30	11.12	± 0.37
HT HCT	%	49.2						49.3	49.5	49.2	± 3.0
	l/l	0.492						0.493	0.495	0.492	± 0.030
VGM MCV	µm ³ ; fl	87						89	88	87	± 4
TGMH MCH	pg	31.7						32.9	32.4	31.7	± 2.0
	fmol	1.97						2.04	2.01	1.97	± 0.12
CCMH MCHC	g/dl	36.4						36.9	36.8	36.4	± 3.0
	g/l	364						369	368	364	± 30
	mmol/l	22.61						22.92	22.85	22.61	± 1.86
IDR RDW	%	17.0						13.0	13.5	17.0	± 4.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	528						* 529	* 545	528	± 45 *± 55
VPM MPV	µm ³ ; fl	7.9						8.2	8.0	7.9	± 2.0
LYMPHO	#	N/A						N/A	N/A	N/A	± 1.00
	%	N/A						N/A	N/A	N/A	± 5.0
MONO	#	N/A						N/A	N/A	N/A	± 0.80
	%	N/A						N/A	N/A	N/A	± 4.0
GRANULO	#	N/A						N/A	N/A	N/A	± 1.40
	%	N/A						N/A	N/A	N/A	± 7.0

* For ABX Pentra 60 Pentra 80 instruments only

Ref: TEMP-0630 Rev.17 BACK / VERSO 9930069-A