

LOT		MX 095 L		CONTROL		L		(Exp.)		2015-12-05		
		Rev 1								(YYYY-MM-DD)		
PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio										TOLERANCES TOLERANCE
		PENTRA	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA		
		DX NEXUS DF NEXUS		EMI CRP			XLR	60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120		
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.1		2.1			2.0	1.9	2.0	2.1	± 0.4	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.49		2.45			2.42	2.48	2.42	2.49	± 0.15	
HB HGB	g/dl	6.1		6.2			6.0	6.0	6.0	6.1	± 0.4	
	g/l	61		62			60	60	60	61	± 4	
	mmol/l	3.79		3.85			3.73	3.73	3.73	3.79	± 0.25	
HT HCT	%	17.2		16.7			16.7	16.6	16.7	17.2	± 2.0	
	l/l	0.172		0.167			0.167	0.166	0.167	0.172	± 0.020	
VGM MCV	µm <sup>3</sup> ; fl	69		68			69	67	69	69	± 4	
TGMH MCH	pg	24.5		25.3			24.8	24.2	24.8	24.5	± 2.0	
	fmol	1.52		1.57			1.54	1.50	1.54	1.52	± 0.12	
CCMH MCHC	g/dl	35.5		37.2			35.9	36.1	35.9	35.5	± 3.0	
	g/l	355		372			359	361	359	355	± 30	
	mmol/l	22.05		23.11			22.31	22.42	22.31	22.05	± 1.86	
IDR RDW	%	19.0		13.5			15.5	16.0	15.5	19.0	± 3.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	70		75			* 62	* 63	* 62	70	± 20 *± 22	
VPM MPV	µm <sup>3</sup> ; fl	10.0		8.3			8.8	9.0	8.8	10.0	± 2.0	
LYMPHO	#	N/A		1.3			N/A	N/A	N/A	N/A	± 0.40	
	%	N/A		61.0			N/A	N/A	N/A	N/A	± 8.0	
MONO	#	N/A		0.2			N/A	N/A	N/A	N/A	± 0.20	
	%	N/A		10.0			N/A	N/A	N/A	N/A	± 6.0	
GRANULO	#	N/A		0.6			N/A	N/A	N/A	N/A	± 0.40	
	%	N/A		29.0			N/A	N/A	N/A	N/A	± 7.0	

\* For ABX Pentra 60 Pentra 80 instruments only

LOT		MX 095 N		CONTROL N						2015-12-05 (YYYY-MM-DD)	
Rev 1											
		ABX Lysebio									
PARAMETRES PARAMETERS	UNITES UNITS	PENTRA	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		DX NEXUS DF NEXUS		EMI CRP			XLR	60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	8.0		7.8			7.6	7.6	7.6	8.0	± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.50		4.50			4.50	4.52	4.50	4.50	± 0.18
HB HGB	g/dl	13.3		13.2			13.3	13.3	13.3	13.3	± 0.5
	g/l	133		132			133	133	133	133	± 5
	mmol/l	8.26		8.20			8.26	8.26	8.26	8.26	± 0.31
HT HCT	%	35.6		36.0			36.0	36.2	36.0	35.6	± 2.5
	l/l	0.356		0.360			0.360	0.362	0.360	0.356	± 0.025
VGM MCV	µm <sup>3</sup> ; fl	79		80			80	80	80	79	± 4
TGMH MCH	pg	29.6		29.3			29.6	29.4	29.6	29.6	± 2.0
	fmol	1.84		1.82			1.84	1.83	1.84	1.84	± 0.12
CCMH MCHC	g/dl	37.4		36.7			36.9	36.8	36.9	37.4	± 3.0
	g/l	374		367			369	368	369	374	± 30
	mmol/l	23.23		22.77			22.94	22.84	22.94	23.23	± 1.86
IDR RDW	%	17.0		13.0			14.5	13.5	14.5	17.0	± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	260		260			* 252	* 253	* 252	260	± 40   *± 40
VPM MPV	µm <sup>3</sup> ; fl	9.3		8.5			9.2	9.2	9.2	9.3	± 2.0
LYMPHO	#	N/A		2.7			N/A	N/A	N/A	N/A	± 0.40
	%	N/A		34.5			N/A	N/A	N/A	N/A	± 5.0
MONO	#	N/A		0.6			N/A	N/A	N/A	N/A	± 0.30
	%	N/A		7.5			N/A	N/A	N/A	N/A	± 4.0
GRANULO	#	N/A		4.5			N/A	N/A	N/A	N/A	± 0.60
	%	N/A		58.0			N/A	N/A	N/A	N/A	± 6.0

\* For ABX Pentra 60 Pentra 80 instruments only

1219930084-B  
BACK / VERSO  
Ret: TEMP-0830 Rev.37

LOT **MX 095 H**  
Rev 1

CONTROL **H**

 (Exp.)

**2015-12-05**  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		PENTRA	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	
		DX NEXUS DF NEXUS		EMI CRP			XLR	60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>20.9</b>		<b>20.0</b>			<b>20.4</b>	<b>20.5</b>	<b>20.4</b>	<b>20.9</b>	± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>5.70</b>		<b>5.60</b>			<b>5.69</b>	<b>5.65</b>	<b>5.69</b>	<b>5.70</b>	± 0.20
HB HGB	g/dl	<b>18.1</b>		<b>17.8</b>			<b>18.2</b>	<b>18.4</b>	<b>18.2</b>	<b>18.1</b>	± 0.6
	g/l	<b>181</b>		<b>178</b>			<b>182</b>	<b>184</b>	<b>182</b>	<b>181</b>	± 6
	mmol/l	<b>11.24</b>		<b>11.05</b>			<b>11.30</b>	<b>11.43</b>	<b>11.30</b>	<b>11.24</b>	± 0.37
HT HCT	%	<b>48.5</b>		<b>48.7</b>			<b>48.9</b>	<b>49.2</b>	<b>48.9</b>	<b>48.5</b>	± 3.0
	l/l	<b>0.485</b>		<b>0.487</b>			<b>0.489</b>	<b>0.492</b>	<b>0.489</b>	<b>0.485</b>	± 0.030
VGM MCV	µm <sup>3</sup> ; fl	<b>85</b>		<b>87</b>			<b>86</b>	<b>87</b>	<b>86</b>	<b>85</b>	± 4
TGMH MCH	pg	<b>31.8</b>		<b>31.8</b>			<b>32.0</b>	<b>32.6</b>	<b>32.0</b>	<b>31.8</b>	± 2.0
	fmol	<b>1.97</b>		<b>1.97</b>			<b>1.99</b>	<b>2.02</b>	<b>1.99</b>	<b>1.97</b>	± 0.12
CCMH MCHC	g/dl	<b>37.4</b>		<b>36.5</b>			<b>37.2</b>	<b>37.4</b>	<b>37.2</b>	<b>37.4</b>	± 3.0
	g/l	<b>374</b>		<b>365</b>			<b>372</b>	<b>374</b>	<b>372</b>	<b>374</b>	± 30
	mmol/l	<b>23.20</b>		<b>22.69</b>			<b>23.10</b>	<b>23.25</b>	<b>23.10</b>	<b>23.20</b>	± 1.86
IDR RDW	%	<b>16.1</b>		<b>12.0</b>			<b>13.7</b>	<b>12.5</b>	<b>13.7</b>	<b>16.1</b>	± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>530</b>		<b>505</b>			<b>* 515</b>	<b>* 515</b>	<b>* 515</b>	<b>530</b>	± 45   *± 55
VPM MPV	µm <sup>3</sup> ; fl	<b>8.9</b>		<b>8.2</b>			<b>9.0</b>	<b>8.5</b>	<b>9.0</b>	<b>8.9</b>	± 2.0
LYMPHO	#	N/A		<b>3.1</b>			N/A	N/A	N/A	N/A	± 1.00
	%	N/A		<b>15.5</b>			N/A	N/A	N/A	N/A	± 5.0
MONO	#	N/A		<b>1.1</b>			N/A	N/A	N/A	N/A	± 0.80
	%	N/A		<b>5.5</b>			N/A	N/A	N/A	N/A	± 4.0
GRANULO	#	N/A		<b>15.8</b>			N/A	N/A	N/A	N/A	± 1.40
	%	N/A		<b>79.0</b>			N/A	N/A	N/A	N/A	± 7.0

\* For ABX Pentra 60 Pentra 80 instruments only