

ABX Minotrol 16



LOT MX 070 L

CONTROL L

(Exp.) **2010-10-05**
 (YYYY - MM - DD)

Rev 1

		ABX Lysebio										
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE	
		ES60	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l							2.0	2.0	2.0	± 0.4	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l							2.38	2.34	2.38	± 0.15	
HB HGB	g/dl							6.0	6.0	6.2	± 0.4	
	g/l							60	60	62	± 4	
	mmol/l							3.73	3.73	3.85	± 0.25	
HT HCT	%							16.9	17.3	17.1	± 2.0	
	l/l							0.169	0.173	0.171	± 0.020	
VGM MCV	µm ³ ; fl							71	74	72	± 4	
TGMH MCH	pg							25.2	25.6	26.1	± 2.0	
	fmol							1.57	1.59	1.62	± 0.12	
CCMH MCHC	g/dl							35.5	34.7	36.2	± 3.0	
	g/l							355	347	362	± 30	
	mmol/l							22.05	21.52	22.47	± 1.86	
IDR RDW	%							15.0	14.5	16.3	± 3.0	
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l							* 78	* 75	80	± 20	*± 22
VPM MPV	µm ³ ; fl							10.0	10.0	9.7	± 2.0	
LYMPHO	#							N/A	N/A	N/A	± 0.40	
	%							N/A	N/A	N/A	± 7.0	
MONO	#							N/A	N/A	N/A	± 0.20	
	%							N/A	N/A	N/A	± 6.0	
GRANULO	#							N/A	N/A	N/A	± 0.40	
	%							N/A	N/A	N/A	± 7.0	

* For ABX Pentra 60 Pentra 80 instruments only

Ref: FORM-0830 Rev.14 BACK / VERSO 9930046-A

ABX Minotrol 16



LOT MX 070 N
 Rev 1

CONTROL N

(Exp.) 2010-10-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l							7.7	7.6	8.0	± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l							4.60	4.61	4.60	± 0.18
HB HGB	g/dl							13.4	13.5	13.5	± 0.5
	g/l							134	135	135	± 5
	mmol/l							8.32	8.38	8.38	± 0.31
HT HCT	%							37.3	37.3	36.8	± 2.5
	l/l							0.373	0.373	0.368	± 0.025
VGM MCV	µm ³ ; fl							81	81	80	± 4
TGMH MCH	pg							29.1	29.3	29.3	± 2.0
	fmol							1.81	1.82	1.82	± 0.12
CCMH MCHC	g/dl							36.0	36.2	36.7	± 3.0
	g/l							360	362	367	± 30
	mmol/l							22.33	22.45	22.78	± 1.86
IDR RDW	%							13.8	14.0	15.5	± 3.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l							* 265	* 265	262	± 40 *± 40
VPM MPV	µm ³ ; fl							9.2	8.9	9.1	± 2.0
LYMPHO	#							N/A	N/A	N/A	± 0.50
	%							N/A	N/A	N/A	± 6.0
MONO	#							N/A	N/A	N/A	± 0.40
	%							N/A	N/A	N/A	± 5.0
GRANULO	#							N/A	N/A	N/A	± 0.60
	%							N/A	N/A	N/A	± 7.0

* For ABX Pentra 60 Pentra 80 instruments only

Ref: FORM-0830 Rev.14 BACK / VERSO 9930047-A



ABX Minotrol 16



LOT MX 070 H

CONTROL H

(Exp.) **2010-10-05**
 (YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l							20.8	20.7	21.5	± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l							5.51	5.58	5.63	± 0.20
HB HGB	g/dl							17.8	17.9	17.7	± 0.6
	g/l							178	179	177	± 6
	mmol/l							11.05	11.12	10.99	± 0.37
HT HCT	%							48.5	48.5	48.4	± 3.0
	l/l							0.485	0.485	0.484	± 0.030
VGM MCV	µm ³ ; fl							88	87	86	± 4
TGMH MCH	pg							32.3	32.1	31.4	± 2.0
	fmol							2.01	1.99	1.95	± 0.12
CCMH MCHC	g/dl							36.7	36.9	36.6	± 3.0
	g/l							367	369	366	± 30
	mmol/l							22.80	22.90	22.70	± 1.86
IDR RDW	%							12.5	13.0	14.5	± 3.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l							* 505	* 525	516	± 45 *± 55
VPM MPV	µm ³ ; fl							9.4	9.3	9.4	± 2.0
LYMPHO	#							N/A	N/A	N/A	± 1.10
	%							N/A	N/A	N/A	± 5.0
MONO	#							N/A	N/A	N/A	± 0.80
	%							N/A	N/A	N/A	± 4.0
GRANULO	#							N/A	N/A	N/A	± 1.40
	%							N/A	N/A	N/A	± 7.0

* For ABX Pentra 60 Pentra 80 instruments only

Ref: FORM-0830 Rev.14 BACK / VERSO 9930048-A