

**LOT** MX 410 L  
Rev 1

**CONTROL** L

 (Exp.)

**2018-06-05**  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		PENTRA	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	
		DX NEXUS DF NEXUS		EMI CRP			XLR	60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.1		2.0			2.0	1.9	2.0	2.1	± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.30		2.30			2.23	2.31	2.23	2.30	± 0.15
HB HGB	g/dl	5.9		6.1			5.8	5.8	5.8	5.9	± 0.4
	g/l	59		61			58	58	58	59	± 4
	mmol/l	3.66		3.79			3.60	3.60	3.60	3.66	± 0.25
HT HCT	%	16.6		16.3			16.1	16.2	16.1	16.6	± 2.0
	l/l	0.166		0.163			0.161	0.162	0.161	0.166	± 0.020
VGM MCV	µm <sup>3</sup> ; fl	72		71			72	70	72	72	± 4
TGMH MCH	pg	25.7		26.5			26.0	25.1	26.0	25.7	± 2.0
	fmol	1.59		1.65			1.62	1.56	1.62	1.59	± 0.12
CCMH MCHC	g/dl	35.6		37.4			36.1	35.9	36.1	35.6	± 3.0
	g/l	356		374			361	359	361	356	± 30
	mmol/l	22.13		23.20			22.43	22.27	22.43	22.13	± 1.86
IDR RDW	%	17.0		13.5			13.5	16.5	15.0	17.0	± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	77		80			* 74	* 78	* 74	77	± 20    *± 22
VPM MPV	µm <sup>3</sup> ; fl	10.1		8.0			10.4	10.3	10.4	10.1	± 2.0
LYMPHO	#	N/A		1.3			N/A	N/A	N/A	N/A	± 0.40
	%	N/A		65.0			N/A	N/A	N/A	N/A	± 8.0
MONO	#	N/A		0.2			N/A	N/A	N/A	N/A	± 0.20
	%	N/A		7.0			N/A	N/A	N/A	N/A	± 6.0
GRANULO	#	N/A		0.6			N/A	N/A	N/A	N/A	± 0.40
	%	N/A		28.0			N/A	N/A	N/A	N/A	± 7.0

\* For ABX Pentra 60 Pentra 80 instruments only

LOT		MX 410 N		CONTROL N						2018-06-05 (YYYY-MM-DD)	
Rev 1											
		ABX Lysebio									
PARAMETRES PARAMETERS	UNITES UNITS	PENTRA	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		DX NEXUS DF NEXUS		EMI CRP			XLR	60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	8.3		8.0			7.7	7.8	7.7	8.3	± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.63		4.58			4.55	4.60	4.55	4.63	± 0.18
HB HGB	g/dl	13.5		13.4			13.4	13.5	13.4	13.5	± 0.5
	g/l	135		134			134	135	134	135	± 5
	mmol/l	8.38		8.32			8.32	8.38	8.32	8.38	± 0.31
HT HCT	%	37.0		36.6			36.4	36.8	36.4	37.0	± 2.5
	l/l	0.370		0.366			0.364	0.368	0.364	0.370	± 0.025
VGM MCV	µm <sup>3</sup> ; fl	80		80			80	80	80	80	± 4
TGMH MCH	pg	29.2		29.3			29.5	29.3	29.5	29.2	± 2.0
	fmol	1.81		1.82			1.83	1.82	1.83	1.81	± 0.12
CCMH MCHC	g/dl	36.4		36.6			36.8	36.7	36.8	36.4	± 3.0
	g/l	364		366			368	367	368	364	± 30
	mmol/l	22.63		22.71			22.86	22.78	22.86	22.63	± 1.86
IDR RDW	%	17.0		13.0			13.5	14.5	14.5	17.0	± 3.0
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	265		265			* 255	* 262	* 255	265	± 40   *± 40
VPM MPV	µm <sup>3</sup> ; fl	9.5		8.3			9.8	9.6	9.8	9.5	± 2.0
LYMPHO	#	N/A		2.9			N/A	N/A	N/A	N/A	± 0.40
	%	N/A		36.0			N/A	N/A	N/A	N/A	± 5.0
MONO	#	N/A		0.5			N/A	N/A	N/A	N/A	± 0.30
	%	N/A		6.0			N/A	N/A	N/A	N/A	± 4.0
GRANULO	#	N/A		4.6			N/A	N/A	N/A	N/A	± 0.60
	%	N/A		58.0			N/A	N/A	N/A	N/A	± 6.0

\* For ABX Pentra 60 Pentra 80 instruments only

1219930084-B BACK / VERSO Ref: TEMP-0830 Rev.36

**LOT** MX 410 H  
Rev 1

**CONTROL** H

(Exp.)

**2018-06-05**  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		PENTRA	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	
		DX NEXUS DF NEXUS		EMI CRP			XLR	60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	21.8		20.7			20.8	21.2	20.8	21.8	± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	5.75		5.58			5.65	5.65	5.65	5.75	± 0.20
HB HGB	g/dl	18.1		17.7			18.2	18.3	18.2	18.1	± 0.6
	g/l	181		177			182	183	182	181	± 6
	mmol/l	11.24		10.99			11.30	11.36	11.30	11.24	± 0.37
HT HCT	%	48.9		48.5			48.6	49.2	48.6	48.9	± 3.0
	l/l	0.489		0.485			0.486	0.492	0.486	0.489	± 0.030
VGM MCV	µm <sup>3</sup> ; fl	85		87			86	87	86	85	± 4
TGMH MCH	pg	31.5		31.7			32.2	32.4	32.2	31.5	± 2.0
	fmol	1.95		1.97			2.00	2.01	2.00	1.95	± 0.12
CCMH MCHC	g/dl	37.0		36.5			37.5	37.2	37.5	37.0	± 3.0
	g/l	370		365			375	372	375	370	± 30
	mmol/l	23.00		22.64			23.26	23.12	23.26	23.00	± 1.86
IDR RDW	%	16.0		12.5			12.0	13.0	13.5	16.0	± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	510		495			* 510	* 505	* 510	510	± 45   *± 55
VPM MPV	µm <sup>3</sup> ; fl	9.4		8.4			9.7	9.5	9.7	9.4	± 2.0
LYMPHO	#	N/A		3.6			N/A	N/A	N/A	N/A	± 1.00
	%	N/A		17.5			N/A	N/A	N/A	N/A	± 5.0
MONO	#	N/A		0.9			N/A	N/A	N/A	N/A	± 0.80
	%	N/A		4.5			N/A	N/A	N/A	N/A	± 4.0
GRANULO	#	N/A		16.1			N/A	N/A	N/A	N/A	± 1.40
	%	N/A		78.0			N/A	N/A	N/A	N/A	± 7.0

\* For ABX Pentra 60 Pentra 80 instruments only