

LOT MX 415 L
Rev 1

CONTROL L

 (Exp.)

2019-04-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		PENTRA	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	
		DX NEXUS DF NEXUS		EMI CRP			XLR	60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.1		2.0			2.0	1.9	2.0	2.1	± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.38		2.38			2.35	2.40	2.35	2.38	± 0.15
HB HGB	g/dl	6.0		6.2			5.9	5.9	5.9	6.0	± 0.4
	g/l	60		62			59	59	59	60	± 4
	mmol/l	3.73		3.85			3.66	3.66	3.66	3.73	± 0.25
HT HCT	%	16.7		16.4			16.5	16.3	16.5	16.7	± 2.0
	l/l	0.167		0.164			0.165	0.163	0.165	0.167	± 0.020
VGM MCV	µm ³ ; fl	70		69			70	68	70	70	± 4
TGMH MCH	pg	25.2		26.1			25.1	24.6	25.1	25.2	± 2.0
	fmol	1.57		1.62			1.56	1.53	1.56	1.57	± 0.12
CCMH MCHC	g/dl	36.0		37.8			35.9	36.2	35.9	36.0	± 3.0
	g/l	360		378			359	362	359	360	± 30
	mmol/l	22.36		23.45			22.27	22.45	22.27	22.36	± 1.86
IDR RDW	%	18.0		13.5			13.5	17.0	15.5	18.0	± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	70		75			* 67	* 71	* 67	70	± 20 *± 22
VPM MPV	µm ³ ; fl	9.5		7.7			10.0	9.2	9.3	9.5	± 2.0
LYMPHO	#	N/A		1.3			N/A	N/A	N/A	N/A	± 0.40
	%	N/A		64.0			N/A	N/A	N/A	N/A	± 8.0
MONO	#	N/A		0.2			N/A	N/A	N/A	N/A	± 0.20
	%	N/A		8.5			N/A	N/A	N/A	N/A	± 6.0
GRANULO	#	N/A		0.6			N/A	N/A	N/A	N/A	± 0.40
	%	N/A		27.5			N/A	N/A	N/A	N/A	± 7.0

* For ABX Pentra 60 Pentra 80 instruments only

LOT		MX 415 N		CONTROL N						2019-04-05 (YYYY-MM-DD)	
Rev 1											
		ABX Lysebio									
PARAMETRES PARAMETERS	UNITES UNITS	PENTRA	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		DX NEXUS DF NEXUS		EMI CRP			XLR	60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	8.4		7.8			7.8	7.7	7.8	8.4	± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.76		4.67			4.75	4.75	4.75	4.76	± 0.18
HB HGB	g/dl	13.6		13.4			13.5	13.6	13.5	13.6	± 0.5
	g/l	136		134			135	136	135	136	± 5
	mmol/l	8.45		8.32			8.38	8.45	8.38	8.45	± 0.31
HT HCT	%	37.1		36.4			37.1	37.1	37.1	37.1	± 2.5
	l/l	0.371		0.364			0.371	0.371	0.371	0.371	± 0.025
VGM MCV	µm ³ ; fl	78		78			78	78	78	78	± 4
TGMH MCH	pg	28.6		28.7			28.4	28.6	28.4	28.6	± 2.0
	fmol	1.77		1.78			1.76	1.78	1.76	1.77	± 0.12
CCMH MCHC	g/dl	36.6		36.8			36.4	36.7	36.4	36.6	± 3.0
	g/l	366		368			364	367	364	366	± 30
	mmol/l	22.75		22.84			22.63	22.80	22.63	22.75	± 1.86
IDR RDW	%	17.5		13.0			14.0	15.0	15.0	17.5	± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	268		265			* 265	* 268	* 265	268	± 40 *± 40
VPM MPV	µm ³ ; fl	9.7		8.5			10.1	9.6	9.6	9.7	± 2.0
LYMPHO	#	N/A		2.7			N/A	N/A	N/A	N/A	± 0.40
	%	N/A		35.0			N/A	N/A	N/A	N/A	± 5.0
MONO	#	N/A		0.5			N/A	N/A	N/A	N/A	± 0.30
	%	N/A		7.0			N/A	N/A	N/A	N/A	± 4.0
GRANULO	#	N/A		4.5			N/A	N/A	N/A	N/A	± 0.60
	%	N/A		58.0			N/A	N/A	N/A	N/A	± 6.0

* For ABX Pentra 60 Pentra 80 instruments only

1300060513 BACK / VERSO Ref: TEMP-0830 Rev.39

LOT MX 415 H
Rev 1

CONTROL H

 (Exp.)

2019-04-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		PENTRA	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	
		DX NEXUS DF NEXUS		EMI CRP			XLR	60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	22.0		20.6			21.0	21.2	21.0	22.0	± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.80		5.66			5.76	5.70	5.76	5.80	± 0.20
HB HGB	g/dl	18.3		17.9			18.4	18.5	18.4	18.3	± 0.6
	g/l	183		179			184	185	184	183	± 6
	mmol/l	11.36		11.12			11.43	11.49	11.43	11.36	± 0.37
HT HCT	%	49.3		49.8			49.0	49.6	49.0	49.3	± 3.0
	l/l	0.493		0.498			0.490	0.496	0.490	0.493	± 0.030
VGM MCV	µm ³ ; fl	85		88			85	87	85	85	± 4
TGMH MCH	pg	31.6		31.6			31.9	32.5	31.9	31.6	± 2.0
	fmol	1.96		1.96			1.98	2.02	1.98	1.96	± 0.12
CCMH MCHC	g/dl	37.1		35.9			37.6	37.3	37.6	37.1	± 3.0
	g/l	371		359			376	373	376	371	± 30
	mmol/l	23.05		22.32			23.34	23.17	23.34	23.05	± 1.86
IDR RDW	%	16.0		12.5			12.0	13.0	13.5	16.0	± 3.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	530		505			* 535	* 535	* 535	530	± 45 *± 55
VPM MPV	µm ³ ; fl	9.2		8.3			9.8	9.3	9.3	9.2	± 2.0
LYMPHO	#	N/A		3.4			N/A	N/A	N/A	N/A	± 1.00
	%	N/A		16.5			N/A	N/A	N/A	N/A	± 5.0
MONO	#	N/A		1.0			N/A	N/A	N/A	N/A	± 0.80
	%	N/A		5.0			N/A	N/A	N/A	N/A	± 4.0
GRANULO	#	N/A		16.2			N/A	N/A	N/A	N/A	± 1.40
	%	N/A		78.5			N/A	N/A	N/A	N/A	± 7.0

* For ABX Pentra 60 Pentra 80 instruments only