

ABX Minotrol 16



LOT MX 090 L

CONTROL L

(Exp.) **2010-12-05**
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l							2.0	2.0	2.0	± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l							2.46	2.41	2.46	± 0.15
HB HGB	g/dl							6.0	6.0	6.2	± 0.4
	g/l							60	60	62	± 4
	mmol/l							3.73	3.73	3.85	± 0.25
HT HCT	%							17.0	17.4	17.5	± 2.0
	l/l							0.170	0.174	0.175	± 0.020
VGM MCV	µm ³ ; fl							69	72	71	± 4
TGMH MCH	pg							24.4	24.9	25.2	± 2.0
	fmol							1.51	1.55	1.57	± 0.12
CCMH MCHC	g/dl							35.3	34.6	35.5	± 3.0
	g/l							353	346	355	± 30
	mmol/l							21.95	21.47	22.04	± 1.86
IDR RDW	%							14.5	14.5	16.2	± 3.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l							* 80	* 72	78	± 20 *± 22
VPM MPV	µm ³ ; fl							10.0	9.7	9.7	± 2.0
LYMPHO	#							N/A	N/A	N/A	± 0.40
	%							N/A	N/A	N/A	± 7.0
MONO	#							N/A	N/A	N/A	± 0.20
	%							N/A	N/A	N/A	± 6.0
GRANULO	#							N/A	N/A	N/A	± 0.40
	%							N/A	N/A	N/A	± 7.0

* For ABX Pentra 60 Pentra 80 instruments only

Ref: FORM-0830 Rev.14 BACK / VERSO 9930046-A

ABX Minotrol 16



LOT MX 090 N
 Rev 1

CONTROL N

(Exp.) 2010-12-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l							7.6	7.5	7.9	± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l							4.68	4.70	4.72	± 0.18
HB HGB	g/dl							13.6	13.6	13.6	± 0.5
	g/l							136	136	136	± 5
	mmol/l							8.45	8.45	8.45	± 0.31
HT HCT	%							37.9	38.1	37.8	± 2.5
	l/l							0.379	0.381	0.378	± 0.025
VGM MCV	µm ³ ; fl							81	81	80	± 4
TGMH MCH	pg							29.1	28.9	28.8	± 2.0
	fmol							1.80	1.80	1.79	± 0.12
CCMH MCHC	g/dl							35.9	35.7	36.0	± 3.0
	g/l							359	357	360	± 30
	mmol/l							22.28	22.18	22.37	± 1.86
IDR RDW	%							13.5	14.0	15.5	± 3.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l							* 260	* 260	260	± 40 *± 40
VPM MPV	µm ³ ; fl							9.0	9.0	8.8	± 2.0
LYMPHO	#							N/A	N/A	N/A	± 0.50
	%							N/A	N/A	N/A	± 6.0
MONO	#							N/A	N/A	N/A	± 0.40
	%							N/A	N/A	N/A	± 5.0
GRANULO	#							N/A	N/A	N/A	± 0.60
	%							N/A	N/A	N/A	± 7.0

* For ABX Pentra 60 Pentra 80 instruments only

Ref: FORM-0830 Rev.14 BACK / VERSO 9930047-A

ABX Minotrol 16



LOT MX 090 H

CONTROL H

(Exp.) **2010-12-05**
 (YYYY - MM - DD)

Rev 1

		ABX Lysebio										
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE	
		ES60	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l							20.8	20.6	21.6	± 1.6	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l							5.62	5.67	5.73	± 0.20	
HB HGB	g/dl							17.9	17.9	17.7	± 0.6	
	g/l							179	179	177	± 6	
	mmol/l							11.12	11.12	10.99	± 0.37	
HT HCT	%							48.9	48.8	48.7	± 3.0	
	l/l							0.489	0.488	0.487	± 0.030	
VGM MCV	µm ³ ; fl							87	86	85	± 4	
TGMH MCH	pg							31.9	31.6	30.9	± 2.0	
	fmol							1.98	1.96	1.92	± 0.12	
CCMH MCHC	g/dl							36.6	36.7	36.3	± 3.0	
	g/l							366	367	363	± 30	
	mmol/l							22.73	22.80	22.57	± 1.86	
IDR RDW	%							12.5	13.0	15.0	± 3.0	
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l							* 510	* 520	520	± 45	* ± 55
VPM MPV	µm ³ ; fl							9.0	8.9	8.6	± 2.0	
LYMPHO	#							N/A	N/A	N/A	± 1.00	
	%							N/A	N/A	N/A	± 5.0	
MONO	#							N/A	N/A	N/A	± 0.80	
	%							N/A	N/A	N/A	± 4.0	
GRANULO	#							N/A	N/A	N/A	± 1.40	
	%							N/A	N/A	N/A	± 7.0	

* For ABX Pentra 60 Pentra 80 instruments only

Ref: FORM-0830 Rev.14 BACK / VERSO 9930048-A