

ABX Minotrol 16



LOT MX 110 L

CONTROL L

(Exp.) **2011-02-05**
 (YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l							1.9	1.9	1.9	± 0.4
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l							2.34	2.27	2.31	± 0.15
HB HGB	g/dl							5.7	5.7	5.9	± 0.4
	g/l							57	57	59	± 4
	mmol/l							3.54	3.54	3.66	± 0.25
HT HCT	%							16.4	16.6	16.4	± 2.0
	l/l							0.164	0.166	0.164	± 0.020
VGM MCV	µm ³ ; fl							70	73	71	± 4
TGMH MCH	pg							24.4	25.1	25.5	± 2.0
	fmol							1.51	1.56	1.59	± 0.12
CCMH MCHC	g/dl							34.8	34.4	36.0	± 3.0
	g/l							348	344	360	± 30
	mmol/l							21.61	21.36	22.34	± 1.86
IDR RDW	%							14.5	14.5	16.5	± 3.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l							* 75	* 73	77	± 20 *± 22
VPM MPV	µm ³ ; fl							9.7	9.3	9.6	± 3.0
LYMPHO	#							N/A	N/A	N/A	± 0.40
	%							N/A	N/A	N/A	± 7.0
MONO	#							N/A	N/A	N/A	± 0.20
	%							N/A	N/A	N/A	± 6.0
GRANULO	#							N/A	N/A	N/A	± 0.40
	%							N/A	N/A	N/A	± 7.0

* For ABX Pentra 60 Pentra 80 instruments only

Ref: FORM-0830 Rev.14 BACK / VERSO 9930046-A



ABX Minotrol 16



LOT MX 110 N

CONTROL N

(Exp.) **2011-02-05**
 (YYYY - MM - DD)

Rev 1

		ABX Lysebio										
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE	
		ES60	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l							7.5	7.5	7.7	± 0.8	
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l							4.65	4.62	4.66	± 0.18	
HB HGB	g/dl							13.4	13.4	13.4	± 0.5	
	g/l							134	134	134	± 5	
	mmol/l							8.32	8.32	8.32	± 0.31	
HT HCT	%							37.7	37.4	37.3	± 2.5	
	l/l							0.377	0.374	0.373	± 0.025	
VGM MCV	µm ³ ; fl							81	81	80	± 4	
TGMH MCH	pg							28.8	29.0	28.8	± 2.0	
	fmol							1.79	1.80	1.79	± 0.12	
CCMH MCHC	g/dl							35.6	35.8	35.9	± 3.0	
	g/l							356	358	359	± 30	
	mmol/l							22.09	22.24	22.32	± 1.86	
IDR RDW	%							13.5	14.0	16.0	± 3.0	
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l							* 260	* 263	264	± 40	*± 40
VPM MPV	µm ³ ; fl							9.0	8.9	8.7	± 2.0	
LYMPHO	#							N/A	N/A	N/A	± 0.50	
	%							N/A	N/A	N/A	± 6.0	
MONO	#							N/A	N/A	N/A	± 0.40	
	%							N/A	N/A	N/A	± 5.0	
GRANULO	#							N/A	N/A	N/A	± 0.60	
	%							N/A	N/A	N/A	± 7.0	

* For ABX Pentra 60 Pentra 80 instruments only

Ref: FORM-0830 Rev.14 BACK / VERSO 9930047-A

ABX Minotrol 16



LOT MX 110 H
 Rev 1

CONTROL H

(Exp.) 2011-02-05
 (YYYY - MM - DD)

		ABX Lysebio									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60		60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l							21.3	21.3	21.5	± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l							5.53	5.53	5.60	± 0.20
HB HGB	g/dl							17.8	17.6	17.5	± 0.6
	g/l							178	176	175	± 6
	mmol/l							11.05	10.93	10.87	± 0.37
HT HCT	%							48.7	48.7	48.2	± 3.0
	l/l							0.487	0.487	0.482	± 0.030
VGM MCV	µm ³ ; fl							88	88	86	± 4
TGMH MCH	pg							32.2	31.8	31.3	± 2.0
	fmol							2.00	1.98	1.94	± 0.12
CCMH MCHC	g/dl							36.6	36.2	36.3	± 3.0
	g/l							366	362	363	± 30
	mmol/l							22.71	22.46	22.57	± 1.86
IDR RDW	%							12.5	13.3	15.0	± 3.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l							* 510	* 530	520	± 45 *± 55
VPM MPV	µm ³ ; fl							8.6	8.6	8.4	± 2.0
LYMPHO	#							N/A	N/A	N/A	± 1.00
	%							N/A	N/A	N/A	± 5.0
MONO	#							N/A	N/A	N/A	± 0.80
	%							N/A	N/A	N/A	± 4.0
GRANULO	#							N/A	N/A	N/A	± 1.40
	%							N/A	N/A	N/A	± 7.0

* For ABX Pentra 60 Pentra 80 instruments only

Ref: FORM-0830 Rev.14 BACK / VERSO 9930048-A