

**LOT** MX 035 L  
Rev 1


**CONTROL** L

 (Exp.)

**2015-06-05**  
(YYYY-MM-DD)


PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE	
		PENTRA	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA		PENTRA
		DX NEXUS DF NEXUS		EMI CRP			XLR	60 60 C+ ES60	80 XL80		120 - 120 RETIC DF120 - DX120
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.1		2.1			2.1	2.0	2.1	2.1	± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.40		2.35			2.30	2.35	2.30	2.40	± 0.15
HB HGB	g/dl	6.0		6.1			5.9	5.9	5.9	6.0	± 0.4
	g/l	60		61			59	59	59	60	± 4
	mmol/l	3.73		3.79			3.66	3.66	3.66	3.73	± 0.25
HT HCT	%	16.8		16.2			16.3	16.0	16.3	16.8	± 2.0
	l/l	0.168		0.162			0.163	0.160	0.163	0.168	± 0.020
VGM MCV	µm <sup>3</sup> ; fl	70		69			71	68	71	70	± 4
TGMH MCH	pg	25.0		26.0			25.7	25.1	25.7	25.0	± 2.0
	fmol	1.55		1.61			1.59	1.56	1.59	1.55	± 0.12
CCMH MCHC	g/dl	35.7		37.6			36.1	36.9	36.1	35.7	± 3.0
	g/l	357		376			361	369	361	357	± 30
	mmol/l	22.18		23.36			22.44	22.93	22.44	22.18	± 1.86
IDR RDW	%	17.6		13.0			15.5	14.5	15.5	17.6	± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	73		78			* 69	* 74	* 69	73	± 20   *± 22
VPM MPV	µm <sup>3</sup> ; fl	10.4		8.8			10.3	10.4	10.3	10.4	± 2.0
LYMPHO	#	N/A		1.3			N/A	N/A	N/A	N/A	± 0.40
	%	N/A		60.0			N/A	N/A	N/A	N/A	± 8.0
MONO	#	N/A		0.2			N/A	N/A	N/A	N/A	± 0.20
	%	N/A		10.5			N/A	N/A	N/A	N/A	± 6.0
GRANULO	#	N/A		0.6			N/A	N/A	N/A	N/A	± 0.40
	%	N/A		29.5			N/A	N/A	N/A	N/A	± 7.0

\* For ABX Pentra 60 Pentra 80 instruments only

<b>LOT</b>		<b>MX 035</b>	<b>N</b>	<b>CONTROL</b>		<b>N</b>	 <b>(Exp.)</b>				<b>2015-06-05</b> (YYYY-MM-DD)	
Rev 1		<b>ABX Lysebio</b>										
PARAMETRES PARAMETERS	UNITES UNITS	PENTRA		MICROS		MICROS		PENTRA		PENTRA		TOLERANCES TOLERANCE
		DX NEXUS DF NEXUS		EMI CRP		XLR	60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120			
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>8.2</b>		<b>8.0</b>		<b>8.0</b>	<b>8.0</b>	<b>8.0</b>	<b>8.0</b>	<b>8.2</b>	± 0.8	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>4.60</b>		<b>4.50</b>		<b>4.54</b>	<b>4.55</b>	<b>4.54</b>	<b>4.60</b>		± 0.18	
HB HGB	g/dl	<b>13.5</b>		<b>13.4</b>		<b>13.6</b>	<b>13.5</b>	<b>13.6</b>	<b>13.5</b>		± 0.5	
	g/l	<b>135</b>		<b>134</b>		<b>136</b>	<b>135</b>	<b>136</b>	<b>135</b>		± 5	
	mmol/l	<b>8.38</b>		<b>8.32</b>		<b>8.45</b>	<b>8.38</b>	<b>8.45</b>	<b>8.38</b>		± 0.31	
HT HCT	%	<b>36.8</b>		<b>36.0</b>		<b>36.8</b>	<b>36.4</b>	<b>36.8</b>	<b>36.8</b>		± 2.5	
	l/l	<b>0.368</b>		<b>0.360</b>		<b>0.368</b>	<b>0.364</b>	<b>0.368</b>	<b>0.368</b>		± 0.025	
VGM MCV	µm <sup>3</sup> ; fl	<b>80</b>		<b>80</b>		<b>81</b>	<b>80</b>	<b>81</b>	<b>80</b>		± 4	
TGMH MCH	pg	<b>29.3</b>		<b>29.8</b>		<b>30.0</b>	<b>29.7</b>	<b>30.0</b>	<b>29.3</b>		± 2.0	
	fmol	<b>1.82</b>		<b>1.85</b>		<b>1.86</b>	<b>1.84</b>	<b>1.86</b>	<b>1.82</b>		± 0.12	
CCMH MCHC	g/dl	<b>36.7</b>		<b>37.2</b>		<b>37.0</b>	<b>37.1</b>	<b>37.0</b>	<b>36.7</b>		± 3.0	
	g/l	<b>367</b>		<b>372</b>		<b>370</b>	<b>371</b>	<b>370</b>	<b>367</b>		± 30	
	mmol/l	<b>22.78</b>		<b>23.12</b>		<b>22.97</b>	<b>23.03</b>	<b>22.97</b>	<b>22.78</b>		± 1.86	
IDR RDW	%	<b>16.0</b>		<b>12.3</b>		<b>14.5</b>	<b>13.5</b>	<b>14.5</b>	<b>16.0</b>		± 3.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>272</b>		<b>263</b>		<b>* 260</b>	<b>* 261</b>	<b>* 260</b>	<b>272</b>	± 40	*± 40	
VPM MPV	µm <sup>3</sup> ; fl	<b>9.8</b>		<b>8.7</b>		<b>9.6</b>	<b>9.6</b>	<b>9.6</b>	<b>9.8</b>		± 2.0	
LYMPHO	#	N/A		<b>2.5</b>		N/A	N/A	N/A	N/A		± 0.40	
	%	N/A		<b>31.5</b>		N/A	N/A	N/A	N/A		± 5.0	
MONO	#	N/A		<b>0.5</b>		N/A	N/A	N/A	N/A		± 0.30	
	%	N/A		<b>6.5</b>		N/A	N/A	N/A	N/A		± 4.0	
GRANULO	#	N/A		<b>5.0</b>		N/A	N/A	N/A	N/A		± 0.60	
	%	N/A		<b>62.0</b>		N/A	N/A	N/A	N/A		± 6.0	

\* For ABX Pentra 60 Pentra 80 instruments only

99300084-B  
BACK / VERSO  
Ret: TEMP-0830 Rev.36

<b>LOT</b>		<b>MX 035 H</b>	<b>CONTROL H</b>				 <b>(Exp.)</b>				<b>2015-06-05</b> (YYYY-MM-DD)
Rev 1		<b>ABX Lysebio</b>									
PARAMETRES PARAMETERS	UNITES UNITS	PENTRA	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		DX NEXUS DF NEXUS		EMI CRP			XLR	60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	21.6		21.0			20.9	21.3	20.9	21.6	± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	5.95		5.81			5.94	5.83	5.94	5.95	± 0.20
HB HGB	g/dl	18.3		18.1			18.6	18.5	18.6	18.3	± 0.6
	g/l	183		181			186	185	186	183	± 6
	mmol/l	11.36		11.24			11.55	11.49	11.55	11.36	± 0.37
HT HCT	%	49.4		49.4			50.5	50.1	50.5	49.4	± 3.0
	l/l	0.494		0.494			0.505	0.501	0.505	0.494	± 0.030
VGM MCV	µm <sup>3</sup> ; fl	83		85			85	86	85	83	± 4
TGMH MCH	pg	30.8		31.2			31.3	31.7	31.3	30.8	± 2.0
	fmol	1.91		1.93			1.94	1.97	1.94	1.91	± 0.12
CCMH MCHC	g/dl	37.1		36.7			36.8	36.9	36.8	37.1	± 3.0
	g/l	371		367			368	369	368	371	± 30
	mmol/l	23.01		22.76			22.88	22.91	22.88	23.01	± 1.86
IDR RDW	%	15.8		11.9			14.0	12.5	14.0	15.8	± 3.0
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	535		505			* 520	* 510	* 520	535	± 45   *± 55
VPM MPV	µm <sup>3</sup> ; fl	9.6		8.7			9.6	9.4	9.6	9.6	± 2.0
LYMPHO	#	N/A		3.8			N/A	N/A	N/A	N/A	± 1.00
	%	N/A		18.0			N/A	N/A	N/A	N/A	± 5.0
MONO	#	N/A		1.2			N/A	N/A	N/A	N/A	± 0.80
	%	N/A		5.5			N/A	N/A	N/A	N/A	± 4.0
GRANULO	#	N/A		16.1			N/A	N/A	N/A	N/A	± 1.40
	%	N/A		76.5			N/A	N/A	N/A	N/A	± 7.0

\* For ABX Pentra 60 Pentra 80 instruments only