

**LOT** MX 422 L  
Rev 1

**CONTROL** L

 (Exp.)

**2020-06-05**  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		PENTRA	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	
		DX NEXUS DF NEXUS		EMI CRP			XLR	60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.1		2.0			2.1	2.0	2.1	2.1	± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.36		2.36			2.30	2.38	2.30	2.36	± 0.15
HB HGB	g/dl	6.1		6.2			5.9	6.0	5.9	6.1	± 0.4
	g/l	61		62			59	60	59	61	± 4
	mmol/l	3.79		3.85			3.66	3.73	3.66	3.79	± 0.25
HT HCT	%	16.8		16.3			16.1	16.2	16.1	16.8	± 2.0
	l/l	0.168		0.163			0.161	0.162	0.161	0.168	± 0.020
VGM MCV	µm <sup>3</sup> ; fl	71		69			70	68	70	71	± 4
TGMH MCH	pg	25.8		26.3			25.7	25.2	25.7	25.8	± 2.0
	fmol	1.61		1.63			1.59	1.57	1.59	1.61	± 0.12
CCMH MCHC	g/dl	36.4		38.1			36.6	37.1	36.6	36.4	± 3.0
	g/l	364		381			366	371	366	364	± 30
	mmol/l	22.61		23.64			22.76	23.02	22.76	22.61	± 1.86
IDR RDW	%	17.8		13.8			13.0	16.0	15.5	17.8	± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	75		72			* 70	* 72	* 70	75	± 20 *± 22
VPM MPV	µm <sup>3</sup> ; fl	9.7		8.1			10.4	10.2	9.7	9.7	± 2.0
LYMPHO	#	N/A		1.3			N/A	N/A	N/A	N/A	± 0.40
	%	N/A		63.0			N/A	N/A	N/A	N/A	± 8.0
MONO	#	N/A		0.2			N/A	N/A	N/A	N/A	± 0.20
	%	N/A		10.0			N/A	N/A	N/A	N/A	± 6.0
GRANULO	#	N/A		0.5			N/A	N/A	N/A	N/A	± 0.40
	%	N/A		27.0			N/A	N/A	N/A	N/A	± 7.0

\* For ABX Pentra 60 Pentra 80 instruments only

LOT		MX 422 N		CONTROL N						2020-06-05 (YYYY - MM - DD)	
Rev 1											
		ABX Lysebio									
PARAMETRES PARAMETERS	UNITES UNITS	PENTRA	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		DX NEXUS DF NEXUS		EMI CRP			XLR	60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	8.2		7.9			7.9	7.9	7.9	8.2	± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.72		4.64			4.70	4.71	4.70	4.72	± 0.18
HB HGB	g/dl	13.4		13.3			13.4	13.4	13.4	13.4	± 0.5
	g/l	134		133			134	134	134	134	± 5
	mmol/l	8.32		8.26			8.32	8.32	8.32	8.32	± 0.31
HT HCT	%	36.3		36.2			36.7	36.7	36.7	36.3	± 2.5
	l/l	0.363		0.362			0.367	0.367	0.367	0.363	± 0.025
VGM MCV	µm <sup>3</sup> ; fl	77		78			78	78	78	77	± 4
TGMH MCH	pg	28.4		28.7			28.5	28.5	28.5	28.4	± 2.0
	fmol	1.76		1.78			1.77	1.77	1.77	1.76	± 0.12
CCMH MCHC	g/dl	36.9		36.7			36.6	36.5	36.6	36.9	± 3.0
	g/l	369		367			366	365	366	369	± 30
	mmol/l	22.90		22.82			22.70	22.65	22.70	22.90	± 1.86
IDR RDW	%	16.6		13.0			13.7	14.2	15.2	16.6	± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	275		260			* 270	* 270	* 270	275	± 40   *± 40
VPM MPV	µm <sup>3</sup> ; fl	9.0		8.4			9.3	9.1	9.0	9.0	± 2.0
LYMPHO	#	N/A		2.7			N/A	N/A	N/A	N/A	± 0.40
	%	N/A		34.5			N/A	N/A	N/A	N/A	± 5.0
MONO	#	N/A		0.5			N/A	N/A	N/A	N/A	± 0.30
	%	N/A		6.0			N/A	N/A	N/A	N/A	± 4.0
GRANULO	#	N/A		4.7			N/A	N/A	N/A	N/A	± 0.60
	%	N/A		59.5			N/A	N/A	N/A	N/A	± 6.0

\* For ABX Pentra 60 Pentra 80 instruments only

1300060513 BACK / VERSO Ref: TEMP-0830 Rev.39

**LOT** MX 422 H  
Rev 1

**CONTROL** H

(Exp.)

**2020-06-05**  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		PENTRA	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	
		DX NEXUS DF NEXUS		EMI CRP			XLR	60 60 C+ ES60	80 XL80	120 - 120 RETIC DF120 - DX120	
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	22.2		21.2			21.4	21.8	21.4	22.2	± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	5.73		5.58			5.70	5.67	5.70	5.73	± 0.20
HB HGB	g/dl	18.3		17.7			18.5	18.5	18.5	18.3	± 0.6
	g/l	183		177			185	185	185	183	± 6
	mmol/l	11.36		10.99			11.49	11.49	11.49	11.36	± 0.37
HT HCT	%	48.1		48.5			49.0	49.3	49.0	48.1	± 3.0
	l/l	0.481		0.485			0.490	0.493	0.490	0.481	± 0.030
VGM MCV	µm <sup>3</sup> ; fl	84		87			86	87	86	84	± 4
TGMH MCH	pg	31.9		31.7			32.5	32.6	32.5	31.9	± 2.0
	fmol	1.98		1.97			2.02	2.03	2.02	1.98	± 0.12
CCMH MCHC	g/dl	38.0		36.5			37.7	37.5	37.7	38.0	± 3.0
	g/l	380		365			377	375	377	380	± 30
	mmol/l	23.61		22.64			23.44	23.29	23.44	23.61	± 1.86
IDR RDW	%	16.5		12.0			13.0	13.5	14.0	16.5	± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	545		515			* 530	* 540	* 530	545	± 45   *± 55
VPM MPV	µm <sup>3</sup> ; fl	9.5		8.6			10.2	9.6	9.7	9.5	± 2.0
LYMPHO	#	N/A		3.6			N/A	N/A	N/A	N/A	± 1.00
	%	N/A		17.0			N/A	N/A	N/A	N/A	± 5.0
MONO	#	N/A		0.9			N/A	N/A	N/A	N/A	± 0.80
	%	N/A		4.1			N/A	N/A	N/A	N/A	± 4.0
GRANULO	#	N/A		16.7			N/A	N/A	N/A	N/A	± 1.40
	%	N/A		78.9			N/A	N/A	N/A	N/A	± 7.0

\* For ABX Pentra 60 Pentra 80 instruments only